

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 12:16

Date of Inspection: 12/28/2024

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.080	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

12/28/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 11:14

Date of Inspection: 09/22/2024

Serial Number: 80-005249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402F Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000 / 0.000	0.049	0.080	0.200	0.079
0.000 / 0.000	0.050	0.080	0.201	0.079
/ 0.000	0.050	0.080	0.201	0.079

Number of Simulators Used: 5

Remarks:

00: RFI Detect UNINTENTIONAL KEYING OF RADIO.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




PATRICK M BAKER

Signature and Printed Name

09/22/2024
Date

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/22/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.000	09:34
Air Blank	0.000	09:35
Control Test	0.000	09:35
Air Blank	0.000	09:36
Control Test	0.000	09:36
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/22/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	INT*	09:40
Air Blank	0.000	09:41
Control Test	INT*	09:42
Air Blank	0.000	09:42
Control Test	INT*	09:43
Air Blank	0.000	09:43
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


*Interferent Detect


Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/22/2024
Software: 8100.27

Test	g/210L	Time


Air Blank	0.000	09:44
Control Test	0.049	09:45
Air Blank	0.000	09:45
Control Test	0.049	09:46
Air Blank	0.000	09:46
Control Test	0.049	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/22/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test	0.079	09:53
Air Blank	0.000	09:54
Control Test	0.080	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	


Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/22/2024
Software: 8100.27

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	09:57
Control Test	0.199	09:57
Air Blank	0.000	09:58
Control Test	0.200	09:59
Air Blank	0.000	09:59
Control Test	0.200	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 20:57

Date of Inspection: 06/25/2024

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.027 / 0.000	0.041 / 0.046	0.074 / 0.078	0.188 / 0.196	0.079
0.007 / 0.000	0.044 / 0.046	0.076 / 0.078	0.193 / 0.197	0.079
0.000 / 0.000	0.045 / 0.047	0.077 / 0.078	0.195 / 0.198	0.079

Number of Simulators Used: 5

Remarks:

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Utilization of new simulators resulted in trial and error. BSW

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

06/25/2024
Date

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
08/15/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.000	09:35
Air Blank	0.000	09:36
Control Test	0.000	09:37
Air Blank	0.000	09:37
Control Test	0.000	09:38
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

 1463

Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
08/15/2024
Software: 8100.27

Test	g/210L	Time

Air Blank	0.000	09:48
Control Test	0.078	09:48
Air Blank	0.000	09:49
Control Test	0.077	09:50
Air Blank	0.000	09:50
Control Test	0.079	09:51
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

Sgt Paul 1465
Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
08/15/2024
Software: 8100.27

Test	g/210L	Time

Air Blank	0.000	09:52
Control Test	0.049	09:53
Air Blank	0.000	09:53
Control Test	0.049	09:54
Air Blank	0.000	09:55
Control Test	0.049	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
08/15/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	INT*	10:00
Air Blank	0.000	10:01
Control Test	INT*	10:02
Air Blank	0.000	10:02
Control Test	INT*	10:03
Air Blank	0.000	10:03
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect


Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
08/15/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.198	10:05
Air Blank	0.000	10:05
Control Test	0.199	10:06
Air Blank	0.000	10:07
Control Test	0.199	10:07
Air Blank	0.000	10:08
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 10:16

Date of Inspection: 08/15/2024

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

A F / M A: Sequence Aborted DUE TO STOPPING BREATH WITH M/A. SEQUENCE RESTARTED DUE TO STOPPING M/A BREATH TOO EARLY

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

08/15/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 12:16

Date of Inspection: 07/23/2024

Serial Number: 80-005249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.078	0.198	0.077
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.079	0.199	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name

07/23/2024
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Serial Number: 80-005249

Time of Inspection: 21:53

Date of Inspection: 05/28/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 28423080A3 Exp: 11/05/2025
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Pat M Baker

PATRICK M BAKER

Signature and Printed Name

05/28/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 17:07

Date of Inspection: 03/19/2024

Serial Number: 80-005249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies : ☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt P. Baker

PATRICK M. BAKER

Signature and Printed Name

03/19/2024
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 20:51

Date of Inspection: 02/28/2024

Serial Number: 80-005249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.079	0.200	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.080	0.201	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

PATRICK M BAKER

02/28/2024
Date