

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 11:50

Date of Inspection: 09/05/2024

Serial Number: 80-006385

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045	0.074 / 0.074		
0.000	0.046	0.075 / 0.074		
0.000	0.046	0.075 / 0.075		

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: AI B. KROME #263389 PRESENT FOR INSPECTION..

BK 263389

*Instrument received a Control Outside Tolerance.
Dept. Inspector Armstrong was contacted.
Instrument will be sent to FDLE for evaluation.*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny Hammond 264409

JENNY J HAMMOND
Signature and Printed Name

09/05/2024
Date

*KB 13893
9-9-24*

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 16:41

Date of Inspection: 08/05/2024

Serial Number: 80-006385

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28423080A2 Exp: 11/05/2025
0.000	0.044 / 0.045	0.075	0.195	0.080
0.000	0.045 / 0.046	0.076	0.196	0.080
0.000	0.045 / 0.046	0.076	0.195	0.080

Number of Simulators Used: 5

Remarks:

05: Control Outside. AI KROME 263389 PRESENT FOR INSPECTION

JK
263389

The connections between the instrument and the simulator was checked. The instrument passed and was in compliance. I spoke with FDLE Department Inspector Taylor Gutschow.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny J Hammond

JENNY J HAMMOND

Signature and Printed Name

08/05/2024
Date

8/5
157693