

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 14:07

Date of Inspection: 08/05/2024

Serial Number: 80-006567

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 29122080A1 Exp: 12/05/2024
0.000	0.045	0.076	0.193	0.090 / 0.081
0.000	0.046	0.076	0.194	0.089 / 0.081
0.000	0.046	0.076	0.194	0.089 / 0.081

Number of Simulators Used: 5

Remarks:

08: Control Outside. AI KROME 263389 PRESENT FOR INSPECTION

BC 263389
The connection between the instrument and the Dry Gas Standard Supply Hose was checked. The instrument passed and was in compliance. I spoke with FDLE Department Inspector Taylor Gutschow.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny Hammond

JENNY J HAMMOND

Signature and Printed Name

08/05/2024

Date

8/10 15293

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CSO
Time of Inspection: 12:53

Date of Inspection: 03/06/2024

Serial Number: 80-006567
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045	0.073 / 0.074		
0.000	0.046	0.074 / 0.074		
0.000	0.046	0.074 / 0.074		

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. SIM TEMP WITHIN RANGE, TWICE. AI J. DAIL #254236 PRESENT FOR INSPECTI
ON.Non-compliance:0.080 CONTROL OUTSIDE TOLERANCE X2./ *inspector*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Stephanie R Knox 26/252

STEPHANIE R KNOX

Signature and Printed Name

03/06/2024
Date