

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 14:19

Date of Inspection: 09/05/2024

Serial Number: 80-006567
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045	0.074 / 0.074		
0.000	0.046	0.074 / 0.075		
0.000	0.046	0.075 / 0.075		

Number of Simulators Used: 5

Remarks:

Time-Date changed. 08: Control Outside Tolerance ut08: CONTROL OUTSIDE TOLERANCE, CONTROL OUTSIDE TOLERANCE. Non-compliance: AI B. KROME #263389 PRESENT FOR INSPECTION. *BC*

Instrument received a Control Outside Tolerance.

*Dept. Inspector Armstrong was contacted.
Instrument will be sent to FDLE for evaluation.*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny Hammond #24409

JENNY J HAMMOND
Signature and Printed Name

09/05/2024
Date

*K/H 13693
9-9-24*