

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TALLAHASSEE PD

Time of Inspection: 09:59

Date of Inspection: 10/22/2024

Serial Number: 80-006926

Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202303J<br>Exp: 03/28/2025 | 0.08g/210L Test (g/210L)<br>Lot#:202307A<br>Exp: 07/11/2025 | 0.20g/210L Test (g/210L)<br>Lot#:202307C<br>Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:33922080A1<br>Exp: 01/05/2025 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000                      | 0.048                                                       | 0.077                                                       | 0.212 / 0.204                                               | 0.082                                                                       |
| 0.000                      | 0.048                                                       | 0.077                                                       | 0.209 / 0.202                                               | 0.082                                                                       |
| 0.000                      | 0.048                                                       | 0.077                                                       | 0.208 / 0.199                                               | 0.083                                                                       |


Number of Simulators Used: 5

**Remarks:**

20: Control Outside Tolerance.

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



T&D 321

JAMES J HARRIS

Signature and Printed Name

10/22/2024  
Date