

AMENDMENT
Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 10:24

Date of Inspection: 02/07/2024

Serial Number: 80-007381
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/14/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402477283 Exp: 02/07/2024
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.077	0.197	0.081
0.000	0.049	0.078	0.197	0.082

Number of Simulators Used: 5

Remarks:


J. GUZMAN IN TRAINING.

QA 0.08 g/210L DGS: EXP: 06/24/2025

QA AMENDMENT DATE: 01/20/2025

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name ANAYA S FRAZIER
Date 02/07/2024