

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO  
Time of Inspection:

Serial Number: 80-007434  
Date of Inspection: December 2024

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: \_\_\_\_\_

Remarks: Inspection was not conducted as the instrument was sent to FDLE for evaluation.

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Anaya Frazier

Signature and Printed Name

12/11/2024  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: Broward County SO

Time of Inspection:

Software: 8100.27

Serial Number: 80-007434

Date of Inspection: November 2024

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: \_\_\_\_\_

Remarks: Inspection was not conducted as the instrument was sent to FDLE for evaluation.

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Anaya Frazier

Signature and Printed Name

11/06/2024

Date

**AMENDMENT**  
Florida Department of Law Enforcement  
Alcohol Testing Program

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: BROWARD COUNTY SO  
Time of Inspection: 10:52

Date of Inspection: 03/06/2024

Serial Number: 80-007434  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: <del>402623838</del> <b>Q4</b> Exp: 12/22/2025
0.000	0.049	0.079	0.203	0.079
0.000	0.050	0.079	0.203	0.079
0.000	0.050	0.079	0.203	0.079

Number of Simulators Used: 5

Remarks:

**Q4 0.08g/210L DGS: LOT #: 402623828**

**Q4 3/11/2024**

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

ANAYA S FRAZIER

03/06/2024  
Date

**AMENDMENT**  
Florida Department of Law Enforcement  
Alcohol Testing Program

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: BROWARD COUNTY SO  
Time of Inspection: 11:09

Date of Inspection: 02/07/2024

Serial Number: 80-007434  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) <del>Lot#:402623830</del> Q4 Exp: 12/22/2025
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.080	0.203	0.080

Number of Simulators Used: 5

**Remarks:**


J. GUZMAN IN TRAINING.

Q4 0.08g/210L DGS: LOT #: 402623828

Q4 3/11/2024

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
ANAYA S FRAZIER  
Signature and Printed Name  
02/07/2024  
Date