

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 12:11

Date of Inspection: 09/05/2024

Serial Number: 80-007480

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:29122080A1 Exp: 12/05/2024
0.000	0.047	0.076	0.194	0.079
0.000	0.048	0.076	0.196	0.079
0.000	0.047	0.076	0.195	0.080

Number of Simulators Used: 5

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION. BK 263389

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny Hammond 264409 JENNY J HAMMOND
Signature and Printed Name

09/05/2024
Date

Handwritten: 264409
9-9-24

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 10:48

Date of Inspection: 09/05/2024

Serial Number: 80-007480

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI ABORTED. DIAGNOSTIC FAIL. Non-compliance: AI ABORTED. DIAGNOSTIC FAIL..

BL 263389

Instrument initially received Diagnostic Fail.
Dept. Inspector Armstrong was contacted and she had me run a Diagnostic. Diagnostic passed all tests.
Per Dept. Inspector Armstrong, an Agency Inspection was conducted. The Agency Inspection passed and was in compliance.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jenny J. Hammond 264409

JENNY J HAMMOND

Signature and Printed Name

09/05/2024
Date

KB 132893
9-9-24

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007480
09/05/2024
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK