

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 09:41

Date of Inspection: 03/06/2024

Serial Number: 80-007482
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.076	0.189 / 0.187	
0.000	0.047	0.075	0.190 / 0.188	
0.000	0.046	0.075	0.189 / 0.187	

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. SIM TEMP WITHIN RANGE., TWICE. AI J. DALL #254236 PRESENT FOR INSPECTI
ON. Non-compliance: 0.200 CONTROL OUTSIDE TOLERANCE X2./

Guidance

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 26/252

STEPHANIE R KNOX

Signature and Printed Name

03/06/2024
Date