

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 20:24

Date of Inspection: 06/26/2024

Serial Number: 80-007492

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24923080A3 Exp: 10/05/2025
0.000	0.048	0.078	0.198	0.000 / 0.000
0.000	0.048	0.078	0.198	0.000 / 0.000
0.000	0.049	0.078	0.198	0.000 / 0.000


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: CLOG. - POSSIBLE REGULATOR FAILURE. SPOKE WITH REGIONAL INSPECTOR, TAYLOR GUTSCHOW, AND I ADVISED HIM OF THE ISSUE. I WILL BE SENDING IT TO HIM ~~FOR REPAIR~~ TO REPLACE THE REGULATOR.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAMON J PULASKI
 Signature and Printed Name

06/26/2024
 Date