



INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law EnforcementS/N 80-000206Florida Department of
Law EnforcementDate In N/ADI Completion Date 09/19/2024☐ Ship☐ P/U☐ H/D☐ CMI☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																														
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>DSP Fail.</u>		<u>09/12/2024</u>	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>197</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		<u>09/17/2024</u>	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	<table border="1"><thead><tr><th colspan="2">Maintenance By TDG</th><th>Date (See Notes)</th></tr></thead><tbody><tr><td colspan="2"><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td colspan="2"><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td colspan="2"><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td colspan="2"><input checked="" type="checkbox"/> Other <u>CMI remotely resolved the DSP Fail on 9/13 and reloaded the forms on 9/17. Tared the tank sensor on 9/17 (it displayed ~11 PSI without a tank attached).</u></td><td></td></tr></tbody></table>			Maintenance By TDG		Date (See Notes)	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input checked="" type="checkbox"/> Other <u>CMI remotely resolved the DSP Fail on 9/13 and reloaded the forms on 9/17. Tared the tank sensor on 9/17 (it displayed ~11 PSI without a tank attached).</u>		
Simulator	Serial #	Lot #/Exp																																				
0.050	MP6286	202303K 03/29/2025																																				
0.080	MP6287	202303L 03/29/2025																																				
0.200	MP6288	202304C 04/05/2025																																				
0.080 DGS	N/A	01923080A3 02/05/2025																																				
Maintenance By TDG		Date (See Notes)																																				
<input type="checkbox"/> Battery Replacement																																						
<input type="checkbox"/> Dry Gas Regulator Replacement																																						
<input type="checkbox"/> Breath Tube Replacement																																						
<input checked="" type="checkbox"/> Other <u>CMI remotely resolved the DSP Fail on 9/13 and reloaded the forms on 9/17. Tared the tank sensor on 9/17 (it displayed ~11 PSI without a tank attached).</u>																																						
Calibration Adjustment			Department Inspection																																			
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>26932</u>																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>							
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
<input type="checkbox"/> Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																					
Simulator	Serial Number																																					
0.000	MP6284																																					
Interferent	MP6285																																					
0.050	MP6286																																					
0.080	MP6287																																					
0.200	MP6288																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____															
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
Notes/Suggested Service: Instrument will be retained at FDLE until requested by an agency. (TDG)			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																			
			Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2024.09.24 09:56:31 -0400</small>																																			
			Shayla Platt <small>Digitally signed by Shayla Platt Date: 2024.09.24 12:55:23 -0400</small>																																			
			Tech Review / Date _____ Admin Review / Date _____																																			

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 11:56

Date of Inspection: 09/17/2024

Serial Number: 80-000206

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *MG*
09/17/2024

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.





Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

09/17/2024
Date

Stability Checks

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																													
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																													
<div>FDLE Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-000205 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr></tbody></table> <div>Operator's Signature </div>		Test	g/210L	Time	Air Blank	0.000	12:52	Control Test	0.050	12:53	Air Blank	0.000	12:53	Control Test	0.050	12:54	Air Blank	0.000	12:54	Control Test	0.049	12:55	Air Blank	0.000	12:56	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<div>FDLE Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-000205 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> <div>Operator's Signature </div>		Test	g/210L	Time	Air Blank	0.000	12:19	Control Test	0.079	12:20	Air Blank	0.000	12:20	Control Test	0.078	12:21	Air Blank	0.000	12:21	Control Test	0.079	12:22	Air Blank	0.000	12:22	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<div>FDLE Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-000205 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.196</td><td>12:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.195</td><td>12:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr><tr><td>Control Test</td><td>0.195</td><td>12:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1953</td><td></td></tr><tr><td>Std Dev</td><td>0.0016</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2956</td><td></td></tr></tbody></table> <div>Operator's Signature </div>		Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.196	12:26	Air Blank	0.000	12:26	Control Test	0.195	12:27	Air Blank	0.000	12:28	Control Test	0.195	12:28	Air Blank	0.000	12:29	Control Test Stats			Average	0.1953		Std Dev	0.0016		Rel Std Dev(%)	0.2956		<div>Printed to external printer. Results in range. Will attach.</div> <div> 9/17/2024</div>	
Test	g/210L	Time																																																																																																																	
Air Blank	0.000	12:52																																																																																																																	
Control Test	0.050	12:53																																																																																																																	
Air Blank	0.000	12:53																																																																																																																	
Control Test	0.050	12:54																																																																																																																	
Air Blank	0.000	12:54																																																																																																																	
Control Test	0.049	12:55																																																																																																																	
Air Blank	0.000	12:56																																																																																																																	
Control Test Stats																																																																																																																			
Average	0.0497																																																																																																																		
Std Dev	0.0006																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																		
Test	g/210L	Time																																																																																																																	
Air Blank	0.000	12:19																																																																																																																	
Control Test	0.079	12:20																																																																																																																	
Air Blank	0.000	12:20																																																																																																																	
Control Test	0.078	12:21																																																																																																																	
Air Blank	0.000	12:21																																																																																																																	
Control Test	0.079	12:22																																																																																																																	
Air Blank	0.000	12:22																																																																																																																	
Control Test Stats																																																																																																																			
Average	0.0787																																																																																																																		
Std Dev	0.0006																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																		
Test	g/210L	Time																																																																																																																	
Air Blank	0.000	12:25																																																																																																																	
Control Test	0.196	12:26																																																																																																																	
Air Blank	0.000	12:26																																																																																																																	
Control Test	0.195	12:27																																																																																																																	
Air Blank	0.000	12:28																																																																																																																	
Control Test	0.195	12:28																																																																																																																	
Air Blank	0.000	12:29																																																																																																																	
Control Test Stats																																																																																																																			
Average	0.1953																																																																																																																		
Std Dev	0.0016																																																																																																																		
Rel Std Dev(%)	0.2956																																																																																																																		
0.077 to 0.083		0.077 to 0.083		0.077 to 0.083		0.077 to 0.083																																																																																																													

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000206

09/17/2024

Software: 8100.27

DGS

Test	g/210L	Time
Air Blānk	0.000	12:12
Control Test	0.078	12:12
Air Blank	0.000	12:13
Control Test	0.079	12:13
Air Blank	0.000	12:13
Control Test	0.079	12:14
Air Blank	0.000	12:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

MC

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Serial Number: 80-000206

Time of Inspection: 13:21

Date of Inspection: 09/19/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 01923080A3 Exp: 02/05/2025
0.000	0.049	0.078	0.196	0.078
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.078	0.196	0.078
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.078	0.196	0.078
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.078	0.196	0.079
0.000	0.050	0.078	0.197	0.079

Standard Deviations	0.0003	0.0000	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/19/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000206, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000206	UNCERTAINTY* \pm
Owning Agency:	FDLE	0.050 g/ 210 L 0.004
Calibration Date:	09/19/2024	0.080 g/ 210 L 0.004
Calibration Time:	13:21	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

09/19/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 10:02

Date of Inspection: 02/06/2024

Serial Number: 80-000206

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/06/2024
Date

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																												
<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 02/06/2024 SN 80-000206 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.052</td><td>10:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0513</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1247</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	10:42	Control Test	0.052	10:43	Air Blank	0.000	10:43	Control Test	0.051	10:44	Air Blank	0.000	10:45	Control Test	0.051	10:45	Air Blank	0.000	10:46	Control Test Stats			Average	0.0513		Std Dev	0.0006		Rel Std Dev(%)	1.1247		<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 02/06/2024 SN 80-000206 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	10:50	Control Test	0.080	10:50	Air Blank	0.000	10:51	Control Test	0.080	10:51	Air Blank	0.000	10:52	Control Test	0.079	10:53	Air Blank	0.000	10:53	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 02/06/2024 SN 80-000206 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	11:16	Control Test	0.198	11:17	Air Blank	0.000	11:17	Control Test	0.197	11:18	Air Blank	0.000	11:18	Control Test	0.197	11:19	Air Blank	0.000	11:20	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		<p>DGS 0.08g/210L 0.077 to 0.083</p> <p>Inadvertently printed to external printer. Results within acceptable range. Will attach.</p> <p>MC 02/06/2024</p>
Test	g/210L	Time																																																																																																													
Air Blank	0.000	10:42																																																																																																													
Control Test	0.052	10:43																																																																																																													
Air Blank	0.000	10:43																																																																																																													
Control Test	0.051	10:44																																																																																																													
Air Blank	0.000	10:45																																																																																																													
Control Test	0.051	10:45																																																																																																													
Air Blank	0.000	10:46																																																																																																													
Control Test Stats																																																																																																															
Average	0.0513																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1247																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	10:50																																																																																																													
Control Test	0.080	10:50																																																																																																													
Air Blank	0.000	10:51																																																																																																													
Control Test	0.080	10:51																																																																																																													
Air Blank	0.000	10:52																																																																																																													
Control Test	0.079	10:53																																																																																																													
Air Blank	0.000	10:53																																																																																																													
Control Test Stats																																																																																																															
Average	0.0797																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.7247																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:16																																																																																																													
Control Test	0.198	11:17																																																																																																													
Air Blank	0.000	11:17																																																																																																													
Control Test	0.197	11:18																																																																																																													
Air Blank	0.000	11:18																																																																																																													
Control Test	0.197	11:19																																																																																																													
Air Blank	0.000	11:20																																																																																																													
Control Test Stats																																																																																																															
Average	0.1973																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2926																																																																																																														

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000206

02/06/2024

Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.080	10:31
Air Blank	0.000	10:32
Control Test	0.079	10:32
Air Blank	0.000	10:32
Control Test	0.080	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

MG

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:24

Date of Inspection: 02/06/2024

Serial Number: 80-000206

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.195	0.079
0.000	0.049	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079


Standard Deviations	0.0011	0.0009	0.0011	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/06/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000206, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000206	UNCERTAINTY* \pm
Owning Agency:	FDLE	0.050 g/ 210 L 0.004
Calibration Date:	02/06/2024	0.080 g/ 210 L 0.004
Calibration Time:	13:24	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/06/2024

Date

Jay D Gutschow
TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality