



INSTRUMENT PROCESSING SHEET

Agency Lafayette County SOS/N 80-000773Florida Department of
Law EnforcementDate In 07/31/2024 DI Completion Date 8/13/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>07/31/2024</u>		Quality Checks By <u>DA</u> Date <u>08/06/24</u>		Flow Calibration By _____ Date _____																																																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>dropoff, with DGS attached</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>162</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																															
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 04/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 03/29/2025	0.080	MP5089	202303L 03/29/2025	0.200	MP5090	202304C 04/05/2025	0.080 DGS	N/A	06723080A5 04/05/2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP5088	202303K 03/29/2025																																																																	
0.080	MP5089	202303L 03/29/2025																																																																	
0.200	MP5090	202304C 04/05/2025																																																																	
0.080 DGS	N/A	06723080A5 04/05/2025																																																																	
Calibration Adjustment By _____				Department Inspection By <u>PN</u>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28427</u> Gauge <u>1012</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Int. Print Check</u></td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Int. Print Check</u>
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP6289																																																																		
Interferent	MP6290																																																																		
0.050	MP6291																																																																		
0.080	MP6292																																																																		
0.200	MP6293																																																																		
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Int. Print Check</u>																																																																		
Notes/Suggested Service: <u>Unplugged external printer during stability checks so first run did not print. Checked interna printer and re-ran 0.05g/210L stability check. DA 8/6/24</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																															
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.08.22 10:55:57 -0400				Shayla Platt Digitally signed by Shayla Platt Date: 2024.08.22 14:54:36 -0400																																																															
Tech Review / Date _____				Admin Review / Date _____																																																															

Stability Checks 80-000773 8/6/24 DA

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
08/06/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:36
Control Test	0.048	10:37
Air Blank	0.000	10:37
Control Test	0.049	10:38
Air Blank	0.000	10:38
Control Test	0.048	10:39
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Wet



Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
08/06/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:42
Control Test	0.078	10:42
Air Blank	0.000	10:43
Control Test	0.078	10:44
Air Blank	0.000	10:44
Control Test	0.077	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Wet



Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
08/06/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.196	10:47
Air Blank	0.000	10:48
Control Test	0.198	10:48
Air Blank	0.000	10:49
Control Test	0.197	10:50
Air Blank	0.000	10:50
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	




Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
08/06/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:54
Control Test	0.079	10:55
Air Blank	0.000	10:55
Control Test	0.080	10:56
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS



Operator's Signature

Internal Printer Check

80-000773 8/6/24 DA

LAFAYETTE CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000773
08/06/2024
Software: 8100.27

***** Printer Test *****

40 characters per line
abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Double Char/line
abcdefghijklmnopqrst
1234567890-=!@#%^&*
ABCDEFGHIJKLMNQRST

WIDE Char/line
abcdefghijklmnopqrst
1234567890-=!@#%^&*
ABCDEFGHIJKLMNQRST

TALL Char/line
abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNQRSTUWXYZ!@#%^&*()_+?

***** Test End *****

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 10:20

Date of Inspection: 08/06/2024

Serial Number: 80-000773
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

NA-Compliance not determined 8/6/24 DA

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DESTINEE N ARMSTRONG

Signature and Printed Name

08/06/2024
Date

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 12:50

Date of Inspection: 08/13/2024

Serial Number: 80-000773
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.078	0.197	0.081
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.082
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.081

Standard Deviations	0.0005	0.0004	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:


The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO
Signature and Printed Name

08/13/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000773 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000773</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAFAYETTE CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/13/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:50</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

08/13/2024

Date

PHIL NICODEMO,

Department Inspector

Service • Integrity • Respect • Quality