



INSTRUMENT PROCESSING SHEET

Agency Haines City PDS/N 80-000796Florida Department of
Law EnforcementDate In 05/02/2024 DI Completion Date 5/6/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE


Intake By <u>ALL</u> Date <u>05/02/2024</u>		Quality Checks By <u>DA</u> Date <u>5/3/24</u>		Flow Calibration By _____ Date _____																																																													
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Bypassed agency inspection to operate instrument</u> <u>DA</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>212</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP5088</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 04/05/2025</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 03/29/2025	0.080	MP5089	202303L 03/29/2025	0.200	MP5090	202304C 04/05/2025	0.080 DGS	N/A	06723080A5 04/05/2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5088	202303K 03/29/2025																																																															
0.080	MP5089	202303L 03/29/2025																																																															
0.200	MP5090	202304C 04/05/2025																																																															
0.080 DGS	N/A	06723080A5 04/05/2025																																																															
Calibration Adjustment By _____				Department Inspection By <u>PN</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6289																																																																
Interferent	MP6290																																																																
0.050	MP6291																																																																
0.080	MP6292																																																																
0.200	MP6293																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																													
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																													
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.05.07 15:56:20 -04'00'		Shayla Platt Digitally signed by Shayla Platt Date: 2024.05.09 08:31:02 -04'00'																																																													
		Tech Review / Date _____		Admin Review / Date _____																																																													

Stability Checks 80-000796 5/3/24 DA

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
05/03/2024
Software: 8100.27

	Test	g/210L	Time
Air Blank	Air Blank	0.000	10:55
Control Test	Control Test	0.049	10:56
Air Blank	Air Blank	0.000	10:56
Control Test	Control Test	0.049	10:57
Air Blank	Air Blank	0.000	10:58
Control Test	Control Test	0.049	10:58
Air Blank	Air Blank	0.000	10:59
Control Test Stats	Control Test Stats		
Average	Average	0.0490	
Std Dev	Std Dev	0.0000	
Rel Std Dev	Rel Std Dev (%)	0.0000	

PN 5/6/2024




Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
05/03/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.078	11:02
Air Blank	0.000	11:02
Control Test	0.078	11:03
Air Blank	0.000	11:03
Control Test	0.077	11:04
Air Blank	0.000	11:05

Control Test Stats	Control Test Stats
Average	Average 0.0777
Std Dev	Std Dev 0.0006
Rel Std Dev	Rel Std Dev(%) 0.7434

PN 5/6/2024



Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
05/03/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.195	11:07
Air Blank	0.000	11:08
Control Test	0.196	11:09
Air Blank	0.000	11:09
Control Test	0.195	11:10
Air Blank	0.000	11:10

Control Test Stats	Control Test Stats
Average	Average 0.1953
Std Dev	Std Dev 0.0006
Rel Std Dev	Rel Std Dev (%) 0.2956

PN 5/6/2024


Operator's Signature

DGS


HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
05/03/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:12
Control Test	0.080	11:13
Air Blank	0.000	11:13
Control Test	0.081	11:14
Air Blank	0.000	11:14
Control Test	0.080	11:14
Air Blank	0.000	11:15

Control Test Stats

Average	0.0803
Std Dev	0.0006
Rel Std Dev	0.7187

PN 5/6/2024


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Time of Inspection: 10:43

Date of Inspection: 05/03/2024

Serial Number: 80-000796

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____


Remarks:

BYPASSED AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

compliance not determined 5/3/24 DA

~~The above instrument complies () does not comply () with Chapter 11D-8, FAC.~~

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

05/03/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Serial Number: 80-000796

Time of Inspection: 10:50

Date of Inspection: 05/06/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 06723080A5 Exp: 04/05/2025
0.000	0.049	0.077	0.205	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080


Standard Deviations	0.0005	0.0004	0.0015	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 PHIL NICODEMO
 Signature and Printed Name
05/06/2024
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000796 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000796</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HAINES CITY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/06/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:50</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/06/2024

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1