



INSTRUMENT PROCESSING SHEET

Agency City of Miramar PoliceS/N 80-001067Florida Department of
Law EnforcementDate In 01/10/2024 DI Completion Date 01/25/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Mini keyboard included.</u>		<u>01/11/2024</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>209</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>01/11/2024</u>	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5095</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5096</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202303K 03/29/2025	0.080	MP5095	202303L 03/29/2025	0.200	MP5096	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025																		
Simulator	Serial #	Lot #/Exp																																		
0.050	MP5094	202303K 03/29/2025																																		
0.080	MP5095	202303L 03/29/2025																																		
0.200	MP5096	202304C 04/05/2025																																		
0.080 DGS	N/A	01923080A3 02/05/2025																																		
Calibration Adjustment			Department Inspection																																	
By TDG			By TDG																																	
Barometric Pressure Gauge <u>1023</u> ID # <u>68639</u>			Barometric Pressure ID# <u>26932</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>23340</td><td>09/18/2025</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	22310	08/11/2024	0.200	MP5100	23340	09/18/2025	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG222203	08/10/2024	Gauge <u>1021</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>					
Simulator	Serial #	Lot #	Expiration																																	
0.000	MP5097	N/A	N/A																																	
0.040	MP5098	23400	10/24/2025																																	
0.100	MP5099	22310	08/11/2024																																	
0.200	MP5100	23340	09/18/2025																																	
0.300	MP5101	22220	06/15/2024																																	
0.080 DGS	N/A	AG222203	08/10/2024																																	
☒ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096																
Simulator	Serial Number																																			
0.000	MP5092																																			
Interferent	MP5093																																			
0.050	MP5094																																			
0.080	MP5095																																			
0.200	MP5096																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202303K</td><td>03/29/2025</td></tr><tr><td>0.080</td><td>MP5095</td><td>202303L</td><td>03/29/2025</td></tr><tr><td>0.200</td><td>MP5096</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3</td><td>02/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202303K	03/29/2025	0.080	MP5095	202303L	03/29/2025	0.200	MP5096	202304C	04/05/2025	0.080 DGS	N/A	01923080A3	02/05/2025	Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____													
Simulator	Serial #	Lot #	Expiration																																	
0.050	MP5094	202303K	03/29/2025																																	
0.080	MP5095	202303L	03/29/2025																																	
0.200	MP5096	202304C	04/05/2025																																	
0.080 DGS	N/A	01923080A3	02/05/2025																																	
Notes/Suggested Service: _____ Tech Review: Added two "/" for 0.300 expiration date. (TDG 1/26/2024)			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
			Digitally signed by Benjamin Siddoway Date: 2024.01.26 10:18:48 -05'00'																																	
			Shayla Platt Date: 2024.01.29 09:09:53 -05'00'																																	
			Tech Review / Date Admin Review / Date																																	

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤ 0.003 of Wet X																																																																																																																																				
CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/11/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:05	Control Test	0.049	14:06	Air Blank	0.000	14:06	Control Test	0.049	14:07	Air Blank	0.000	14:07	Control Test	0.048	14:08	Air Blank	0.000	14:09	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/11/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:12	Control Test	0.078	14:13	Air Blank	0.000	14:13	Control Test	0.077	14:14	Air Blank	0.000	14:14	Control Test	0.078	14:15	Air Blank	0.000	14:16	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/11/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:19	Control Test	0.198	14:19	Air Blank	0.000	14:20	Control Test	0.197	14:20	Air Blank	0.000	14:21	Control Test	0.197	14:22	Air Blank	0.000	14:22	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		065 CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/11/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:56</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:57</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:56	Control Test	0.081	13:56	Air Blank	0.000	13:57	Control Test	0.081	13:57	Air Blank	0.000	13:58	Control Test	0.081	13:58	Air Blank	0.000	13:59	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	14:05																																																																																																																																					
Control Test	0.049	14:06																																																																																																																																					
Air Blank	0.000	14:06																																																																																																																																					
Control Test	0.049	14:07																																																																																																																																					
Air Blank	0.000	14:07																																																																																																																																					
Control Test	0.048	14:08																																																																																																																																					
Air Blank	0.000	14:09																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0487																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1863																																																																																																																																						
Air Blank	0.000	14:12																																																																																																																																					
Control Test	0.078	14:13																																																																																																																																					
Air Blank	0.000	14:13																																																																																																																																					
Control Test	0.077	14:14																																																																																																																																					
Air Blank	0.000	14:14																																																																																																																																					
Control Test	0.078	14:15																																																																																																																																					
Air Blank	0.000	14:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0777																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7434																																																																																																																																						
Air Blank	0.000	14:19																																																																																																																																					
Control Test	0.198	14:19																																																																																																																																					
Air Blank	0.000	14:20																																																																																																																																					
Control Test	0.197	14:20																																																																																																																																					
Air Blank	0.000	14:21																																																																																																																																					
Control Test	0.197	14:22																																																																																																																																					
Air Blank	0.000	14:22																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1973																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2926																																																																																																																																						
Air Blank	0.000	13:56																																																																																																																																					
Control Test	0.081	13:56																																																																																																																																					
Air Blank	0.000	13:57																																																																																																																																					
Control Test	0.081	13:57																																																																																																																																					
Air Blank	0.000	13:58																																																																																																																																					
Control Test	0.081	13:58																																																																																																																																					
Air Blank	0.000	13:59																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0810																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

CITY OF MIRAMAR POL
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2024
SN 80-001057
09:46:37

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 20

Optical Calibration
Adjustment
By: TDG

Page 1 of 2

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	-0.000	0.000	
0.040	0.040	-0.0000	
0.100	0.100	-0.0003	
0.200	0.200	0.0003	
0.300	0.300	-0.0001	

1 Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0002
0.100	0.100	-0.0003
0.200	0.200	0.0004
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%% *
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 2827.00
 Sample #2 = 2781.00
 Sample #3 = 2805.00
 Sample #4 = 2830.00
 Average Result = 2805.3333
 STD DEV = 24.5017
 REL STD DEV = 0.873

***** CHANNEL 2 *****
 Sample #1 = 3341.00
 Sample #2 = 3346.00
 Sample #3 = 3354.00
 Sample #4 = 3379.00
 Average Result = 3359.6667
 STD DEV = 17.2143
 REL STD DEV = 0.512

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1022
 3 um H2O Adjust (mg/l*10,000) = 1004
 9 um H2O Adjust (mg/l*10,000) = 450
 ***** AUTO CAL PASS *****

Page 2/2

80-001067
 01/25/2024

Optical Calibration Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓ CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/25/2024 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:45 Control Test 0.050 12:45 Air Blank 0.000 12:46 Control Test 0.050 12:47 Air Blank 0.000 12:47 Control Test 0.049 12:48 Air Blank 0.000 12:48 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 Operator's Signature 	✓ CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/25/2024 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:38 Control Test 0.080 12:39 Air Blank 0.000 12:39 Control Test 0.079 12:40 Air Blank 0.000 12:41 Control Test 0.079 12:41 Air Blank 0.000 12:42 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 Operator's Signature 	✓ CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/25/2024 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:31 Control Test 0.199 12:32 Air Blank 0.000 12:33 Control Test 0.198 12:33 Air Blank 0.000 12:34 Control Test 0.197 12:35 Air Blank 0.000 12:35 Control Test Stats Average 0.1980 Std Dev 0.0010 Rel Std Dev(%) 0.5051 Operator's Signature 	005 CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 01/25/2024 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:27 Control Test 0.080 12:27 Air Blank 0.000 12:28 Control Test 0.080 12:28 Air Blank 0.000 12:28 Control Test 0.080 12:29 Air Blank 0.000 12:29 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CITY OF MIRAMAR POL
Time of Inspection: 14:45

Date of Inspection: 01/25/2024

Serial Number: 80-001067
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078


Standard Deviations	0.0003	0.0000	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

01/25/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001067, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001067</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CITY OF MIRAMAR POL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/25/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:45</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/25/2024

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality