

Agency Seminole County SO

S/N 80-001271

Date In 12/21/2023 DI Completion Date 1/26/2024

☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Calibration Adjustment

By _____

☐ Post Calibration Adjustment Stability ChecksDepartment InspectionBy BS

Notes/Suggested Service: Instrument would not print forms, and after CMI uploaded forms to instrument, it would still fail to print forms. Send to repair. (BS 1/26/2024)

Conduct an Agency Inspection Before Evidentiary

Shayla Digitally signed by **Shayla Platt**
Date: 2024.02.01 13:01:48 -05'00'

Taylor Digitally signed by **Taylor Gutschow**
Date: 2024.02.01 13:01:48 -05'00'

Gutschow Digitally signed by **Taylor Gutschow**
Date: 2024.02.01 13:01:48 -05'00'

Platt Digitally signed by **Shayla Platt**
Date: 2024.02.01 13:01:48 -05'00'

Taylor
Gutschow

Digitally signed by Taylor Gutschor
Date: 2024.02.01 13:01:48 -05'00'

Platt

Date: 2024.02.02
07:41:49 -05'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY SO

Serial Number: 80-001271

Time of Inspection: 15:50

Date of Inspection: 01/26/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG229803 Exp: 10/25/2024
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.202	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.202	0.079

Standard Deviations	0.0004	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY

Signature and Printed Name

01/26/2024

Date

Stability Checks

SEMINOLE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001271
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:24
Control Test	0.050	13:25
Air Blank	0.000	13:26
Control Test	0.049	13:26
Air Blank	0.000	13:27
Control Test	0.049	13:28
Air Blank	0.000	13:28
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Benjamin Sibley
Operator's Signature

SEMINOLE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001271
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:32
Control Test	0.080	13:33
Air Blank	0.000	13:33
Control Test	0.079	13:34
Air Blank	0.000	13:34
Control Test	0.078	13:35
Air Blank	0.000	13:36
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2558	

Benjamin Sibley
Operator's Signature

SEMINOLE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001271
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:42
Control Test	0.203	13:43
Air Blank	0.000	13:43
Control Test	0.200	13:44
Air Blank	0.000	13:44
Control Test	0.199	13:45
Air Blank	0.000	13:46
Control Test Stats		
Average	0.2007	
Std Dev	0.0021	
Rel Std Dev(%)	1.0374	

Benjamin Sibley
Operator's Signature

SEMINOLE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001271
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.080	13:51
Air Blank	0.000	13:52
Control Test	0.080	13:52
Air Blank	0.000	13:53
Control Test	0.081	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

Benjamin Sibley
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001271 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001271</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SEMINOLE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/26/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:50</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin Siddoway
Digitally signed by Benjamin Siddoway
Date: 2024.01.31 13:59:39 -05'00'

01/26/2024

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Keith Betham on 1/31/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001271

Bill To Address:

Keith Betham

Seminole County Sheriff's Office

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Dept of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument will not print forms, even after CMI connected remotely to upload forms. The
external printer will work with other instruments, so the instrument may have a bad printer port.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Keith Betham

Phone #: 407-665-3131 Email: kbetham@seminolesheriff.org

ATP Contact Name: Benjamin Siddoway ATP Email: BenjaminSiddoway@fdle.state.fl.us