



Agency Hillsborough County SO

S/N 80-003388

Date In 01-05-2024 DI Completion Date 1/26/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By BS

Barometric Pressure Gauge	1019	ID # 30793
---------------------------	------	------------

■ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Wet and Dry 0.080 g/210L
Stability Check avg. differed greater than 0.003. BS
~~Performed~~* calibration adjustment 1/25/2024. RFI from
cell phone during calibration adjustment, removed it
from the area. Fault detected, confirmed simulator was
not leaking and repeated 0.040 and 0.100 adjustments.
(BS). Simulator port loose during 0.200 g/210L test;
connection was secured, test repeated. PN 1/26/2024

*Performed (BS 1/29/2024)

Department Inspection

By PN

Barometric Pressure ID# 28421

Gauge 1019	Instrument 1019
------------	-----------------

Mouth Alcohol Solution Lot # 2023-A

Acetone Stock Solution Lot # 2022-B

Attachments

- | | |
|--|--|
| ☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☒ Calibration Adjustment | ☒ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other |
|--|--|

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Tech Review / Date

Shayla
Platt

Digitally signed by

Shayla Platt

Date: 2024.01.31

09:54:31 -05'00

Admin Review / Date 09:54:31 -05

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:06

Date of Inspection: 01/26/2024

Serial Number: 80-003388
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.052	0.080	0.000 / 0.201	0.080
0.000	0.051	0.081	0.200 / 0.202	0.080
0.000	0.051	0.080	0.199 / 0.201	0.080
0.000	0.051	0.080	0.198 / 0.200	0.080
0.000	0.051	0.083	0.201 / 0.202	0.080
0.000	0.051	0.084	0.200 / 0.201	0.080
0.000	0.051	0.083	0.200 / 0.202	0.079
0.000	0.051	0.084	0.201 / 0.202	0.080
0.000	0.053	0.083	0.201 / 0.201	0.079
0.000	0.053	0.084	0.201 / 0.202	0.080
Standard Deviations	0.0008	0.0017	0.0632 / 0.0006	0.0004


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance SIMULATOR PORT WAS LOOSE

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHTL NICODEMO

Signature and Printed Name

01/26/2024
Date

Stability Checks

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.049	12:22
Air Blank	0.000	12:23
Control Test	0.050	12:23
Air Blank	0.000	12:24
Control Test	0.048	12:25
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.079	12:32
Air Blank	AMB*	12:33
Air Blank	0.000	12:34

*Ambient Fail

wet #1

Cleared area and
repeated test

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:39
Control Test	0.079	12:40
Air Blank	0.000	12:40
Control Test	0.079	12:41
Air Blank	0.000	12:42
Control Test	0.079	12:42
Air Blank	0.000	12:43
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet #2

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:44
Control Test	0.197	12:45
Air Blank	0.000	12:45
Control Test	0.196	12:46
Air Blank	0.000	12:47
Control Test	0.196	12:47
Air Blank	0.000	12:48
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:49
Control Test	0.083	12:49
Air Blank	0.000	12:50
Control Test	0.083	12:50
Air Blank	0.000	12:51
Control Test	0.082	12:51
Air Blank	0.000	12:51
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel Std Dev(%)	0.6984	

Operator's Signature

Dry

Optical Bench Calibration Adjustment

```

<<<< CHANNEL 2 >>>>
Sample      % A95      (% Abs Ref)
Sample #1 = 7.4370  (0.0450)
Sample #2 = 9.3550  (0.0400)
Sample #3 = 9.0190  (0.0000)
Sample #4 = 9.2190  (0.0710)
Avg % A95 = 9.2943  (0.0370)
STD DEV = 0.0692  (0.0356)
REL STD DEV = 0.744  (96.2045)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.4040   (-0.0130)
Sample #2 = 3.3540   (0.0020)
Sample #3 = 3.4040   (0.0210)
Sample #4 = 3.3810   (0.0540)
Avg % Abs = 3.3935   (0.0257)
STD DEV = 0.0115   (0.0253)
REL STD DEV = 0.340   (0.02515)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.4100      (-0.00430)
Sample #2 = 3.3650      (-0.0240)
Sample #3 = 3.4150      (-0.0150)
Sample #4 = 3.3310      (0.0050)
Avg. % Abs = 3.3717      (-0.0117)
STD DEV = 0.0421      (0.0150)
REL STD DEV = 1.249      (128.361)

```

```

<<<<< CHANNEL_2 >>>>>
% Abs      (% Abs Ref)
e
p #1 = 1.3090  (0.0720)
p #2 = 1.4520  (-0.0240)
p #3 = 1.4160  (0.0240)
p #4 = 1.4540  (0.0100)
Abs5 = 1.4440  (0.0033)
U = 0.0245  (0.0247)
DO DEU = 1.702  (740.560)

```

UGH CO. SO
er - Alcohol Analyzer
SN 80-003368
12:08:13
bration
Res Value = 57
e Res Value = 51

HILL	Auto
Intro	Max
Mode	Auto
01/2	

```
***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.139
Std Dev = 0.02 Rel Std Dev = 51.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.745
Std Dev = 0.03 Rel Std Dev = 4.31
Sol Val = 0.4782 mg/l or 0.100 g/210L
```

```
Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12510, Sum to = 13624
<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 3.2530      (0.1560)
Sample #2 = 3.3030      (0.0220)
Sample #3 = 3.2470      (0.1187)
```

```
Sol Value = 0.100 g/210L ***
Fit value = 0.4762 ng/L %%%
Samples Taken = 1
Sum lo = 12556, Sum lo = 13654
<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 1.7250 (0.0870)
Sample #2 = 1.6350 (0.2410)
Sample #3 = 1.5060 (0.2670)
```

```

value = 0.040 g/210L ***
value = 0.1905 mg/L %%%
Gas Taken = 4, Discarded = 1
Sum To = 13649
<<<<< CHANNEL 1 >>>>
e      % Abs      (% Abs Ref)
#1 = 0.7750      (-0.0370)
#2 = 0.7240      (-0.0710)
#3 = 0.7290      (-0.0310)

```

```

= 0.000 g/210L ****
= 0.0000 mg/l %22%
aken = 4, Discarded = 1
12628, sum_io = 13692
< CHANNEL 1 >>>>
% Abs (% Abs Ref)
= -0.0090 (0.0240)
= -0.0570 (0.0850)
= -0.0440 (0.2090)
= -0.0170 (0.2410)

```

Scd	Fit	Samp	Sum	Samp	Samp	Samp	-Samp
1	1	1	1	1	1	1	1
2	2	2	2	2	2	2	2
3	3	3	3	3	3	3	3
4	4	4	4	4	4	4	4
5	5	5	5	5	5	5	5
6	6	6	6	6	6	6	6
7	7	7	7	7	7	7	7
8	8	8	8	8	8	8	8
9	9	9	9	9	9	9	9
10	10	10	10	10	10	10	10
11	11	11	11	11	11	11	11
12	12	12	12	12	12	12	12
13	13	13	13	13	13	13	13
14	14	14	14	14	14	14	14
15	15	15	15	15	15	15	15
16	16	16	16	16	16	16	16
17	17	17	17	17	17	17	17
18	18	18	18	18	18	18	18
19	19	19	19	19	19	19	19
20	20	20	20	20	20	20	20
21	21	21	21	21	21	21	21
22	22	22	22	22	22	22	22
23	23	23	23	23	23	23	23
24	24	24	24	24	24	24	24
25	25	25	25	25	25	25	25
26	26	26	26	26	26	26	26
27	27	27	27	27	27	27	27
28	28	28	28	28	28	28	28
29	29	29	29	29	29	29	29
30	30	30	30	30	30	30	30
31	31	31	31	31	31	31	31
32	32	32	32	32	32	32	32
33	33	33	33	33	33	33	33
34	34	34	34	34	34	34	34
35	35	35	35	35	35	35	35
36	36	36	36	36	36	36	36
37	37	37	37	37	37	37	37
38	38	38	38	38	38	38	38
39	39	39	39	39	39	39	39
40	40	40	40	40	40	40	40
41	41	41	41	41	41	41	41
42	42	42	42	42	42	42	42
43	43	43	43	43	43	43	43
44	44	44	44	44	44	44	44
45	45	45	45	45	45	45	45
46	46	46	46	46	46	46	46
47	47	47	47	47	47	47	47
48	48	48	48	48	48	48	48
49	49	49	49	49	49	49	49
50	50	50	50	50	50	50	50
51	51	51	51	51	51	51	51
52	52	52	52	52	52	52	52
53	53	53	53	53	53	53	53
54	54	54	54	54	54	54	54
55	55	55	55	55	55	55	55
56	56	56	56	56	56	56	56
57	57	57	57	57	57	57	57
58	58	58	58	58</			

% A95 = 1.717
Std Dev = 0.03 Rel Std Dev = 1.89
Sol Val = 0.9524 mg/L or 0.200 g/210L
% A95 = 3.279
Std Dev = 0.03 Rel Std Dev = 0.88
Sol Val = 1.4296 mg/L or 0.300 g/210L
% A95 = 4.715
Std Dev = 0.12 Rel Std Dev = 2.52
Zero Order Coef = -116.15
First Order Coef = 2693.27

```

Sample #4 = 3.2870      (0.0300)
Avg % Abs = 3.2791      (0.0567)
STD DEV = 0.0288      (0.0533)
REL STD DEV = 0.880      (94.000)
-----
<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 6.5060      (0.0190)
Sample #2 = 6.4610      (0.0610)

```

```

Sample #1 = 1.6240 (0.3270)
Avg % Abs = 1.5863 (0.2783)
STD DEV = 0.0715 (0.0441)
REL STD DEV = 4.502 (15.846)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 3.4330 (0.0670)
Sample #2 = 3.3750 (0.1760)

```

```

R#4 = 0.7620 (-0.0080)
RDS = 0.7450 (-0.0313)
CU = 0.0321 (0.1395)
DO DEU = 4.314 (126.067)
-----
***** CHANNEL 2 *****
e % RDS (% Abs Ref)
R#1 = 1.4900 (-0.0110)
R#2 = 1.4590 (-0.0250)

```

```

= 0.0393 (0.1783)
0.0204 (0.0824)
EU = 51.875 (46.204)
-----
< CHANNEL 2 >>>>
      % Abs      (% Abs Ref)
= 0.0870 (0.0000)
= 0.1370 (0.0270)
= 0.1430 (0.0730)
= 0.1130

```

Aug	STD	REL	Sam
			Sam
			Sam
			Sam
			Sam

```

Second Order Coef = 76.40
Standard Deviation = 21.29738
-----
<<<< CHANNEL 2 >>>>
Sol Ua1 = 0.0000 mg/L or 0.000 g/2.0L
% Abs = 0.133
Std Dev = 0.01 Rel Std Dev = 9.84
Sol Ua1 = 0.1935 mg/L or 0.040 g/2.0L
% Abs = 1.452
Std Dev = 0.01 Rel Std Dev = 0.57

```

Sample #3 = 6.4200 (0.0700)
Sample #4 = 6.4160 (0.0790)
Avg % Abs = 6.4290 (0.0823)
STD DEV = 0.0286 (0.0232)
REL STD DEV = 0.445 (28.154)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/L %%%
Samples Taken = 4, Discarded = 1

```

Sample #3 = 3.4370 (0.1610)
Sample #4 = 3.3970 (0.2230)
Avg % APS = 3.3263 (0.1933)
STD DEV = 0.1424 (0.0258)
REL STD DEV = 1.276 (13.352)
-----
Sol value = 0.100 g/21.0L ***
Tit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 1.2522 g/mL = 13632

```

AD5 = 1.4530	(0.0050)
AD4 = 1.4550	(0.0410)
AD5 = 1.4523	(0.0050)
ADU = 0.0083	(0.0338)
AD DEU = 0.573	(675.574)

Value = 0.100	g/210L ***
Value = 0.4762	ng/l %22%
As Taken = 4,	Discarded = 1
12538	91m 10 = 13639

```

= 0.1180 (0.1030)
= 0.1327 (0.0677)
0.0131 (0.0383)
EU = 9.838 (56.571)
-----
= 0.040 g/21.0L ***
= 0.1905 mg/l 24%
taken = 4, Discarded = 1
12557, 9um 10 = 13652
CHANNEL >>>>

```

Samp	Sol
Aug	Fit
STD	Samp
REL	Sum

300.001 = 0.462 mg/l or 0.100 g/210L
 % ABS = 3.393
 Std Dev = 0.01 Rel Std Dev = 0.34
 Sol. Val = 0.9524 mg/l or 0.200 g/210L
 % ABS = 6.429
 Std Dev = 0.03 Rel Std Dev = 0.44
 Sol. Val = 1.4286 mg/l or 0.300 g/210L
 % ABS = 9.294
 Std Dev = 0.07 Rel Std Dev = 0.74
 Zero Order Coef = -180.50
 First Order Coef = 1473.15

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.7100  (0.1000)
Sample #2 = 4.7540  (0.0340)
Sample #3 = 4.8150  (-0.0620)
Sample #4 = 4.5820  (0.0780)
Avg % Abs = 4.7153  (0.0167)
STD DEV = 0.1189  (0.0716)
REL STD DEV = 2.520  (429.549)
-----

```

```
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
sample #1 = 1.6930    (0.0230)
sample #2 = 1.7170    (-0.0090)
sample #3 = 1.7500    (0.0060)
sample #4 = 1.6850    (0.1030)
vg % Abs = 1.7173    (0.0333)
TD DEV = 0.0325    (0.0608)
EL STD DEV = 1.893    (182.392)
=====
```

```

<<<<< CHANNEL 1 >>>>>
a      % Abs      (% Abs Ref)
p #1 = 1.5910      (-0.0360)
p #2 = 1.5610      (-0.0450)
p #3 = 1.7320      (-0.0420)
p #4 = 1.5680      (-0.0030)
Abs = 1.6537      (-0.0300)
EU = 0.0822      (0.2234)
10 DEV = 4.974      (78.102)
-----

```

% Abs	(% Abs Ref)
= 0.4680	(0.0810)
= 0.7600	(-0.0740)
= 0.6260	(0.0470)
= 0.7000	(-0.0060)
= 0.6953	(-0.0110)
0.0671	(0.0607)
EU = 9.653	(551.407)

Samp
Samp
Samp
Samp
Samp
Aug
STD
REL

Optical Bench Calibration Adjustment

80-003388
1/25/2024

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	-0.000	0.0002	
0.040	0.041	-0.0006	
0.100	0.099	0.0006	
0.200	0.200	-0.0003	
0.300	0.300	0.0001	

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.000	-0.0001	
0.040	0.040	0.0003	
0.100	0.100	-0.0002	
0.200	0.200	0.0000	
0.300	0.300	0.0000	

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 3432.00
Sample #2 = 3330.00
Sample #3 = 3383.00
Sample #4 = 3348.00
Average Result = 3353.6667
STD DEV = 26.9506
REL STD DEV = 0.804

***** CHANNEL 2
Sample #1 = 3382.00
Sample #2 = 3310.00
Sample #3 = 3342.00
Sample #4 = 3321.00
Average Result = 3324.3333
STD DEV = 16.2583
REL STD DEV = 0.489

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l*10,000) = 456
9 um H2O Adjust (mg/l*10,000) = 485
**** AUTO CAL PASS

Benjamin S. Solis

Post Calibration Adjustment Stability Checks

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:32
Control Test	0.050	16:33
Air Blank	0.000	16:34
Control Test	0.049	16:34
Air Blank	0.000	16:35
Control Test	0.049	16:36
Air Blank	0.000	16:36
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Bayan Siddiq
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:37
Control Test	0.080	16:38
Air Blank	0.000	16:39
Control Test	0.079	16:39
Air Blank	0.000	16:40
Control Test	0.078	16:41
Air Blank	0.000	16:41
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Bayan Siddiq
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:42
Control Test	0.199	16:43
Air Blank	0.000	16:43
Control Test	0.198	16:44
Air Blank	0.000	16:44
Control Test	0.200	16:45
Air Blank	0.000	16:46
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Bayan Siddiq
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:59
Control Test	0.079	16:59
Air Blank	0.000	17:00
Control Test	0.079	17:00
Air Blank	0.000	17:00
Control Test	0.080	17:01
Air Blank	0.000	17:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Bayan Siddiq
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003388, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003388</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/26/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:06</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodemo

Digitally signed by Phil Nicodemo
Date: 2024.01.26 14:32:59 -05'00'

01/26/2024

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality