



Agency ~~Mount Dora PD~~ Lake County Sheriff's Office 10/3/2024 DA s/N 80-000825

Agency ~~Mount Dora PD~~ Lake County Sheriff's Office 10/3/2024 DA S/N 80-000825

Date In 09/16/2024 DI Completion Date 09/18/2024 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

10/3/24 DA 10/15/24

Calibration Adjustment	By _____	Department Inspection	By DA _____
------------------------	----------	-----------------------	-------------

Notes/Suggested Service: Noticed that the instrument was not holding the correct time after department inspection was completed on 9/18/2024. Battery was changed on 10/2/2024 and department inspection repeated. DA 10/3/2024.

Instrument was de-registered in 2021. Lake County Sheriff's Office took possession of the instrument and the instrument had not been re-registered to LCSO. Repeated department inspection reflects new agency information. Agency name was changed in instrument. DA 10/3/24

☒	Instrument Complies with Chapter 11D-8, FAC
☐	Instrument Does Not Comply with Chapter 11D-8, FAC
☒	Return to/Place into Evidentiary Use
☐	Remain Out of Evidentiary Use
☒	Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow
Digitally signed by Taylor Gutschow
Date: 2024.10.15 13:04:25 -0400'

Shayla Platt
Digitally signed by Shayla Platt
Date: 2024.10.16 14:11:04 -0400'

Tech Review / Date
Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO

Time of Inspection: 12:33

Date of Inspection: 10/03/2024

Serial Number: 80-000825

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.078	0.199	0.078

Standard Deviations	0.0005	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

10/03/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MOUNT DORA P.D.
Time of Inspection: 16:12

Date of Inspection: 09/18/2024

Serial Number: 80-000825
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C-202304C Exp: 04/05/2025 DA 10/15/24	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
Standard Deviations	0.0005	0.0004	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

09/18/2024
Date

Stability Checks 80-000825 9/18/24 DA

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
09/18/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:56
Control Test	0.050	13:57
Air Blank	0.000	13:57
Control Test	0.049	13:58
Air Blank	0.000	13:59
Control Test	0.049	13:59
Air Blank	0.000	14:00
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
09/18/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:02
Control Test	0.079	14:03
Air Blank	0.000	14:03
Control Test	0.079	14:04
Air Blank	0.000	14:04
Control Test	0.078	14:05
Air Blank	0.000	14:06
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Wet



Operator's Signature

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
09/18/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:07
Control Test	0.198	14:08
Air Blank	0.000	14:08
Control Test	0.198	14:09
Air Blank	0.000	14:09
Control Test	0.197	14:10
Air Blank	0.000	14:11
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	



Operator's Signature

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
09/18/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:13
Control Test	0.079	14:13
Air Blank	0.000	14:14
Control Test	0.080	14:14
Air Blank	0.000	14:15
Control Test	0.079	14:15
Air Blank	0.000	14:15
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite A2025
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000825 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000825</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAKE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/03/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:33</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee
Armstrong
Digitally signed by Destinee
Armstrong
Date: 2024.10.04 12:48:53 -04'00'

10/03/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-000825

OWNING AGENCY: Lake County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 09/18/2024 10/3/2024

DA
10/15/2024

AGENCY INSPECTOR: Matthew Bowden; Jeff Desantis

ADDRESS: 32400 County Rd 473

CITY, STATE, ZIP: Leesburg, FL 34788

TELEPHONE NUMBER: 352-267-3292; 352-602-9722

FAX NUMBER: N/A

10/15/24 DA

EMAIL ADDRESS (if available): Matthew.Bowden@lcsso.org; Jeff.Desantis@lcsso.org

For Program Office Use Only:



- ☒ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database