



INSTRUMENT PROCESSING SHEET

Agency Blountstown PDS/N 80-000862Florida Department of
Law EnforcementDate In 02-29-2024 DI Completion Date 3/5/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>02-29-2024</u>		Quality Checks By <u>ALL</u> Date <u>02-29-2024</u>		Flow Calibration By _____ Date _____																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>162</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP6291</td><td>202303K 03-29-2025</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L 03-29-2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C 04-05-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901 04-19-2025</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 03-29-2025	0.080	MP6292	202303L 03-29-2025	0.200	MP6293	202304C 04-05-2025	0.080 DGS	N/A	AG310901 04-19-2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP6291	202303K 03-29-2025																																																															
0.080	MP6292	202303L 03-29-2025																																																															
0.200	MP6293	202304C 04-05-2025																																																															
0.080 DGS	N/A	AG310901 04-19-2025																																																															
Calibration Adjustment By _____				Department Inspection By <u>PN</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28662</u> Gauge <u>1014</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6289																																																																
Interferent	MP6290																																																																
0.050	MP6291																																																																
0.080	MP6292																																																																
0.200	MP6293																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				Attachments <table border="1" style="width:100%"><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></table>		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																										
				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																												
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.03.08 11:56:39 +05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2024.03.08 12:38:35 -05'00' Tech Review / Date _____ Admin Review / Date _____																																																																	

Stability checks 80-000862 02/29/24

BLOUNTSTOWN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000862
02/29/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.051	12:10
Air Blank	0.000	12:11
Control Test	0.051	12:12
Air Blank	0.000	12:12
Control Test	0.051	12:13
Air Blank	0.000	12:13
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AZ
Operator's Signature

BLOUNTSTOWN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000862
02/29/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.081	12:16
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Wet

AZ
Operator's Signature

BLOUNTSTOWN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000862
02/29/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:22
Control Test	0.203	12:23
Air Blank	0.000	12:23
Control Test	0.200	12:24
Air Blank	0.000	12:24
Control Test	0.200	12:25
Air Blank	0.000	12:25
Control Test Stats		
Average	0.2010	
Std Dev	0.0017	
Rel Std Dev(%)	0.8617	

AZ
Operator's Signature

BLOUNTSTOWN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000862
02/29/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.081	12:28
Air Blank	0.000	12:29
Control Test	0.082	12:29
Air Blank	0.000	12:30
Control Test	0.082	12:30
Air Blank	0.000	12:30
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

065

AZ
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BLOUNTSTOWN PD

Serial Number: 80-000862

Time of Inspection: 12:27

Date of Inspection: 03/05/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 06723080A5 Exp: 04/05/2025
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.079
Standard Deviations	0.0003	0.0004	0.0000	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

03/05/2024

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000862, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000862</u>	UNCERTAINTY * $\pm$	
Owning Agency:	<u>BLOUNTSTOWN PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/05/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:27</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

03/05/2024

Date


PHIL NICODEMO,
Department Inspector

Service • Integrity • Respect • Quality