

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Miami Police Department
Time of Inspection:

Date of Inspection:

Serial Number: 80-000874
Software:

| Check or Test | YES | NO | Check or Test | YES | NO |
|--|-----|----|---|-----|----|
| Diagnostic Check (Pre-Inspection): OK | | | Date and/or Time Adjusted | | |
| Minimum Sample Volume Check: OK | | | Barometric Pressure Sensor Check: OK | | |
| Alcohol Free Subject Test: 0.000 | | | Mouth Alcohol Test: Slope Not Met | | |
| Interferent Detect Test: Interferent Detect | | | Diagnostic Check (Post-Inspection): OK | | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: Exp: | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------------|--|--|--|--|
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

| Standard Deviations | | | | |
|---------------------|--|--|--|--|
| | | | | |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

As of 12/11/2024, this instrument is currently out for repair. Unable to perform 2024 Department Inspection.

The above instrument complies (☐) does not comply (☒) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Taylor D. Gutschow

Signature and Printed Name

12/11/2024

Date