



Agency Miami Dade PD

S/N 80-000881

Date In 05/28/2024 DI Completion Date 06/06/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG Date 06/04/2024

- ☒ Annual
- ☐ Registration
- ☒ Return from CMI / EE

- Case
- Keyboard
- Feet
- Ports
- Handle
- Dry Gas Shelf
- Breath Tube
- Screws Tight

- ☐ Power cord
- ☐ Printer Cable
- ☒ Static Bag
- ☐ 12V DC Cable

Notes:

Quality Checks By TDG Date 06/04/2024

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 172
☒ Flow Verification (L/s)
 Flow Column # ATP104
 32 mm 0.148 (.139 - .169)
 36 mm 0.167 (.156 - .190)
 53 mm 0.230 (.228 - .278)
 103 mm 0.488 (.447 - .547)

■ Barometric Pressure Check
Gauge ID # 26932
■ Stability Checks

Flow Calibration By_____ Date_____

Flow Column # _____

☐ 5L/min – 17mm

☐ 15L/min – 53mm

☐ 30L/min – 103mm

☐ R-Value _____

☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm	_____	(.139 - .169)
36 mm	_____	(.156 - .190)
53 mm	_____	(.228 - .278)
103 mm	_____	(.447 - .547)

Maintenance By _____ Date _____

☐ Battery Replacement

☐ Dry Gas Regulator Replacement

☐ Breath Tube Replacement

☐ Other _____

Calibration Adjustment	By _____	Department Inspection	By <u>TDG</u>
------------------------	----------	-----------------------	---------------

Barometric Pressure Gauge _____ ID # _____

☐ Post Calibration Adjustment Stability Checks

Department Inspection By TDG

Barometric Pressure ID# 26932 (x2)
 Gauge 1012 / 1011 Instrument 1011 / 1010
 Mouth Alcohol Solution Lot # 2023-A (x2)
 Acetone Stock Solution Lot # 2023-B (x2)

Attachments

☒ Form 41 (x2) ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Extra Stabilities</u>
--	--

Notes/Suggested Service: Conducted second discretionary Department Inspection after running the extra stabilities. Both inspections complied with Chapter 11D-8. (TDG)

☒ Instrument Complies with Chapter 11D-8, FAC

☐ Instrument Does Not Comply with Chapter 11D-8, FAC

☒ Return to/Place into Evidentiary Use

☐ Remain Out of Evidentiary Use

☒ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo	Digitally signed by Phil Nicodemo Date: 2024.06.11 13:32:25 -0400	Shayla Platt	by Shayla Platt Date: 2024.06.12 07:32:09 -04'00'
Tech Review / Date		Admin Review / Date	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 13:42

Date of Inspection: 06/04/2024

Serial Number: 80-000881
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined ^{MC} 06/04/2024

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

06/04/2024
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
VIEW PAGE 20 Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-000881 16/04/2024 Software: 8101.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:11	Control Test	0.049	14:11	Air Blank	0.000	14:12	Control Test	0.049	14:13	Air Blank	0.000	14:13	Control Test	0.049	14:14	Air Blank	0.000	14:14	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		VIEW PAGE 20 Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-000881 16/04/2024 Software: 8101.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0026</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:21	Control Test	0.079	14:22	Air Blank	0.000	14:23	Control Test	0.079	14:23	Air Blank	0.000	14:24	Control Test	0.078	14:25	Air Blank	0.000	14:25	Control Test Stats			Average	0.0787		Std Dev	0.0026		Rel. Std Dev(%)	0.7339		VIEW PAGE 20 Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-000881 16/04/2024 Software: 8101.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2936</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:51	Control Test	0.197	13:51	Air Blank	0.000	13:52	Control Test	0.196	13:53	Air Blank	0.000	13:53	Control Test	0.197	13:54	Air Blank	0.000	13:54	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel. Std Dev(%)	0.2936		VIEW PAGE 20 Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-000881 16/04/2024 Software: 8101.27 012 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.4314</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:58	Control Test	0.082	13:58	Air Blank	0.000	13:58	Control Test	0.080	13:59	Air Blank	0.000	13:59	Control Test	0.080	14:00	Air Blank	0.000	14:00	Control Test Stats			Average	0.0807		Std Dev	0.0012		Rel. Std Dev(%)	1.4314	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.049	14:11																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test	0.049	14:13																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test	0.049	14:14																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.079	14:22																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.079	14:23																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test	0.078	14:25																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0026																																																																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test	0.197	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.196	13:53																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test	0.197	13:54																																																																																																																																																	
Air Blank	0.000	13:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:58																																																																																																																																																	
Control Test	0.082	13:58																																																																																																																																																	
Air Blank	0.000	13:58																																																																																																																																																	
Control Test	0.080	13:59																																																																																																																																																	
Air Blank	0.000	13:59																																																																																																																																																	
Control Test	0.080	14:00																																																																																																																																																	
Air Blank	0.000	14:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel. Std Dev(%)	1.4314																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 13:35

Date of Inspection: 06/05/2024

Serial Number: 80-000881
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.049 / 0.049	0.078	0.195	0.082
0.000	0.049 / 0.049	0.078	0.195	0.083
0.000	0.049 / 0.049	0.078	0.196	0.082
0.000	0.049 / 0.049	0.078	0.196	0.082
0.000	0.049 / 0.049	0.078	0.196	0.083
0.000	0.049 / 0.049	0.078	0.195	0.082
0.000	0.055 / 0.049	0.078	0.196	0.082
0.000	0.056 / 0.050	0.078	0.195	0.082
0.000	0.055 / 0.049	0.078	0.196	0.082
0.000	0.055 / 0.049	0.078	0.196	0.082
Standard Deviations	0.0032 / 0.0003	0.0000	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

Unknown. Secured sim connection to instrument and sim power cord and repeated. Results of repeated test were within the acceptable range. Will conduct additional stabilities.

ML 06/05/2024

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

06/05/2024
Date

0.05

Extra Stabilities

MIAMI DASH PC
 Intoxilyzer - Alconol Analyzer
 Model 8000 SN 60-000891
 06/06/2024
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.049	10:01
Air Blank	0.000	10:01
Control Test	0.049	10:02
Air Blank	0.000	10:02
Control Test	0.049	10:03
Air Blank	0.000	10:04
Control Test	0.049	10:04
Air Blank	0.000	10:05
Control Test	0.049	10:06
Air Blank	0.000	10:06
Control Test	0.050	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:08
Control Test	0.050	10:09
Air Blank	0.000	10:10
Control Test	0.049	10:10
Air Blank	0.000	10:11
Control Test	0.049	10:12
Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:13
Control Test	0.049	10:14
Air Blank	0.000	10:15
Control Test	0.049	10:15
Air Blank	0.000	10:16
Control Test	0.050	10:16
Air Blank	0.000	10:17
Control Test	0.050	10:18
Air Blank	0.000	10:18
Control Test	0.049	10:19
Air Blank	0.000	10:19
Control Test	0.048	10:20
Air Blank	0.000	10:21
Control Test	0.049	10:21
Air Blank	0.000	10:22
Control Test	0.049	10:23
Air Blank	0.000	10:23
Control Test	0.049	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0492	
Std Dev	0.0005	
Rel Std Dev(%)	0.9956	

Ran extra stabilities using the same 0.05
 sim and ARS lot. Results were within the
 acceptable range.

ML
 06/06/2024

ML
 Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 14:26

Date of Inspection: 06/06/2024

Serial Number: 80-000881

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.049	0.078	0.196	0.082
0.000	0.049	0.078	0.196	0.082
0.000	0.050	0.078	0.196	0.082
0.000	0.050	0.078	0.196	0.082
0.000	0.049	0.078	0.196	0.082
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.078	0.196	0.082
0.000	0.050	0.078	0.196	0.082
0.000	0.050	0.078	0.196	0.082
0.000	0.050	0.078	0.197	0.082


Standard Deviations	0.0005	0.0000	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/06/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000881, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000881</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/06/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:26</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/06/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Miami Dade PDS/N 80-000881Florida Department of
Law EnforcementDate In 03/12/2024 DI Completion Date n/a☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>TDG</u> Date <u>03/12/2024</u>		Quality Checks By <u>TDG</u> Date <u>03/12/2024</u>		Flow Calibration By _____ Date _____																																																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Calibration Adjustment By _____				Department Inspection By _____																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tr><td>☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u></td></tr></table>		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u>
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000																																																																			
Interferent																																																																			
0.050																																																																			
0.080																																																																			
0.200																																																																			
☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u>																																																																		
Notes/Suggested Service: <u>Instrument gives DSP Fail upon start-up and will not go into Ready Mode. Sending to CMI. (TDG)</u> _____ _____ _____ _____ _____ _____ _____ _____ _____				☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use _____ Tech Review / Date Admin Review / Date																																																															

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Richard Closius on 03/12/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000881

Bill To Address:

Miami Dade PD

Attn: Richard Closius

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument gives DSP Fail upon start-up and won't go into Ready Mode.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☐ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Richard Closius

Phone #: 305-458-3991

Email: rlclosius@mdpd.com

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us