



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-000902Florida Department of
Law EnforcementDate In 03-19-2024 DI Completion Date 3-27-2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>03-19-2024</u>		Quality Checks By <u>ALL</u> Date <u>03-20-2024</u>		Flow Calibration By _____ Date _____																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>drop off, no box</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>193</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.464</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP6295</td><td>202303K 03-29-2025</td></tr><tr><td>0.080</td><td>MP6296</td><td>202303L 03-29-2025</td></tr><tr><td>0.200</td><td>MP6298</td><td>202304C 04-05-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901 04-19-2025</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6295	202303K 03-29-2025	0.080	MP6296	202303L 03-29-2025	0.200	MP6298	202304C 04-05-2025	0.080 DGS	N/A	AG310901 04-19-2025	Maintenance By <u>IS</u> Date <u>3-21-2024</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																										
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6295	202303K 03-29-2025																																											
0.080	MP6296	202303L 03-29-2025																																											
0.200	MP6298	202304C 04-05-2025																																											
0.080 DGS	N/A	AG310901 04-19-2025																																											
Calibration Adjustment By _____		Department Inspection By <u>SP</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>30793</u> Gauge <u>1010</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP6291																																												
0.080	MP6292																																												
0.200	MP6293																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 x3 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Diagnostic Check</u>																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.03.29 11:07:53 -0400 Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.04.04 13:51:51 -0400																																											
		Tech Review / Date _____ Admin Review / Date _____																																											

Stability checks 80-000902 03-20-24

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000902
03/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.051	11:30
Air Blank	0.000	11:31
Control Test	0.051	11:32
Air Blank	0.000	11:32
Control Test	0.051	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AZ
Operator's Signature

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000902
03/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:25
Control Test	0.081	11:25
Air Blank	0.000	11:26
Control Test	0.081	11:26
Air Blank	0.000	11:27
Control Test	0.081	11:28
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet
AZ
Operator's Signature

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000902
03/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:35
Control Test	0.201	11:35
Air Blank	0.000	11:36
Control Test	0.201	11:37
Air Blank	0.000	11:37
Control Test	0.201	11:38
Air Blank	0.000	11:38
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AZ
Operator's Signature

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000902
03/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.082	11:41
Air Blank	0.000	11:41
Control Test	0.082	11:42
Air Blank	0.000	11:42
Control Test	0.082	11:43
Air Blank	0.000	11:43
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

065
AZ
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA

Time of Inspection: 12:56

Date of Inspection: 03/21/2024

Serial Number: 80-000902

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted	Yes	
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Time-Date changed. TIME/DATE INCORRECT, BATTERY CHANGE

Stopped inspection, time/date were incorrect.
Battery will be changed.

[Signature]

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

03/21/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA
Time of Inspection: 13:33

Date of Inspection: 03/27/2024

Serial Number: 80-000902
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Non-compliance: COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

03/27/2024
Date

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000902
03/27/2024
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

SP

Instrument experienced diagnostic fail during initial attempt of department inspection. Ran separate diagnostic check in attempt to reproduce the issue. All Diagnostics OK.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA
Time of Inspection: 16:33

Date of Inspection: 03/27/2024

Serial Number: 80-000902
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.049	0.079	0.199	0.000 / 0.081
0.000	0.050	0.079	0.199	0.000 / 0.081
0.000	0.050	0.080	0.199	0.000 / 0.081
0.000	0.050	0.079	0.199	0.000 / 0.082
0.000	0.050	0.079	0.199	0.000 / 0.082
0.000	0.050	0.080	0.199	0.000 / 0.081
0.000	0.050	0.080	0.199	0.000 / 0.082
0.000	0.050	0.080	0.199	0.000 / 0.082
0.000	0.051	0.080	0.200	0.000 / 0.082
0.000	0.050	0.080	0.199	0.000 / 0.082

Standard Deviations	0.0004	0.0005	0.0003	0.0000 / 0.0005
---------------------	--------	--------	--------	-----------------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance CORD NOT PLUGGED IN.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SHAYLA D PLATT

Signature and Printed Name

03/27/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000902, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000902</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC TAMPA</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/27/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:33</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Shayla Platt

03/27/2024

Date

SHAYLA D PLATT,

Department Inspector

Service • Integrity • Respect • Quality