



INSTRUMENT PROCESSING SHEET

Agency Orange County SOS/N 80-000964Florida Department of
Law EnforcementDate In 06/27/2024 DI Completion Date 7/2/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>06/27/2024</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>ALL</u> Date <u>07/01/2024</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>212</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901 04/19/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 03/29/2025	0.080	MP5089	202303L 03/29/2025	0.200	MP5090	202304C 04/05/2025	0.080 DGS	N/A	AG310901 04/19/2025	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																																			
Simulator	Serial #	Lot #/Exp																																																																		
0.050	MP5088	202303K 03/29/2025																																																																		
0.080	MP5089	202303L 03/29/2025																																																																		
0.200	MP5090	202304C 04/05/2025																																																																		
0.080 DGS	N/A	AG310901 04/19/2025																																																																		
Calibration Adjustment By <u>SP</u> Barometric Pressure Gauge <u>1014</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>23400</td><td>10/24/25</td></tr><tr><td>0.100</td><td>MP6296</td><td>23450</td><td>12/5/25</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>2/13/26</td></tr><tr><td>0.300</td><td>MP6298</td><td>23410</td><td>11/1/25</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>4/5/25</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K</td><td>3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L</td><td>3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C</td><td>4/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901</td><td>4/19/2025</td></tr></tbody></table> Notes/Suggested Service: _____ Cal adjusted due to barometric pressure greater than 1% SP _____ _____ _____ _____ _____		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	23400	10/24/25	0.100	MP6296	23450	12/5/25	0.200	MP6297	24080	2/13/26	0.300	MP6298	23410	11/1/25	0.080 DGS	N/A	06723080A5	4/5/25	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202303K	3/29/2025	0.080	MP5089	202303L	3/29/2025	0.200	MP5090	202304C	4/05/2025	0.080 DGS	N/A	AG310901	4/19/2025	Department Inspection By <u>SP</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%; border-collapse: collapse;"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"><tr><td style="width:50%; text-align: center;">Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.07.09 13:46:55 +04'00'</td><td style="width:50%; text-align: center;">Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.07.12 13:00:58 -04'00'</td></tr><tr><td style="text-align: center;">Tech Review / Date _____</td><td style="text-align: center;">Admin Review / Date _____</td></tr></table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.07.09 13:46:55 +04'00'	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.07.12 13:00:58 -04'00'	Tech Review / Date _____	Admin Review / Date _____
Simulator	Serial #	Lot #	Expiration																																																																	
0.000	MP6294	N/A	N/A																																																																	
0.040	MP6295	23400	10/24/25																																																																	
0.100	MP6296	23450	12/5/25																																																																	
0.200	MP6297	24080	2/13/26																																																																	
0.300	MP6298	23410	11/1/25																																																																	
0.080 DGS	N/A	06723080A5	4/5/25																																																																	
Simulator	Serial #	Lot #	Expiration																																																																	
0.050	MP5088	202303K	3/29/2025																																																																	
0.080	MP5089	202303L	3/29/2025																																																																	
0.200	MP5090	202304C	4/05/2025																																																																	
0.080 DGS	N/A	AG310901	4/19/2025																																																																	
Simulator	Serial Number																																																																			
0.000	MP5086																																																																			
Interferent	MP5087																																																																			
0.050	MP5088																																																																			
0.080	MP5089																																																																			
0.200	MP5090																																																																			
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																																			
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.07.09 13:46:55 +04'00'	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.07.12 13:00:58 -04'00'																																																																			
Tech Review / Date _____	Admin Review / Date _____																																																																			

Stability checks 80-000964 07/01/24

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000964
07/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.050	11:10
Air Blank	0.000	11:11
Control Test	0.050	11:12
Air Blank	0.000	11:12
Control Test	0.050	11:13
Air Blank	0.000	11:13
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *AZ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000964
07/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:15
Control Test	0.079	11:16
Air Blank	0.000	11:17
Control Test	0.079	11:17
Air Blank	0.000	11:18
Control Test	0.079	11:18
Air Blank	0.000	11:18
Control Test Stats		11:19
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

Operator's Signature *AZ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000964
07/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.198	11:24
Air Blank	0.000	11:24
Control Test	0.198	11:25
Air Blank	0.000	11:25
Control Test	0.198	11:26
Air Blank	0.000	11:27
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *AZ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000964
07/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:37
Control Test	0.079	11:38
Air Blank	0.000	11:38
Control Test	0.080	11:38
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

OGS

Operator's Signature *AZ*

Post Cal Stability Checks

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.050	10:55
Air Blank	0.000	10:55
Control Test	0.049	10:56
Air Blank	0.000	10:57
Control Test	0.049	10:57
Air Blank	0.000	10:58
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SP

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.079	10:50
Air Blank	0.000	10:51
Control Test	0.078	10:52
Air Blank	0.000	10:52
Control Test	0.078	10:53
Air Blank	0.000	10:53
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

SP

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.197	11:01
Air Blank	0.000	11:01
Control Test	0.197	11:02
Air Blank	0.000	11:03
Control Test	0.197	11:03
Air Blank	0.000	11:04
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:23
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DES

SP

Operator's Signature

ORANGE COUNTY SO

Toxins Analyzer - Alcohol Analyzer

Model 8000

07/02/2024

08:17:38

Auto Calibration

Max Power Res Value = 94

Auto Range Res Value = 68

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12592, Sum Io = 13946

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.0670 (-0.0070)

Sample #2 = 0.1130 (-0.0160)

Sample #3 = 0.0750 (0.0040)

Sample #4 = 0.1270 (0.0000)

Aug % Abs = 0.1050 (-0.0040)

STD DEV = 0.0269 (0.0106)

REL STD DEV = 25.626 (264.575)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.0530 (-0.0080)

Sample #2 = 0.1010 (-0.0290)

Sample #3 = 0.0610 (-0.0210)

Sample #4 = 0.0960 (-0.0340)

Aug % Abs = 0.0860 (-0.0280)

STD DEV = 0.0218 (0.0066)

REL STD DEV = 25.342 (23.419)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12577, Sum Io = 13847

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.7940 (-0.0100)

Sample #2 = 0.8180 (0.0000)

Sample #3 = 0.8000 (0.0000)

Sample #4 = 0.7930 (0.0150)

Aug % Abs = 0.8037 (0.0050)

STD DEV = 0.0129 (0.0097)

REL STD DEV = 1.605 (173.205)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 1.4930 (0.0030)

Sample #2 = 1.5120 (0.0110)

Sample #3 = 1.5030 (0.0130)

Sample #4 = 1.4860 (0.0140)

Aug % Abs = 1.5010 (0.0127)

STD DEV = 0.0121 (0.0015)

REL STD DEV = 0.808 (12.059)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12576, Sum Io = 13847

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 1.9030 (-0.0030)

Sample #2 = 1.8580 (0.0230)

Sample #3 = 1.8600 (0.0360)

Sample #4 = 1.8710 (0.0290)

Aug % Abs = 1.8697 (0.0293)

STD DEV = 0.0111 (0.0065)

REL STD DEV = 0.592 (22.181)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 3.6520 (-0.0110)

Sample #2 = 3.6140 (0.0180)

Sample #3 = 3.5910 (0.0360)

Sample #4 = 3.5530 (0.0340)

Aug % Abs = 3.5953 (0.0293)

STD DEV = 0.0127 (0.0099)

REL STD DEV = 0.354 (33.633)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12572, Sum Io = 13844

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 3.6040 (-0.0120)

Sample #2 = 3.5500 (0.0250)

Sample #3 = 3.5320 (0.0350)

Sample #4 = 3.5330 (0.0460)

Aug % Abs = 3.5383 (0.0360)

STD DEV = 0.0101 (0.0085)

REL STD DEV = 0.286 (22.484)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 6.9200 (-0.0080)

Sample #2 = 6.8520 (0.0460)

Sample #3 = 6.8410 (0.0500)

Sample #4 = 6.8720 (0.0400)

Aug % Abs = 6.8550 (0.0453)

STD DEV = 0.0157 (0.0050)

REL STD DEV = 0.229 (11.103)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12569, Sum Io = 13842

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 5.2780 (-0.0200)

Sample #2 = 5.2520 (0.0140)

Sample #3 = 5.2200 (0.0230)

Sample #4 = 5.2320 (0.0260)

Aug % Abs = 5.2347 (0.0210)

STD DEV = 0.0162 (0.0062)

REL STD DEV = 0.309 (29.738)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 10.1020 (-0.0150)

Sample #2 = 10.0400 (0.0520)

Sample #3 = 10.0340 (0.0480)

Sample #4 = 10.0260 (0.0540)

Aug % Abs = 10.0333 (0.0513)

STD DEV = 0.0070 (0.0031)

REL STD DEV = 0.070 (5.951)

***** AUTO CAL DATA *****

***** CHANNEL 1 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.105

Std Dev = 0.03 Rel Std Dev = 25.63

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.804

Std Dev = 0.01 Rel Std Dev = 1.60

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.871

Std Dev = 0.01 Rel Std Dev = 0.59

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.538

Std Dev = 0.01 Rel Std Dev = 0.29

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.235

Std Dev = 0.02 Rel Std Dev = 1.31

Zero Order Coef = -291.91

First Order Coef = 2702.43

Second Order Coef = 16.31

Standard Deviation = 38.963562

***** CHANNEL 2 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.06

Std Dev = 0.02 Rel Std Dev = 25.34

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.501

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.599

Std Dev = 0.01 Rel Std Dev = 0.35

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.855

Std Dev = 0.02 Rel Std Dev = 0.23

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.033

Std Dev = 0.01 Rel Std Dev = 0.07

Zero Order Coef = -119.69

First Order Coef = 1326.71

Second Order Coef = 10.97

Standard Deviation = 25.504511

CAL ADJUSTMENT

88

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l x 10,000) = 445
9 um H2O Adjust (mg/l x 10,000) = 170
***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0002
0.040 0.040 0.0003
0.100 0.101 -0.0012
0.200 0.199 0.0010
0.300 0.300 -0.0003

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.040 0.040 0.0002
0.100 0.101 -0.0008
0.200 0.199 0.0007
0.300 0.300 -0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1

Sample #1 = 3398.00
Sample #2 = 3327.00
Sample #3 = 3367.00
Sample #4 = 3398.00
Average Result = 3364.0000
STD DEV = 35.5949
REL STD DEV = 1.058
***** CHANNEL 2

Sample #1 = 3630.00
Sample #2 = 3608.00
Sample #3 = 3663.00
Sample #4 = 3648.00
Average Result = 3639.6667
STD DEV = 28.4312
REL STD DEV = 0.781
***** CHANNEL 1

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l x 10,000) = 445
9 um H2O Adjust (mg/l x 10,000) = 170
***** AUTO CAL PASS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 13:00

Date of Inspection: 07/02/2024

Serial Number: 80-000964
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.079	0.197	0.080
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.051	0.079	0.198	0.079


Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


SHAYLA D PLATT
Signature and Printed Name

07/02/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000964, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000964</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/02/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:00</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Platt
Digitally signed by Shayla
Date: 2024.07.09 11:17:39
-04'00'

07/02/2024

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality