



Agency Titusville PD

S/N 80-000996

Date In 09/19/2024 DI Completion Date 10/9/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By DA

Barometric Pressure Gauge 1007/1006 ID # 28662

■ Post Calibration Adjustment Stability Checks 04/05/2025

AG310901 DA 10/15/24

Tech review: corrected expiration date for DGS used during calibration adjustment and the lot number of the DGS used during post stability checks. DA 10/15/24

Department Inspection

By DA

Barometric Pressure ID# 30793

Gauge 1008	<u>Instrument 1007</u>
------------	------------------------

Mouth Alcohol Solution Lot # 2024-A

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|--|---|
| ☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☒ Calibration Adjustment | ☒ Post-Stability Checks
☐ Flow Calibration
☒ Form 40
☐ Other |
|--|---|

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Phil
Nicodemo


Tech Review / Date

Shayla
Platt
Admin Review / Date

Stability Checks 80-000996 10/2/24 DA

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.051	08:41
Air Blank	0.000	08:42
Control Test	0.050	08:43
Air Blank	0.000	08:43
Control Test	0.050	08:44
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	




Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.079	08:47
Air Blank	0.000	08:47
Control Test	0.078	08:48
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:50
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	


Wet



Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:54
Control Test	0.198	08:54
Air Blank	0.000	08:55
Control Test	0.196	08:56
Air Blank	0.000	08:56
Control Test	0.195	08:57
Air Blank	0.000	08:58
Control Test Stats		
Average	0.1963	
Std Dev	0.0015	
Rel Std Dev(%)	0.7780	




Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.082	09:00
Air Blank	0.000	09:00
Control Test	0.083	09:01
Air Blank	0.000	09:01
Control Test	0.083	09:02
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel Std Dev(%)	0.6984	

DGS



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD

Time of Inspection: 08:27

Date of Inspection: 10/02/2024

Serial Number: 80-000996

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____


Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

NA- Didn't determine compliance DA 10/2/24

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

10/02/2024
Date

Optical Bench Calibration Adjustment #1 10/8/24 DA

TITUSVILLE PD
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000996
10/08/2024 09:36:44

Auto Calibration
Max Power Res Value = 74
Auto Range Res Value = 62
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12698, Sum Io = 13635

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0080)
Sample #2 = 0.1830 (0.0060)
Sample #3 = 0.1510 (0.0630)
Sample #4 = 0.1240 (0.0800)
Avg % Abs = 0.1527 (0.0497)
STD DEV = 0.0295 (0.0388)
REL STD DEV = 19.346 (78.040)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1670 (-0.0060)
Sample #2 = 0.1690 (0.0110)
Sample #3 = 0.1550 (0.0300)
Sample #4 = 0.1570 (0.0380)
Avg % Abs = 0.1603 (0.0263)
STD DEV = 0.0076 (0.0139)
REL STD DEV = 4.723 (52.665)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12658, Sum Io = 13610

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.9440 (-0.0120)
Sample #2 = 0.8530 (0.0550)
Sample #3 = 0.9050 (0.0500)
Sample #4 = 0.9090 (0.0700)
Avg % Abs = 0.8890 (0.0583)
STD DEV = 0.0312 (0.0104)
REL STD DEV = 3.514 (17.843)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5740 (0.0000)
Sample #2 = 1.5180 (0.0510)
Sample #3 = 1.5470 (0.0400)
Sample #4 = 1.5310 (0.0560)
Avg % Abs = 1.5320 (0.0490)
STD DEV = 0.0145 (0.0082)
REL STD DEV = 0.948 (16.705)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12643, Sum Io = 13599
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 2.0320 (-0.0060)
Sample #2 = 1.9810 (0.0590)
Sample #3 = 1.9810 (0.0830)
Sample #4 = 1.9740 (0.0980)
Avg % Abs = 1.9787 (0.0800)
STD DEV = 0.0040 (0.0197)
REL STD DEV = 0.204 (24.590)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.6310 (-0.0140)
Sample #2 = 3.5570 (0.0680)
Sample #3 = 3.5630 (0.0730)
Sample #4 = 3.5390 (0.0860)
Avg % Abs = 3.5530 (0.0757)
STD DEV = 0.0125 (0.0093)
REL STD DEV = 0.352 (12.280)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12634, Sum Io = 13596

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.9100 (-0.0160)
Sample #2 = 3.7670 (0.1030)
Sample #3 = 3.7430 (0.1350)
Sample #4 = 3.7980 (0.1270)
Avg % Abs = 3.7693 (0.1217)
STD DEV = 0.0276 (0.0167)
REL STD DEV = 0.732 (13.688)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.8910 (0.0000)
Sample #2 = 6.7430 (0.1360)
Sample #3 = 6.7380 (0.1610)
Sample #4 = 6.7440 (0.1740)
Avg % Abs = 6.7417 (0.1570)
STD DEV = 0.0032 (0.0193)
REL STD DEV = 0.048 (12.301)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12618, Sum Io = 13584

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.5850 (-0.0040)
Sample #2 = 5.4780 (0.1270)
Sample #3 = 5.4830 (0.1510)
Sample #4 = 5.5200 (0.1460)
Avg % Abs = 5.4937 (0.1413)
STD DEV = 0.0229 (0.0127)
REL STD DEV = 0.418 (8.959)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.9320 (-0.0050)
Sample #2 = 9.7690 (0.1900)
Sample #3 = 9.7560 (0.2150)
Sample #4 = 9.7790 (0.2130)
Avg % Abs = 9.7680 (0.2060)
STD DEV = 0.0115 (0.0139)
REL STD DEV = 0.118 (6.744)

***** AUTO CAL DATA *****
Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.153
Std Dev = 0.03 Rel Std Dev = 19.35
Sol Val = 0.1905 mg/l or 0.040 g/210L

Optical Bench Calibration Adjustment # 2 10/8/24 DA

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
10/08/2024 10:56:03

Auto Calibration
Max Power Res Value = 75
Auto Range Res Value = 62
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12613, Sum Io = 13588

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1090 (-0.0070)
Sample #2 = 0.1090 (0.0070)
Sample #3 = 0.1280 (0.0460)
Sample #4 = 0.1590 (0.0520)
Avg % Abs = 0.1320 (0.0517)
STD DEV = 0.0252 (0.0055)
REL STD DEV = 19.120 (10.660)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1580 (-0.0090)
Sample #2 = 0.1130 (0.0410)
Sample #3 = 0.1260 (0.0280)
Sample #4 = 0.1540 (0.0260)
Avg % Abs = 0.1310 (0.0317)
STD DEV = 0.0210 (0.0081)
REL STD DEV = 15.934 (25.720)

Sol Value = 0.041 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12602, Sum Io = 13582

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8800 (-0.0080)
Sample #2 = 0.8610 (0.0270)
Sample #3 = 0.8590 (0.0210)
Sample #4 = 0.8660 (0.0300)
Avg % Abs = 0.8620 (0.0260)
STD DEV = 0.0036 (0.0046)
REL STD DEV = 0.418 (17.625)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5380 (-0.0140)
Sample #2 = 1.4570 (0.0290)
Sample #3 = 1.5160 (0.0220)
Sample #4 = 1.5230 (0.0150)
Avg % Abs = 1.5120 (0.0220)
STD DEV = 0.0135 (0.0071)
REL STD DEV = 0.890 (31.818)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.9940 (-0.0120)
Sample #2 = 9.7820 (0.1930)
Sample #3 = 9.7510 (0.2170)
Sample #4 = 9.6990 (0.2610)
Avg % Abs = 9.7440 (0.2237)
STD DEV = 0.0419 (0.0345)
REL STD DEV = 0.430 (15.419)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.132
Std Dev = 0.03 Rel Std Dev = 19.12
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.862
Std Dev = 0.01 Rel Std Dev = 0.42
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.012
Std Dev = 0.00 Rel Std Dev = 0.18
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.758
Std Dev = 0.03 Rel Std Dev = 0.92
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.464
Std Dev = 0.02 Rel Std Dev = 0.41
Zero Order Coef = -306.12
First Order Coef = 2470.88
Second Order Coef = 36.78
Standard Deviation = 41.065052

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.131
Std Dev = 0.02 Rel Std Dev = 15.99
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.512
Std Dev = 0.01 Rel Std Dev = 0.89
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.569
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.727
Std Dev = 0.05 Rel Std Dev = 0.71
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.744
Std Dev = 0.04 Rel Std Dev = 0.43
Zero Order Coef = -170.34
First Order Coef = 1335.73
Second Order Coef = 15.23
Standard Deviation = 20.856242

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12599, Sum Io = 13580

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 2.0750 (-0.0100)
Sample #2 = 2.0080 (0.0710)
Sample #3 = 2.0150 (0.0580)
Sample #4 = 2.0130 (0.0740)
Avg % Abs = 2.0120 (0.0677)
STD DEV = 0.0036 (0.0085)
REL STD DEV = 0.179 (12.569)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6350 (-0.0030)
Sample #2 = 3.5640 (0.0810)
Sample #3 = 3.5750 (0.0700)
Sample #4 = 3.5690 (0.0780)
Avg % Abs = 3.5693 (0.0763)
STD DEV = 0.0055 (0.0057)
REL STD DEV = 0.154 (7.449)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12592, Sum Io = 13577

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.8760 (-0.0170)
Sample #2 = 3.7970 (0.0700)
Sample #3 = 3.7430 (0.1080)
Sample #4 = 3.7330 (0.1260)
Avg % Abs = 3.7577 (0.1013)
STD DEV = 0.0344 (0.0286)
REL STD DEV = 0.916 (28.213)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.8570 (0.0070)
Sample #2 = 6.7750 (0.1260)
Sample #3 = 6.7260 (0.1600)
Sample #4 = 6.6790 (0.1720)
Avg % Abs = 6.7267 (0.1527)
STD DEV = 0.0480 (0.0239)
REL STD DEV = 0.714 (15.629)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12587, Sum Io = 13573
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.6230 (-0.0200)
Sample #2 = 5.4840 (0.1010)
Sample #3 = 5.4670 (0.1180)
Sample #4 = 5.4400 (0.1530)
Avg % Abs = 5.4637 (0.1240)
STD DEV = 0.0222 (0.0265)
REL STD DEV = 0.406 (21.382)

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.039	0.0011
0.100	0.101	-0.0011
0.200	0.199	0.0005
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0004
0.100	0.101	-0.0006
0.200	0.200	0.0004
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3041.00
 Sample #2 = 3111.00
 Sample #3 = 3089.00
 Sample #4 = 3134.00
 Average Result = 3111.3333
 STD DEV = 22.5019
 REL STD DEV = 0.723

***** CHANNEL 2

Sample #1 = 3343.00
 Sample #2 = 3369.00
 Sample #3 = 3369.00
 Sample #4 = 3390.00
 Average Result = 3376.0000
 STD DEV = 12.1244
 REL STD DEV = 0.359

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1006
 3 um H2O Adjust (mg/l*10,000) = 698
 9 um H2O Adjust (mg/l*10,000) = 433

**** AUTO CAL PASS

Post-Stability Checks 80-000996 10/8/24 DA

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000996
10/08/2024
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.051	11:45
Air Blank	0.000	11:45
Control Test	0.050	11:46
Air Blank	0.000	11:47
Control Test	0.050	11:47
Air Blank	0.000	11:48
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	


Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000996
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:50
Control Test	0.081	11:50
Air Blank	0.000	11:51
Control Test	0.080	11:52
Air Blank	0.000	11:52
Control Test	0.080	11:53
Air Blank	0.000	11:53
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Wet


Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000996
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:55
Control Test	0.213	11:56
Air Blank	0.000	11:57
Control Test	0.210	11:57
Air Blank	0.000	11:58
Control Test	0.200	11:59
Air Blank	0.000	11:59
Control Test Stats		
Average	0.2010	
Std Dev	0.0017	
Rel Std Dev(%)	0.8617	


Operator's Signature

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000996
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.079	12:03
Air Blank	0.000	12:04
Control Test	0.079	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:05
Air Blank	0.000	12:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 13:43

Date of Inspection: 10/09/2024

Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.079	0.201	0.081
0.000	0.051	0.079	0.202	0.080
0.000	0.051	0.080	0.202	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.079	0.202	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.202	0.080
0.000	0.052	0.079	0.202	0.080
0.000	0.051	0.080	0.202	0.081
0.000	0.051	0.080	0.202	0.080
Standard Deviations	0.0004	0.0005	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DESTINEE N ARMSTRONG

Signature and Printed Name

10/09/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000996 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000996</u>	Serial Number:	<u>80-000996</u>
Owning Agency:	<u>TITUSVILLE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/09/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee
Armstrong

Digitally signed by Destinee
Armstrong
Date: 2024.10.09 14:16:11
+0400

10/09/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality