



INSTRUMENT PROCESSING SHEET

Agency Coconut Creek PDS/N 80-001044Florida Department of
Law EnforcementDate In 05/30/2024 DI Completion Date 06/05/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>06/04/2024</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>TDG</u> Date <u>06/04/2024</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>260</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203 08/10/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG222203 08/10/2024	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP6286	202303K 03/29/2025																																																
0.080	MP6287	202303L 03/29/2025																																																
0.200	MP6288	202304C 04/05/2025																																																
0.080 DGS	N/A	AG222203 08/10/2024																																																
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection By <u>TDG</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1012</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table border="1" style="width:100%"><tr><td>Phil Nicodem Digitally signed by Phil Nicodem Date: 2024.06.07 10:36:43 +04'00'</td><td>Shayla Platt Digitally signed by Shayla Platt Date: 2024.06.13 12:10:18 -04'00'</td></tr></table> Tech Review / Date _____ Admin Review / Date _____			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	Phil Nicodem Digitally signed by Phil Nicodem Date: 2024.06.07 10:36:43 +04'00'	Shayla Platt Digitally signed by Shayla Platt Date: 2024.06.13 12:10:18 -04'00'																																
Simulator	Serial Number																																																	
0.000	MP6284																																																	
Interferent	MP6285																																																	
0.050	MP6286																																																	
0.080	MP6287																																																	
0.200	MP6288																																																	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																	
Phil Nicodem Digitally signed by Phil Nicodem Date: 2024.06.07 10:36:43 +04'00'	Shayla Platt Digitally signed by Shayla Platt Date: 2024.06.13 12:10:18 -04'00'																																																	

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
COCONUT CREEK PD Intoxilyzer - Alcolizer Model 8000 06/04/2024 Software: 8100.27 SN 80-001044 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:01	Control Test	0.050	14:01	Air Blank	0.000	14:02	Control Test	0.050	14:02	Air Blank	0.000	14:03	Control Test	0.050	14:04	Air Blank	0.000	14:04	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COCONUT CREEK PD Intoxilyzer - Alcolizer Model 8000 06/04/2024 Software: 8100.27 SN 80-001044 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:56</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:57</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:54	Control Test	0.080	13:55	Air Blank	0.000	13:56	Control Test	0.080	13:56	Air Blank	0.000	13:57	Control Test	0.080	13:57	Air Blank	0.000	13:58	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COCONUT CREEK PD Intoxilyzer - Alcolizer Model 8000 06/04/2024 Software: 8100.27 SN 80-001044 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:12	Control Test	0.199	14:13	Air Blank	0.000	14:14	Control Test	0.199	14:14	Air Blank	0.000	14:15	Control Test	0.199	14:16	Air Blank	0.000	14:16	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COCONUT CREEK PD Intoxilyzer - Alcolizer Model 8000 06/04/2024 Software: 8100.27 SN 80-001044 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:51	Control Test	0.080	13:52	Air Blank	0.000	13:52	Control Test	0.081	13:52	Air Blank	0.000	13:53	Control Test	0.081	13:53	Air Blank	0.000	13:54	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Air Blank	0.000	14:01																																																																																																																																					
Control Test	0.050	14:01																																																																																																																																					
Air Blank	0.000	14:02																																																																																																																																					
Control Test	0.050	14:02																																																																																																																																					
Air Blank	0.000	14:03																																																																																																																																					
Control Test	0.050	14:04																																																																																																																																					
Air Blank	0.000	14:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0500																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:54																																																																																																																																					
Control Test	0.080	13:55																																																																																																																																					
Air Blank	0.000	13:56																																																																																																																																					
Control Test	0.080	13:56																																																																																																																																					
Air Blank	0.000	13:57																																																																																																																																					
Control Test	0.080	13:57																																																																																																																																					
Air Blank	0.000	13:58																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:12																																																																																																																																					
Control Test	0.199	14:13																																																																																																																																					
Air Blank	0.000	14:14																																																																																																																																					
Control Test	0.199	14:14																																																																																																																																					
Air Blank	0.000	14:15																																																																																																																																					
Control Test	0.199	14:16																																																																																																																																					
Air Blank	0.000	14:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1990																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:51																																																																																																																																					
Control Test	0.080	13:52																																																																																																																																					
Air Blank	0.000	13:52																																																																																																																																					
Control Test	0.081	13:52																																																																																																																																					
Air Blank	0.000	13:53																																																																																																																																					
Control Test	0.081	13:53																																																																																																																																					
Air Blank	0.000	13:54																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0807																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7157																																																																																																																																						

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Serial Number: 80-001044
Software: 8100.27

FDLE/ATP Form 41 – Revised August 2005



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001044, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001044</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>COCONUT CREEK PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/05/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:12</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/05/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality