



INSTRUMENT PROCESSING SHEET

Agency Coral Springs PDS/N 80-001050Florida Department of
Law EnforcementDate In 07/03/2024 DI Completion Date 07/09/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ Date <u>07/08/2024</u>		Quality Checks By TDG _____ Date <u>07/09/2024</u>		Flow Calibration By _____ Date _____																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Missing right middle foot</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>178</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203 08/10/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG222203 08/10/2024	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																														
Simulator	Serial #	Lot #/Exp																																															
0.050	MP6286	202303K 03/29/2025																																															
0.080	MP6287	202303L 03/29/2025																																															
0.200	MP6288	202304C 04/05/2025																																															
0.080 DGS	N/A	AG222203 08/10/2024																																															
Calibration Adjustment By _____		Department Inspection By TDG _____																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																														
0.000		N/A	N/A																																														
0.040																																																	
0.100																																																	
0.200																																																	
0.300																																																	
0.080 DGS	N/A																																																
Simulator	Serial #	Lot #	Expiration																																														
0.050																																																	
0.080																																																	
0.200																																																	
0.080 DGS	N/A																																																

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																														
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																														
✓ CORAL SPRINGS PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001050 07/09/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	11:17	Control Test	0.050	11:17	Air Blank	0.000	11:18	Control Test	0.050	11:19	Air Blank	0.000	11:19	Control Test	0.050	11:20	Air Blank	0.000	11:21	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ CORAL SPRINGS PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001050 07/09/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	11:35	Control Test	0.078	11:40	Air Blank	0.000	11:40	Control Test	0.078	11:41	Air Blank	0.000	11:42	Control Test	0.079	11:42	Air Blank	0.000	11:43	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		✓ CORAL SPRINGS PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001050 07/09/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.195</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.195</td><td>11:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7780</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	11:31	Control Test	0.198	11:32	Air Blank	0.000	11:33	Control Test	0.195	11:33	Air Blank	0.000	11:34	Control Test	0.195	11:35	Air Blank	0.000	11:35	Control Test Stats			Average	0.1963		Std Dev	0.0015		Rel Std Dev(%)	0.7780		✓ CORAL SPRINGS PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001050 07/09/2024 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr></table>	Air Blank	0.000	10:29	Control Test	0.080	10:29	Air Blank	0.000	10:29	Control Test	0.079	10:30	Air Blank	0.000	10:30	Control Test	0.080	10:31	Air Blank	0.000	10:31	Control Test Stats			Average	0.0797	
Air Blank	0.000	11:17																																																																																																																															
Control Test	0.050	11:17																																																																																																																															
Air Blank	0.000	11:18																																																																																																																															
Control Test	0.050	11:19																																																																																																																															
Air Blank	0.000	11:19																																																																																																																															
Control Test	0.050	11:20																																																																																																																															
Air Blank	0.000	11:21																																																																																																																															
Control Test Stats																																																																																																																																	
Average	0.0500																																																																																																																																
Std Dev	0.0000																																																																																																																																
Rel Std Dev(%)	0.0000																																																																																																																																
Air Blank	0.000	11:35																																																																																																																															
Control Test	0.078	11:40																																																																																																																															
Air Blank	0.000	11:40																																																																																																																															
Control Test	0.078	11:41																																																																																																																															
Air Blank	0.000	11:42																																																																																																																															
Control Test	0.079	11:42																																																																																																																															
Air Blank	0.000	11:43																																																																																																																															
Control Test Stats																																																																																																																																	
Average	0.0783																																																																																																																																
Std Dev	0.0006																																																																																																																																
Rel Std Dev(%)	0.7370																																																																																																																																
Air Blank	0.000	11:31																																																																																																																															
Control Test	0.198	11:32																																																																																																																															
Air Blank	0.000	11:33																																																																																																																															
Control Test	0.195	11:33																																																																																																																															
Air Blank	0.000	11:34																																																																																																																															
Control Test	0.195	11:35																																																																																																																															
Air Blank	0.000	11:35																																																																																																																															
Control Test Stats																																																																																																																																	
Average	0.1963																																																																																																																																
Std Dev	0.0015																																																																																																																																
Rel Std Dev(%)	0.7780																																																																																																																																
Air Blank	0.000	10:29																																																																																																																															
Control Test	0.080	10:29																																																																																																																															
Air Blank	0.000	10:29																																																																																																																															
Control Test	0.079	10:30																																																																																																																															
Air Blank	0.000	10:30																																																																																																																															
Control Test	0.080	10:31																																																																																																																															
Air Blank	0.000	10:31																																																																																																																															
Control Test Stats																																																																																																																																	
Average	0.0797																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL SPRINGS PD
Time of Inspection: 14:27

Date of Inspection: 07/09/2024

Serial Number: 80-001050
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.079

Standard Deviations	0.0000	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001050, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001050	UNCERTAINTY* $\pm$
Owning Agency:	CORAL SPRINGS PD	0.050 g/ 210 L 0.004
Calibration Date:	07/09/2024	0.080 g/ 210 L 0.004
Calibration Time:	14:27	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

07/09/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

Service • Integrity • Respect • Quality