



INSTRUMENT PROCESSING SHEET

Agency Hollywood PDS/N 80-001063Florida Department of
Law EnforcementDate In 04/11/2024 DI Completion Date 04/17/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ Date <u>04/11/2024</u>		Quality Checks By TDG _____ Date <u>04/12/2024</u>		Flow Calibration By _____ Date _____																													
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off. No box.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____														
Simulator	Serial #	Lot #/Exp																															
0.050	MP6286	202303K 03/29/2025																															
0.080	MP6287	202303L 03/29/2025																															
0.200	MP6288	202304C 04/05/2025																															
0.080 DGS	N/A	01923080A3 02/05/2025																															
Calibration Adjustment By TDG _____				Department Inspection By TDG _____																													
Barometric Pressure Gauge <u>1021</u> ID # <u>28199</u>				Barometric Pressure ID# <u>68639</u>																													
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>23390</td><td>10/17/2025</td></tr><tr><td>0.200</td><td>MP5100</td><td>23340</td><td>09/18/2025</td></tr><tr><td>0.300</td><td>MP5101</td><td>23070</td><td>03/06/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	23390	10/17/2025	0.200	MP5100	23340	09/18/2025	0.300	MP5101	23070	03/06/2025	0.080 DGS	N/A	06723080A5	04/05/2025	Gauge <u>1020</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>	
Simulator	Serial #	Lot #	Expiration																														
0.000	MP5097	N/A	N/A																														
0.040	MP5098	23400	10/24/2025																														
0.100	MP5099	23390	10/17/2025																														
0.200	MP5100	23340	09/18/2025																														
0.300	MP5101	23070	03/06/2025																														
0.080 DGS	N/A	06723080A5	04/05/2025																														
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K</td><td>03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L</td><td>03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG222203	08/10/2024	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____									
Simulator	Serial #	Lot #	Expiration																														
0.050	MP6286	202303K	03/29/2025																														
0.080	MP6287	202303L	03/29/2025																														
0.200	MP6288	202304C	04/05/2025																														
0.080 DGS	N/A	AG222203	08/10/2024																														
Notes/Suggested Service: _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																													
Shayla Platt Digitally signed by Shayla Platt Date: 2024.04.18 16:02:32 -04'00'				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.04.19 07:18:33 -04'00'																													
Tech Review / Date _____				Admin Review / Date _____																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
<p>✓</p> <p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/12/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.047</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	11:30	Control Test	0.049	11:31	Air Blank	0.000	11:31	Control Test	0.048	11:32	Air Blank	0.000	11:32	Control Test	0.049	11:33	Air Blank	0.000	11:34	Control Test Stats			Average	0.047		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>✓</p> <p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/12/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	11:37	Control Test	0.077	11:38	Air Blank	0.000	11:38	Control Test	0.077	11:39	Air Blank	0.000	11:39	Control Test	0.077	11:40	Air Blank	0.000	11:41	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>✓</p> <p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/12/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.194</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5911</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	11:44	Control Test	0.194	11:44	Air Blank	0.000	11:45	Control Test	0.196	11:45	Air Blank	0.000	11:46	Control Test	0.196	11:47	Air Blank	0.000	11:47	Control Test Stats			Average	0.1953		Std Dev	0.0012		Rel Std Dev(%)	0.5911		<p>065</p> <p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/12/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	11:50	Control Test	0.080	11:50	Air Blank	0.000	11:50	Control Test	0.080	11:51	Air Blank	0.000	11:51	Control Test	0.081	11:52	Air Blank	0.000	11:52	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.049	11:31																																																																																																																																					
Air Blank	0.000	11:31																																																																																																																																					
Control Test	0.048	11:32																																																																																																																																					
Air Blank	0.000	11:32																																																																																																																																					
Control Test	0.049	11:33																																																																																																																																					
Air Blank	0.000	11:34																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.047																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1863																																																																																																																																						
Air Blank	0.000	11:37																																																																																																																																					
Control Test	0.077	11:38																																																																																																																																					
Air Blank	0.000	11:38																																																																																																																																					
Control Test	0.077	11:39																																																																																																																																					
Air Blank	0.000	11:39																																																																																																																																					
Control Test	0.077	11:40																																																																																																																																					
Air Blank	0.000	11:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0770																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:44																																																																																																																																					
Control Test	0.194	11:44																																																																																																																																					
Air Blank	0.000	11:45																																																																																																																																					
Control Test	0.196	11:45																																																																																																																																					
Air Blank	0.000	11:46																																																																																																																																					
Control Test	0.196	11:47																																																																																																																																					
Air Blank	0.000	11:47																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1953																																																																																																																																						
Std Dev	0.0012																																																																																																																																						
Rel Std Dev(%)	0.5911																																																																																																																																						
Air Blank	0.000	11:50																																																																																																																																					
Control Test	0.080	11:50																																																																																																																																					
Air Blank	0.000	11:50																																																																																																																																					
Control Test	0.080	11:51																																																																																																																																					
Air Blank	0.000	11:51																																																																																																																																					
Control Test	0.081	11:52																																																																																																																																					
Air Blank	0.000	11:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0803																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7187																																																																																																																																						

CHANEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5310 (-0.0180)
Sample #2 = 1.5520 (-0.0200)
Sample #3 = 1.5310 (-0.0110)
Sample #4 = 1.5440 (-0.0070)
Avg % Abs = 1.5423 (-0.0127)
STD DEV = 0.0106 (0.0067)
REL STD DEV = 0.687 (52.566)

JLLYWOOD PD
Toxilogizer - Alcohol Analyzer
Jodel 8000
4/17/2024
SN 80-001063
10:34:08

Jto Calibration
Avg Power Res Value = 80
Jto Range Res Value = 47
Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12607, Sum Io = 13698
CHANEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9640 (-0.0280)
Sample #2 = 2.0140 (-0.0250)
Sample #3 = 1.9860 (-0.0100)
Sample #4 = 1.9610 (0.0310)
Avg % Abs = 1.9870 (-0.0013)
STD DEV = 0.0265 (0.0290)
REL STD DEV = 1.334 (2174.137)

CHANEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6800 (-0.0040)
Sample #2 = 3.7000 (-0.0140)
Sample #3 = 3.6790 (-0.0070)
Sample #4 = 3.6650 (0.0140)
Avg % Abs = 3.6813 (-0.0023)
STD DEV = 0.0176 (0.0146)
REL STD DEV = 0.479 (524.500)
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12533, Sum Io = 13660
CHANEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7550 (-0.0050)
Sample #2 = 3.7670 (0.0300)
Sample #3 = 3.6870 (0.0470)
Sample #4 = 3.7680 (0.0260)
Avg % Abs = 3.7407 (0.0343)
STD DEV = 0.0465 (0.0112)
REL STD DEV = 1.243 (32.477)

CHANEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8390 (-0.0100)
Sample #2 = 0.8870 (-0.0050)
Sample #3 = 0.8630 (0.0190)
Sample #4 = 0.8570 (0.0540)
Avg % Abs = 0.8650 (0.0227)
STD DEV = 0.0159 (0.0297)
REL STD DEV = 1.827 (130.899)

CHANEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.0180 (-0.00300)
Sample #2 = 7.0100 (0.0140)
Sample #3 = 6.8730 (0.0230)
Sample #4 = 7.0260 (0.0060)
Avg % Abs = 6.9697 (0.0143)
STD DEV = 0.0841 (0.0085)
REL STD DEV = 1.207 (59.337)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12512, Sum Io = 13649
CHANEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.5620 (-0.0220)
Sample #2 = 5.5670 (0.00300)
Sample #3 = 5.5700 (-0.00800)
Sample #4 = 5.5620 (0.00000)
Avg % Abs = 5.5663 (-0.0017)
STD DEV = 0.0040 (0.0057)
REL STD DEV = 0.073 (341.174)

CHANEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.2830 (-0.0120)
Sample #2 = 10.3130 (0.0070)
Sample #3 = 10.3030 (-0.0020)
Sample #4 = 10.3030 (0.0000)
Avg % Abs = 10.3063 (0.0017)
STD DEV = 0.0058 (0.0047)
REL STD DEV = 0.056 (283.549)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12533, Sum Io = 13660
CHANEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7550 (-0.0050)
Sample #2 = 3.7670 (0.0300)
Sample #3 = 3.6870 (0.0470)
Sample #4 = 3.7680 (0.0260)
Avg % Abs = 3.7407 (0.0343)
STD DEV = 0.0465 (0.0112)
REL STD DEV = 1.243 (32.477)

Optical Calibration Adjustment

By: TDG

AUTO CAL DATA *****
CHANEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.122
Std Dev = 0.02 Rel Std Dev = 14.22
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.869
Std Dev = 0.02 Rel Std Dev = 1.83
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.987
Std Dev = 0.03 Rel Std Dev = 1.33
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.741
Std Dev = 0.05 Rel Std Dev = 1.24
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.566
Std Dev = 0.00 Rel Std Dev = 0.07
Zero Order Coef = -343.02
First Order Coef = 2588.24
Second Order Coef = 7.94
Standard Deviation = 53.970482

CHANEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.112
Std Dev = 0.00 Rel Std Dev = 3.37
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.542
Std Dev = 0.01 Rel Std Dev = 0.69
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.681
Std Dev = 0.02 Rel Std Dev = 0.48
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.970
Std Dev = 0.08 Rel Std Dev = 1.21
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.306
Std Dev = 0.01 Rel Std Dev = 0.06
Zero Order Coef = -168.55
First Order Coef = 1331.38
Second Order Coef = 7.10
Standard Deviation = 49.928169

CHANEL 1 >>>>
Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3030.00
Sample #2 = 3069.00
Sample #3 = 3055.00
Sample #4 = 3032.00
Average Result = 3052.0000
STD DEV = 18.6815
REL STD DEV = 0.612





***** CHANNEL 2 *****
Sample #1 = 3423.00
Sample #2 = 3422.00
Sample #3 = 3450.00
Sample #4 = 3449.00
Average Result = 3440.3333
STD DEV = 15.8850
REL STD DEV = 0.462

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1021
3 um H2O Adjust (mg/l*10,000) = 757
9 um H2O Adjust (mg/l*10,000) = 369
***** AUTO CAL PASS *****

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.0001 0.0006
0.040 0.040 -0.0002
0.100 0.101 -0.0015
0.200 0.198 0.0016
0.300 0.301 -0.0005

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0004
0.040 0.040 0.0001
0.100 0.101 -0.0014
0.200 0.199 0.0014
0.300 0.300 -0.0005

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/17/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:56</p> <p>Control Test 0.050 11:57</p> <p>Air Blank 0.000 11:57</p> <p>Control Test 0.049 11:58</p> <p>Air Blank 0.000 11:58</p> <p>Control Test 0.050 11:59</p> <p>Air Blank 0.000 12:00</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p> <p>Operator's Signature </p>	<p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/17/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:06</p> <p>Control Test 0.080 12:07</p> <p>Air Blank 0.000 12:07</p> <p>Control Test 0.080 12:08</p> <p>Air Blank 0.000 12:08</p> <p>Control Test 0.080 12:09</p> <p>Air Blank 0.000 12:09</p> <p>Control Test Stats</p> <p>Average 0.0800</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/17/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:14</p> <p>Control Test 0.198 12:15</p> <p>Air Blank 0.000 12:16</p> <p>Control Test 0.197 12:16</p> <p>Air Blank 0.000 12:17</p> <p>Control Test 0.198 12:17</p> <p>Air Blank 0.000 12:18</p> <p>Control Test Stats</p> <p>Average 0.1977</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2921</p> <p>Operator's Signature </p>	<p>HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/17/2024 Software: 8100.27</p> <p>SN 80-001063</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:03</p> <p>Control Test 0.081 12:03</p> <p>Air Blank 0.000 12:04</p> <p>Control Test 0.081 12:04</p> <p>Air Blank 0.000 12:04</p> <p>Control Test 0.082 12:05</p> <p>Air Blank 0.000 12:05</p> <p>Control Test Stats</p> <p>Average 0.0813</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7099</p> <p>Operator's Signature </p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD
Time of Inspection: 14:36

Date of Inspection: 04/17/2024

Serial Number: 80-001063
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.050	0.080	0.196	0.081
0.000	0.049	0.080	0.197	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.050	0.080	0.196	0.083
0.000	0.050	0.080	0.197	0.083
0.000	0.050	0.080	0.197	0.082
0.000	0.050	0.080	0.196	0.082
0.000	0.050	0.080	0.196	0.083
0.000	0.050	0.080	0.196	0.083
0.000	0.050	0.080	0.196	0.083

Standard Deviations	0.0004	0.0000	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/17/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001063, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001063	UNCERTAINTY* \pm
Owning Agency:	HOLLYWOOD PD	0.050 g/ 210 L 0.004
Calibration Date:	04/17/2024	0.080 g/ 210 L 0.004
Calibration Time:	14:36	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor D Guttschow

04/17/2024

Date

TAYLOR D GUTTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality