



Agency Pasco County SO

S/N 80-001074

Date In 10/02/2024 DI Completion Date 10/10/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge _____ ID #

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By DA

Barometric Pressure ID# 28427Gauge 1009 Instrument 1008Mouth Alcohol Solution Lot # 2024-A

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|---|---|
| ☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☐ Calibration Adjustment | ☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other _____ |
|---|---|

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2024.10.15 15:30:38
-04'00'

Tech Review / Date

Shayla
Platt

Digitally signed by

~~Shayla Platt~~


Date: 2024.10.16
13:39:14 04'00'

Admin Review / Date

Stability Checks 80-001074 10/10/24 DA

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
10/10/2024
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.050	12:26
Air Blank	0.000	12:27
Control Test	0.050	12:27
Air Blank	0.000	12:28
Control Test	0.050	12:29
Air Blank	0.000	12:29
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

A handwritten signature in blue ink, consisting of a stylized 'A' followed by a vertical line and a horizontal crossbar.

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
10/10/2024
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	12:31
Control Test	0.079	12:32
Air Blank	0.000	12:32
Control Test	0.079	12:33
Air Blank	0.000	12:34
Control Test	0.079	12:34
Air Blank	0.000	12:35
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
10/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.202	12:37
Air Blank	0.000	12:37
Control Test	0.201	12:38
Air Blank	0.000	12:39
Control Test	0.201	12:39
Air Blank	0.000	12:40
Control Test Stats		
Average	0.2013	
Std Dev	0.0006	
Rel Std Dev(%)	0.2868	




Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
10/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:43
Control Test	0.079	12:43
Air Blank	0.000	12:44
Control Test	0.079	12:44
Air Blank	0.000	12:45
Control Test	0.080	12:45
Air Blank	0.000	12:46
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:47

Date of Inspection: 10/10/2024

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.079	0.202	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.051	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.078

Standard Deviations	0.0003	0.0005	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DESTINEE N ARMSTRONG

Signature and Printed Name

10/10/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001074, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001074</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/10/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:47</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee Armstrong
Digitally signed by Destinee Armstrong
Date: 2024.10.14 12:50:50 -04'00'

10/10/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Pasco County SOS/N 80-001074Florida Department of
Law EnforcementDate In 4/3/2024

DI Completion Date _____

☐ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake	By <u>BS</u>	Date <u>4/4/2024</u>	Quality Checks	By <u>DA</u>	Date <u>4/11/2024</u>	Flow Calibration	By _____	Date _____																																
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Keyboard not working reliably -- R value = 100 so I recommend sending to repair DA _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>100</u> ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																		
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A				<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____	
Simulator	Serial #	Lot #/Exp																																						
0.050																																								
0.080																																								
0.200																																								
0.080 DGS	N/A																																							
Maintenance By _____ Date _____																																								
☐ Battery Replacement																																								
☐ Dry Gas Regulator Replacement																																								
☐ Breath Tube Replacement																																								
☐ Other _____																																								

Calibration Adjustment By _____			Department Inspection By _____																																					
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# _____																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____									
Simulator	Serial #	Lot #	Expiration																																					
0.000		N/A	N/A																																					
0.040																																								
0.100																																								
0.200																																								
0.300																																								
0.080 DGS	N/A																																							
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>						Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																					
Simulator	Serial Number																																							
0.000																																								
Interferent																																								
0.050																																								
0.080																																								
0.200																																								
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr><tr><td>☐ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☐ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☐ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></thead></table>						Attachments		☐ Form 41	☐ Post-Stability Checks	☐ Stability Checks	☐ Flow Calibration	☐ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____		
Simulator	Serial #	Lot #	Expiration																																					
0.050																																								
0.080																																								
0.200																																								
0.080 DGS	N/A																																							
Attachments																																								
☐ Form 41	☐ Post-Stability Checks																																							
☐ Stability Checks	☐ Flow Calibration																																							
☐ Calibration Certificate	☐ Form 40																																							
☐ Calibration Adjustment	☐ Other _____																																							
Notes/Suggested Service: _____ Instrument sent to repair 4/15/24. _____ _____ _____ _____ _____ _____			☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																					
			Tech Review / Date _____ Admin Review / Date _____																																					

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Deputy Barry Nixon on 4/14/24

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: 100

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001074

Bill To Address:

ATTN: Deputy Barry Nixon

Pasco County Sheriffs Office

7432 Little Rd

New Port Richey, FL 34654

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department of Law Enforcement

ATTN: Alcohol Testing Program

813 B Lake Bradford Rd

Tallahassee, FL 32304

Reason for Return:

Flow-sensor R value is 100.

Instrument keyboard not working reliably.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Barry Nixon

Phone #: 727-514-6335

Email: BNixon@pascosheriff.org

ATP Contact Name: Destinee Armstrong

ATP Email: destinee.armstrong@fdle.state.fl.us