

Flow Cal Adjust

ML

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 60-001122
03/20/2024
Software: 6100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 7.211
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.746
3: Rate (Liters/min) = 30
SQRT(Diff) = 21.047

*Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 695
Rounded Intercept = -704317
Correlation = 0.99669

Stability Checks

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	
✓		✓		✓		✓	
FL HIGHWAY PATROL Intoxilyzer - Alcotest Analyzer Model 8000 03/20/2024 Software: 8100.27 Test Time g/210L 0.010 11:10 0.049 11:11 0.000 11:11 0.049 11:12 0.000 11:13 0.049 11:13 0.000 11:14 Control Test Stats Average 0.0490 Std Dev 0.010 Rel Std Dev(%) 0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcotest Analyzer Model 8000 03/20/2024 Software: 8100.27 Test Time g/210L 0.000 11:18 0.079 11:18 0.000 11:18 0.079 11:19 0.000 11:20 0.079 11:20 0.000 11:21 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcotest Analyzer Model 8000 03/20/2024 Software: 8100.27 Test Time g/210L 0.000 11:24 0.198 11:25 0.000 11:25 0.197 11:26 0.000 11:27 0.196 11:28 0.000 11:28 Control Test Stats Average 0.1970 Std Dev 0.0010 Rel Std Dev(%) 0.5076		FL HIGHWAY PATROL Intoxilyzer - Alcotest Analyzer Model 8000 03/20/2024 Software: 8100.27 Test Time g/210L 0.000 10:59 0.079 10:59 0.000 11:00 0.079 11:00 0.000 11:01 0.079 11:01 0.000 11:01 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000	
Operator's Signature		Operator's Signature		Operator's Signature		Operator's Signature	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 13:17

Date of Inspection: 03/20/2024

Serial Number: 80-001122
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.080	0.199	0.078


Standard Deviations	0.0004	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/20/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001122, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001122</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/20/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:17</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/20/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality