



Agency Orange County SO

S/N 80-001257

Date In 02-14-2024 DI Completion Date 4/11/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge _____ ID #

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Inspection Completed on 4/11. When attempting to connect the printer cable, the power cable was accidentally hit, causing the instrument to lose power and the inspection. Repeated the Inspection on 4/11 with same simulators and solutions.

PN 4/11/2024

Tech review: added note for tank sensor tare. SP 4/17/2024

Department Inspection

By PNBarometric Pressure ID# 28662/28662

Gauge 1000/1002 Instrument 994/996

Mouth Alcohol Solution Lot # 2023-A

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☒ Form 40 |
| ☐ Calibration Adjustment | ☐ Other _____ |

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2024.04.18 08:52:52
+04'00'

Tech Review / Date

Shayla
Platt

Digitally signed

by Shayla Platt

Date: 2024.04.18

Admin Review / Date 11:05:19-0

Stability Checks

80-001257 03/28/24

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001257
03/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:12
Control Test	0.052	12:13
Air Blank	0.000	12:13
Control Test	0.051	12:14
Air Blank	0.000	12:15
Control Test	0.050	12:15
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001257
03/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:17
Control Test	0.082	12:18
Air Blank	0.000	12:19
Control Test	0.081	12:19
Air Blank	0.000	12:20
Control Test	0.082	12:20
Air Blank	0.000	12:21
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Wet

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001257
03/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:23
Control Test	0.203	12:23
Air Blank	0.000	12:24
Control Test	0.203	12:25
Air Blank	0.000	12:25
Control Test	0.202	12:26
Air Blank	0.000	12:26
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel Std Dev(%)	0.2849	

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001257
03/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.081	12:30
Air Blank	0.000	12:31
Control Test	0.081	12:31
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:32
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 14:23

Date of Inspection: 04/11/2024

Serial Number: 80-001257
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.051	0.081	0.200	0.079 Lot# AG310901
0.000	0.050	0.080	0.200	0.079 Exp: 4/19/2025
0.000	0.050	0.080	0.200	0.079 PN 4/12/2024
0.000	0.051	0.081	0.199	0.079
0.000	0.051	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.081	0.200	0.079
Standard Deviations	0.0005	0.0004	0.0003	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHIL NICODEMO

Signature and Printed Name

04/11/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 08:10

Date of Inspection: 04/11/2024

Serial Number: 80-001257
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

Compliance not Determined
PN 4/11/2024

The above instrument ~~complies~~ (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Phil Nicodemus

PHIL NICODEMO

Signature and Printed Name

04/11/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001257 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001257</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/11/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/11/2024

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality