



Agency **Palm Bay PD**

S/N 80-001265

Date In 07/31/2024 DI Completion Date 8/13/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By

Barometric Pressure Gauge

ID #

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Department Inspection

By PN

Barometric Pressure ID# 28427

Gauge 1012	<u>Instrument 1014</u>
------------	------------------------

Mouth Alcohol Solution Lot # 2024-A

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|---|---|
| ☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☐ Calibration Adjustment | ☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other |
|---|---|

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor Gutschow
Date: 2024.08.22 13:51:04 -04'00'

Platt

Date: 2024.08.22
15:22:42 -04'00'

Tech Review / Date

Admin Review / Date

Stability checks 80-001265 08/01/24

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001265
08/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:34
Control Test	0.050	13:34
Air Blank	0.000	13:35
Control Test	0.050	13:35
Air Blank	0.000	13:36
Control Test	0.050	13:37
Air Blank	0.000	13:37
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001265
08/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:38
Control Test	0.080	13:39
Air Blank	0.000	13:40
Control Test	0.080	13:40
Air Blank	0.000	13:41
Control Test	0.080	13:41
Air Blank	0.000	13:42
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001265
08/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:44
Control Test	0.197	13:44
Air Blank	0.000	13:45
Control Test	0.196	13:46
Air Blank	0.000	13:46
Control Test	0.196	13:47
Air Blank	0.000	13:47
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Operator's Signature

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001265
08/01/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:49
Control Test	0.081	13:50
Air Blank	0.000	13:50
Control Test	0.081	13:50
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:51
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY PD
Time of Inspection: 16:12

Date of Inspection: 08/13/2024

Serial Number: 80-001265
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.049	0.078	0.195	0.080
0.000	0.050	0.079	0.195	0.081
0.000	0.049	0.079	0.196	0.080
0.000	0.049	0.079	0.195	0.080


Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 PHIL NICODEMO
Signature and Printed Name

08/13/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001265, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001265</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PALMBAY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/13/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/13/2024

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1