



Agency Gilchrist County SO

S/N 80-001285

Date In 12/04/2024 DI Completion Date 12/05/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

0.05 g/210 stability check did not print properly. All values were nominal, but repeated test. PN 12/5/2024

Department Inspection

By PN

Attachments

■ Instrument Complies with Chapter 11D-8, FAC

☐ Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

☐ **Remain Out of Evidentiary Use**

Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2024.12.06 13:52:32
+05'00'

Shayla Platt

Shayla Platt

• Date: 2024.12.10

Tech Review / Date

Admin Review / Date

0.05g/2/02

1st attempt

- printout was

obscured. PN 12/5/2024

GILCHRIST COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 3000 SN 80-001285
12/05/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.001	11:20
Control Test	0.050	11:21
Air Blank	0.001	11:21
Control Test	0.049	11:22
Air Blank	0.000	11:23
Control Test	0.049	11:23
Air Blank	0.000	11:24

Control Test Status

Std Dev 0.0006
Rel Std Dev(%) 1.1703


Operator's Signature

STABILITY CHECKS

GILCHRIST COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000
12/05/2024
Software: 8100.27
SN 80-001285

Test	g/210L	Time
Air Blank	0.000	11:42
Control Test	0.078	11:43
Air Blank	0.000	11:43
Control Test	0.079	11:43
Air Blank	0.000	11:44
Control Test	0.078	11:44
Air Blank	0.000	11:45
Control Test Stats		
Average	0.0783	-
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

D65

Operator's Signature

GILCHRIST COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000
12/05/2024
Software: 8100.27
SN 80-001285

Test	g/210L	Time
Air Blank	0.000	11:37
Control Test	0.201	11:37
Air Blank	0.000	11:38
Control Test	0.200	11:39
Air Blank	0.000	11:39
Control Test	0.199	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

Operator's Signature

GILCHRIST COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000
12/05/2024
Software: 8100.27
SN 80-001285

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Control Test	0.079	11:33
Air Blank	0.000	11:33
Control Test	0.078	11:34
Air Blank	0.000	11:34
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Operator's Signature

GILCHRIST COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000
12/05/2024
Software: 8100.27
SN 80-001285

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.049	11:26
Air Blank	0.000	11:27
Control Test	0.049	11:28
Air Blank	0.000	11:28
Control Test	0.049	11:29
Air Blank	0.000	11:29
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GILCHRIST COUNTY SO
Time of Inspection: 14:11

Date of Inspection: 12/05/2024

Serial Number: 80-001285
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:2024062K <i>PN 12/5/24</i> Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/27/2026
0.000	0.050	0.079	0.200	0.077 <i>2 PN 12/5/24</i>
0.000	0.050	0.079	0.200	0.077
0.000	0.050	0.079	0.201	0.077
0.000	0.050	0.079	0.201	0.077
0.000	0.050	0.079	0.201	0.077
0.000	0.050	0.079	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.050	0.079	0.201	0.076
0.000	0.050	0.079	0.201	0.077

Standard Deviations	0.0000	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

PHIL NICODEMO

12/05/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001285, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001285</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>GILCHRIST COUNTY SO</u>	0.050 g/210 L 0.004
Calibration Date:	<u>12/05/2024</u>	0.080 g/210 L 0.004
Calibration Time:	<u>14:11</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/05/2024

Date


PHIL NICODEMO,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality