



Agency Walton County

S/N 80-001311

Date In 08/02/2024 DI Completion Date 10/9/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge _____ ID #

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Mouth alcohol not detected on first attempt during department inspection. Rinsed mouth with additional mouth alcohol and repeated the check. DA

Department Inspection

By DA

Barometric Pressure ID# 30793

Gauge 1008	Instrument 1004
------------	-----------------

Mouth Alcohol Solution Lot # 2024-A

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☒ Form 40 |
| ☐ Calibration Adjustment | ☐ Other _____ |

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo Digitally signed by Phil Nicodemo
Date: 2024.10.15 09:03:43 -04'00'

Tech Review / Date

Shayla

Platt

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 12:51

Date of Inspection: 10/08/2024

Serial Number: 80-001311
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

NA- compliance not determined 10/8/24 DA

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG


Signature and Printed Name

10/08/2024
Date

Stability Checks 80-001311 10/8/24 DA

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
10/08/2024
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	12:53	
Control Test	0.050	12:53	
Air Blank	0.000	12:54	
Control Test	0.050	12:55	
Air Blank	0.000	12:55	
Control Test	0.050	12:56	
Air Blank	0.000	12:56	
Control Test Stats			
Average	0.0500		
Std Dev	0.0000		
Rel Std Dev(%)	0.0000		




Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:58
Control Test	0.078	12:58
Air Blank	0.000	12:59
Control Test	0.078	13:00
Air Blank	0.000	13:00
Control Test	0.078	13:01
Air Blank	0.000	13:01
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Wet



Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:05
Control Test	0.198	13:06
Air Blank	0.000	13:06
Control Test	0.195	13:07
Air Blank	0.000	13:08
Control Test	0.196	13:08
Air Blank	0.000	13:09
Control Test Stats		
Average	0.1963	
Std Dev	0.0015	
Rel Std Dev(%)	0.7780	




Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
10/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:12
Control Test	0.078	13:12
Air Blank	0.000	13:13
Control Test	0.079	13:13
Air Blank	0.000	13:13
Control Test	0.079	13:14
Air Blank	0.000	13:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS



Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 13:30

Date of Inspection: 10/09/2024

Serial Number: 80-001311
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.080	0.197	0.079
0.000	0.051	0.079	0.197	0.079
0.000	0.051	0.079	0.197	0.079
0.000	0.051	0.080	0.197	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.051	0.079	0.198	0.079
0.000	0.051	0.080	0.197	0.078

Standard Deviations	0.0005	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

A F / M A:MOUTH ALCOHOL NOT DETECTED.

Mouth alcohol not detected on first attempt- Rinsed
mouth with additional mouth alcohol and repeated.
DA 10/9/24

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

10/09/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001311, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001311</u>	Serial Number:	<u>80-001311</u>
Owning Agency:	<u>WALTON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/09/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:30</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

 Digitally signed by Destinee Armstrong
Date: 2024.10.10 07:36:22 -0400

10/09/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality