



INSTRUMENT PROCESSING SHEET

Agency Okeechobee CSOS/N 80-001322Florida Department of
Law EnforcementDate In 10/30/2024DI Completion Date 11/04/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>11/01/2024</u>		Quality Checks By <u>TDG</u> Date <u>11/01/2024</u>		Flow Calibration By _____ Date _____																													
☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>172</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.535</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____														
Simulator	Serial #	Lot #/Exp																															
0.050	MP6286	202303K 03/29/2025																															
0.080	MP6287	202303L 03/29/2025																															
0.200	MP6288	202304C 04/05/2025																															
0.080 DGS	N/A	01923080A3 02/05/2025																															
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																													
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u>																													
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>	
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks				<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 47</u>									
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Digitally signed by _____ Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.11.07 14:49:10 -0500 Shayla Platt Shayla Platt Date: 2024.11.08 12:26:02 -05'00' Tech Review / Date _____ Admin Review / Date _____																													



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-001322

OWNING AGENCY: Okeechobee County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 11/04/2024

AGENCY INSPECTOR: Mark Wright

ADDRESS: 504 NW 4th Street

CITY, STATE, ZIP: Okeechobee, FL 34972

TELEPHONE NUMBER: 863-484-3630

FAX NUMBER: N/A

EMAIL ADDRESS (if available): mwright@okeesherriff.com

For Program Office Use Only:



- ☒ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2024 Software: 8100.27 SN 80-001322 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr><tr><td>Control Test</td><td>0.050</td><td>15:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:04</td></tr><tr><td>Control Test</td><td>0.051</td><td>15:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr><tr><td>Control Test</td><td>0.050</td><td>15:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0503</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:03	Control Test	0.050	15:04	Air Blank	0.000	15:04	Control Test	0.051	15:05	Air Blank	0.000	15:06	Control Test	0.050	15:06	Air Blank	0.000	15:07	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2024 Software: 8100.27 SN 80-001322 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:42	Control Test	0.079	14:43	Air Blank	0.000	14:43	Control Test	0.079	14:44	Air Blank	0.000	14:45	Control Test	0.079	14:45	Air Blank	0.000	14:46	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2024 Software: 8100.27 SN 80-001322 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:49	Control Test	0.197	14:50	Air Blank	0.000	14:50	Control Test	0.197	14:51	Air Blank	0.000	14:51	Control Test	0.196	14:52	Air Blank	0.000	14:53	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2936		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2024 Software: 8100.27 SN 80-001322 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:56	Control Test	0.079	14:56	Air Blank	0.000	14:57	Control Test	0.080	14:57	Air Blank	0.000	14:58	Control Test	0.079	14:58	Air Blank	0.000	14:59	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test	0.050	15:04																																																																																																																																																	
Air Blank	0.000	15:04																																																																																																																																																	
Control Test	0.051	15:05																																																																																																																																																	
Air Blank	0.000	15:06																																																																																																																																																	
Control Test	0.050	15:06																																																																																																																																																	
Air Blank	0.000	15:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.079	14:43																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.079	14:44																																																																																																																																																	
Air Blank	0.000	14:45																																																																																																																																																	
Control Test	0.079	14:45																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:49																																																																																																																																																	
Control Test	0.197	14:50																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test	0.197	14:51																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.196	14:52																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.079	14:56																																																																																																																																																	
Air Blank	0.000	14:57																																																																																																																																																	
Control Test	0.080	14:57																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																	
Control Test	0.079	14:58																																																																																																																																																	
Air Blank	0.000	14:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 11:46

Date of Inspection: 11/04/2024

Serial Number: 80-001322
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.051	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.051	0.079	0.197	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.051	0.079	0.196	0.080
0.000	0.051	0.079	0.196	0.079


Standard Deviations	0.0005	0.0000	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/04/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001322, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001322</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OKEECHOBEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/04/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:46</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

11/04/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality