



INSTRUMENT PROCESSING SHEET

Agency Gainesville PDS/N 80-001364Florida Department of
Law EnforcementDate In 10/23/2024 DI Completion Date 10/23/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>10/23/2024</u>		Quality Checks By <u>ALL</u> Date <u>10/23/2024</u>		Flow Calibration By _____ Date _____																																																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Dropoff, with box</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>216</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																															
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901 04/19/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202303L 03/29/2025	0.200	MP5090	202304C 04/05/2025	0.080 DGS	N/A	AG310901 04/19/2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP5088	202406K 06/19/2026																																																																	
0.080	MP5089	202303L 03/29/2025																																																																	
0.200	MP5090	202304C 04/05/2025																																																																	
0.080 DGS	N/A	AG310901 04/19/2025																																																																	
Calibration Adjustment By _____				Department Inspection By <u>DA</u> 10/26/24 DA																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28662</u> Gauge <u>1018</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP5086																																																																		
Interferent	MP5087																																																																		
0.050	MP5088																																																																		
0.080	MP5089																																																																		
0.200	MP5090																																																																		
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																																		
Notes/Suggested Service: <u>RFI detect during minimum sample volume check. Removed device from instrument vicinity and repeated the minimum sample volume check</u> <u>DA</u> Admin Review: <u>Added initials to the "by" line for the department inspection. DA 10/28/24</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Taylor Gutschow Date: 2024.10.24 10:27:58 +04'00' Taylor Gutschow Digitally signed by Shayla Platt Date: 2024.10.29 15:05:09 -04'00' Shayla Platt Tech Review / Date _____ Admin Review / Date _____																																																															

Stability Checks 80-001364

10/23/2024

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001364
10/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.049	11:29
Air Blank	0.000	11:29
Control Test	0.049	11:30
Air Blank	0.000	11:31
Control Test	0.049	11:31
Air Blank	0.000	11:32
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature 

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001364
10/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:33
Control Test	0.081	11:34
Air Blank	0.000	11:34
Control Test	0.080	11:35
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:37
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Operator's Signature 

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001364
10/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:38
Control Test	0.200	11:39
Air Blank	0.000	11:39
Control Test	0.199	11:40
Air Blank	0.000	11:40
Control Test	0.198	11:41
Air Blank	0.000	11:42
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Operator's Signature 

Wet

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001364
10/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.080	11:45
Air Blank	0.000	11:45
Control Test	0.081	11:45
Air Blank	0.000	11:46
Control Test	0.081	11:46
Air Blank	0.000	11:46
Control Test Stats		11:47
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator's Signature 

065

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD

Serial Number: 80-001364

Time of Inspection: 14:14

Date of Inspection: 10/23/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG325603 Exp: 09/13/2025
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.079

Standard Deviations	0.0004	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

REPEATED MIN SAMPLE VOLUME CHECK DUE TO RFI REMOVED DEVICE

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

10/23/2024

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001364 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001364</u>	Serial Number:	<u>80-001364</u>
Owning Agency:	<u>GAINESVILLE PD</u>	Owning Agency:	<u>GAINESVILLE PD</u>
Calibration Date:	<u>10/23/2024</u>	Calibration Date:	<u>10/23/2024</u>
Calibration Time:	<u>14:14</u>	Calibration Time:	<u>14:14</u>
		UNCERTAINTY* $\pm$	
		0.050 g/ 210 L	0.004
		0.080 g/ 210 L	0.004
		0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee Armstrong
Digitally signed by Destinee Armstrong
Date: 2024.10.23 14:50:28 -04'00'

10/23/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality