



Agency Orange County SO

S/N 80-001420

Date In 05/17/2024 DI Completion Date 06/04/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: First DGS stability check gave no values, ensured connections were secure and reran stability 05/20/2024 ALL

Department Inspection

By SP

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2024.06.05 12:20:00
-04'00'

Phil Nicodemo

Digitally signed by Phil Nicodemo
Date: 2024.06.10 14:06:05 -04'00'

Tech Review / Date

Admin Review / Date

stability checks 80-001420 05/20/24

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
05/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.050	11:00
Air Blank	0.000	11:00
Control Test	0.050	11:01
Air Blank	0.000	11:02
Control Test	0.050	11:02
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *AJ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
05/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:05
Control Test	0.079	11:06
Air Blank	0.000	11:07
Control Test	0.079	11:07
Air Blank	0.000	11:08
Control Test	0.079	11:09
Air Blank	0.000	11:09
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

Operator's Signature *AJ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
05/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:11
Control Test	0.198	11:11
Air Blank	0.000	11:12
Control Test	0.197	11:12
Air Blank	0.000	11:13
Control Test	0.197	11:14
Air Blank	0.000	11:14
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Operator's Signature *AJ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
05/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.000	11:18
Air Blank	0.000	11:18
Control Test	0.000	11:19
Air Blank	0.000	11:19
Control Test	0.000	11:20
Air Blank	0.000	11:20
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D6S

Operator's Signature *AJ*

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
05/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.082	11:28
Air Blank	0.000	11:28
Control Test	0.081	11:28
Air Blank	0.000	11:29
Control Test	0.082	11:29
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

D6S

Operator's Signature *AJ*

*ensured connections
perman stability
05/20/24*

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 12:31

Date of Inspection: 06/04/2024

Serial Number: 80-001420
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.080	0.196	0.080
0.000	0.049	0.080	0.196	0.080
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.196	0.080
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.080	0.196	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.080	0.196	0.081
Standard Deviations	0.0004	0.0005	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/04/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001420, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001420</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/04/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:31</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by
Shayla Platt
Date: 2024.06.05
09:25:13 -04'00'

06/04/2024

Date

SHAYLA D PLATT,

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality