



INSTRUMENT PROCESSING SHEET

Agency Palm Bay PDS/N 80-001721Florida Department of
Law EnforcementDate In 10/02/2024 DI Completion Date 10/03/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>10/02/2024</u>		Quality Checks By <u>ALL</u> Date <u>10/02/2024</u>		Flow Calibration By _____ Date _____																																																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>115</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6294</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6295</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6296</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG325603 09/13/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6294	202303K 03/29/2025	0.080	MP6295	202303L 03/29/2025	0.200	MP6296	202304C 04/05/2025	0.080 DGS	N/A	AG325603 09/13/2025	Flow Calibration Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP6294	202303K 03/29/2025																																																																	
0.080	MP6295	202303L 03/29/2025																																																																	
0.200	MP6296	202304C 04/05/2025																																																																	
0.080 DGS	N/A	AG325603 09/13/2025																																																																	
Calibration Adjustment By _____		Department Inspection By <u>DA</u>																																																																	
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28662</u> Gauge <u>1012</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP5086																																																																		
Interferent	MP5087																																																																		
0.050	MP5088																																																																		
0.080	MP5089																																																																		
0.200	MP5090																																																																		
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																																																																		
Notes/Suggested Service: <u>inadvertently ran .050 stability twice 10/02/2024 ALL</u> _____ _____ _____ _____ _____ _____ _____ _____ _____		Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.10.08 11:59:41 +04'00' Shayla Platt Shayla Platt Digitally signed by Shayla Platt Date: 2024.10.09 15:39:32 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 14:14

Date of Inspection: 10/02/2024

Serial Number: 80-001721

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 0


Remarks:

BYPASS AI FOR INSPECTION COMPLIANCE NOT DETERMINED

10/02/24 AL

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALFONSO L LOWRY

Signature and Printed Name

10/02/2024
Date

Stability checks 80-001721 10/02/24

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001721
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:01
Control Test	0.050	15:02
Air Blank	0.000	15:03
Control Test	0.050	15:03
Air Blank	0.000	15:04
Control Test	0.050	15:04
Air Blank	0.000	15:05
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *AZ*

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001721
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:08
Control Test	0.080	15:09
Air Blank	0.000	15:09
Control Test	0.079	15:10
Air Blank	0.000	15:10
Control Test	0.079	15:11
Air Blank	0.000	15:11
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Wet

Operator's Signature *AZ*

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001721
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:28
Control Test	0.196	15:28
Air Blank	0.000	15:29
Control Test	0.196	15:30
Air Blank	0.000	15:30
Control Test	0.197	15:31
Air Blank	0.000	15:31
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Operator's Signature *AZ*

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001721
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:16
Control Test	0.050	15:16
Air Blank	0.000	15:17
Control Test	0.049	15:18
Air Blank	0.000	15:18
Control Test	0.050	15:19
Air Blank	0.000	15:19
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature *AZ*

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001721
10/02/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:33
Control Test	0.079	15:33
Air Blank	0.000	15:34
Control Test	0.079	15:34
Air Blank	0.000	15:34
Control Test	0.080	15:35
Air Blank	0.000	15:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

065

Operator's Signature *AZ*

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY PD

Time of Inspection: 12:57

Date of Inspection: 10/03/2024

Serial Number: 80-001721

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080

Standard Deviations	0.0003	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name DESTINEE N ARMSTRONG

10/03/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite A2025
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001721, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001721</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PALM BAY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/03/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Date
10/03/2024

DESTINEE N ARMSTRONG,
Department Inspector

Service • Integrity • Respect • Quality