



INSTRUMENT PROCESSING SHEET

Agency Lakeland PDS/N 80-001865Florida Department of
Law EnforcementDate In 05/07/2024 DI Completion Date 05/17/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>AI reports the instrument needs a battery change. Missing front left foot on DGS shelf.</u>		<u>05/16/2024</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>226</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>05/17/2024</u>	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																											
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203 08/10/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG222203 08/10/2024													
Simulator	Serial #	Lot #/Exp																															
0.050	MP6286	202303K 03/29/2025																															
0.080	MP6287	202303L 03/29/2025																															
0.200	MP6288	202304C 04/05/2025																															
0.080 DGS	N/A	AG222203 08/10/2024																															
Calibration Adjustment			Department Inspection																														
By _____			By <u>TDG</u>																														
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>26932</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1013</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>		
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____										
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: <u>Checked breath tube screen, replaced o-rings, and verified set-up on 5/16. Finished the rest of the Quality Checks on 5/17. (TDG)</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																														
<u>Added plastic caps/static bag for return shipping. (TDG)</u>			☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																														
			☒ Conduct an Agency Inspection Before Evidentiary Use																														
			Digitally signed by Shayla Platt Phil Nicodemo Date: 2024.05.23 15:00:29 -0400																														
			Tech Review / Date _____ Admin Review / Date _____																														

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓
LAKELAND PD Intoxilyzer - Alcohol Analyzer Model: 8000 05/17/2024 Software: 8100.27 SN 80-001865 Test g/210L Time Air Blank 0.000 12:03 Control Test 0.051 12:04 Air Blank 0.000 12:04 Control Test 0.050 12:05 Air Blank 0.000 12:05 Control Test 0.050 12:06 Air Blank 0.000 12:07 Control Test Stats Average 0.0503 Std Dev 0.0006 Rel. Std Dev(%) 1.1471	LAKELAND PD Intoxilyzer - Alcohol Analyzer Model: 8000 05/17/2024 Software: 8100.27 SN 80-001865 Test g/210L Time Air Blank 0.000 12:10 Control Test 0.080 12:11 Air Blank 0.000 12:11 Control Test 0.079 12:12 Air Blank 0.000 12:13 Control Test 0.079 12:13 Air Blank 0.000 12:14 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel. Std Dev(%) 0.7277	LAKELAND PD Intoxilyzer - Alcohol Analyzer Model: 8000 05/17/2024 Software: 8100.27 SN 80-001865 Test g/210L Time Air Blank 0.000 12:17 Control Test 0.197 12:17 Air Blank 0.000 12:18 Control Test 0.197 12:19 Air Blank 0.000 12:19 Control Test 0.197 12:20 Air Blank 0.000 12:20 Control Test Stats Average 0.1970 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	065 LAKELAND PD Intoxilyzer - Alcohol Analyzer Model: 8000 05/17/2024 Software: 8100.27 SN 80-001865 Test g/210L Time Air Blank 0.000 11:53 Control Test 0.080 11:53 Air Blank 0.000 11:54 Control Test 0.080 11:54 Air Blank 0.000 11:54 Control Test 0.081 11:55 Air Blank 0.000 11:55 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel. Std Dev(%) 0.7187
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 14:50

Date of Inspection: 05/17/2024

Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.080	0.197	0.080
0.000	0.051	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081

Standard Deviations	0.0006	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 TAYLOR D GUTSCHOW
 Signature and Printed Name

05/17/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001865, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001865</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LAKE LAND PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/17/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:50</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


05/17/2024 Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality