



INSTRUMENT PROCESSING SHEET

Agency NAS Key WestS/N 80-002170Florida Department of
Law EnforcementDate In 08/27/2024 DI Completion Date 09/06/2024☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>TDG</u> Date <u>09/05/2024</u>		Quality Checks By <u>TDG</u> Date <u>09/06/2024</u>		Flow Calibration By _____ Date _____																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Had difficulty powering up the instrument.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>248</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP5099</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5100</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5101</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5099	202303K 03/29/2025	0.080	MP5100	202303L 03/29/2025	0.200	MP5101	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5099	202303K 03/29/2025																																																															
0.080	MP5100	202303L 03/29/2025																																																															
0.200	MP5101	202304C 04/05/2025																																																															
0.080 DGS	N/A	01923080A3 02/05/2025																																																															
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP5097</td></tr><tr><td>Interferent</td><td>MP5098</td></tr><tr><td>0.050</td><td>MP5099</td></tr><tr><td>0.080</td><td>MP5100</td></tr><tr><td>0.200</td><td>MP5101</td></tr></table>				Simulator	Serial Number	0.000	MP5097	Interferent	MP5098	0.050	MP5099	0.080	MP5100	0.200	MP5101
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP5097																																																																
Interferent	MP5098																																																																
0.050	MP5099																																																																
0.080	MP5100																																																																
0.200	MP5101																																																																
Notes/Suggested Service: <u>Added plastic caps for shipping to CMI. (TDG)</u>		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u>																																																															
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																																															
		Digitally signed by <u>Shayla Platt</u> Date: 2024.09.20 09:47:50 -04'00' Destinee Armstrong Tech Review / Date _____ Admin Review / Date _____																																																															

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Lee Anderson on 09/13/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-002170

Bill To Address:

Naval Air Station Key West

Attn: Lee Anderson

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Had difficulty powering on and required several attempts to turn on. Records were uploaded by
FDLE.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Lee Anderson

Phone #: 863-521-2041 Email: lee.b.anderson.civ@us.navy.mil

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST
Time of Inspection: 11:23

Date of Inspection: 09/06/2024

Serial Number: 80-002170
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:


AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined

MG
09/06/2024

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/06/2024
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
US 450 4507 PROVLABEL - Accrol Analyser Model: 8000 SN: 80-002170 19/10/2024 Software: 8100.27 <table><thead><tr><th>Test</th><th>9/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr><tr><td>Control Test</td><td>0.049</td><td>13:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel. Std Dev (%)</td><td>1.1625</td><td></td></tr></tbody></table>	Test	9/210L	Time	Air Blank	0.000	13:15	Control Test	0.050	13:16	Air Blank	0.000	13:17	Control Test	0.050	13:17	Air Blank	0.000	13:18	Control Test	0.049	13:18	Air Blank	0.000	13:19	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel. Std Dev (%)	1.1625		US 450 4507 PROVLABEL - Accrol Analyser Model: 8000 SN: 80-002170 19/10/2024 Software: 8100.27 <table><thead><tr><th>Test</th><th>9/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:21</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev (%)</td><td>1.2658</td><td></td></tr></tbody></table>	Test	9/210L	Time	Air Blank	0.000	13:19	Control Test	0.078	13:19	Air Blank	0.000	13:20	Control Test	0.079	13:20	Air Blank	0.000	13:21	Control Test	0.080	13:21	Air Blank	0.000	13:22	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev (%)	1.2658		US 450 4507 PROVLABEL - Accrol Analyser Model: 8000 SN: 80-002172 19/10/2024 Software: 8100.27 <table><thead><tr><th>Test</th><th>9/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr><tr><td>Control Test</td><td>0.098</td><td>12:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr><tr><td>Control Test</td><td>0.098</td><td>12:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr><tr><td>Control Test</td><td>0.098</td><td>12:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0987</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev (%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	9/210L	Time	Air Blank	0.000	12:49	Control Test	0.098	12:50	Air Blank	0.000	12:51	Control Test	0.098	12:51	Air Blank	0.000	12:52	Control Test	0.098	12:52	Air Blank	0.000	12:53	Control Test Stats			Average	0.0987		Std Dev	0.0000		Rel. Std Dev (%)	0.0000		US 450 4507 PROVLABEL - Accrol Analyser Model: 8000 SN: 80-002171 19/10/2024 Software: 8100.27 <table><thead><tr><th>Test</th><th>9/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev (%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	9/210L	Time	Air Blank	0.000	12:49	Control Test	0.079	12:50	Air Blank	0.000	12:51	Control Test	0.079	12:51	Air Blank	0.000	12:52	Control Test	0.079	12:52	Air Blank	0.000	12:53	Control Test Stats			Average	0.0793		Std Dev	0.0000		Rel. Std Dev (%)	0.0000	
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.050	13:16																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.050	13:17																																																																																																																																																	
Air Blank	0.000	13:18																																																																																																																																																	
Control Test	0.049	13:18																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev (%)	1.1625																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.078	13:19																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test	0.079	13:20																																																																																																																																																	
Air Blank	0.000	13:21																																																																																																																																																	
Control Test	0.080	13:21																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev (%)	1.2658																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.098	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.098	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test	0.098	12:52																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0987																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev (%)	0.0000																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.079	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.079	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test	0.079	12:52																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev (%)	0.0000																																																																																																																																																		
065	ML	ML	ML																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST
Time of Inspection: 15:25

Date of Inspection: 09/06/2024

Serial Number: 80-002170
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.051	0.080	0.198	0.078
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.079	0.196	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.079


Standard Deviations	0.0004	0.0003	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/06/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002170, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002170</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>NAS KEY WEST</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/06/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:25</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/06/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality