



INSTRUMENT PROCESSING SHEET

Agency Manatee CSOS/N 80-006236Florida Department of
Law EnforcementDate In 07/15/2024 DI Completion Date 07/15/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>07/15/2024</u>		Quality Checks By <u>TDG</u> Date <u>07/15/2024</u>		Flow Calibration By _____ Date _____																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Missing left middle foot.</u> <u>Dropped off.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>235</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP6286	202303K 03/29/2025																																																															
0.080	MP6287	202303L 03/29/2025																																																															
0.200	MP6288	202304C 04/05/2025																																																															
0.080 DGS	N/A	01923080A3 02/05/2025																																																															
				Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																													
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1019</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6284																																																																
Interferent	MP6285																																																																
0.050	MP6286																																																																
0.080	MP6287																																																																
0.200	MP6288																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																													
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemus Digitally signed by Phil Nicodemus Date: 2024.07.25 09:00:45 +04'00' Shayla Platt Digitally signed by Shayla Platt Date: 2024.07.26 14:52:21 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																							
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																							
✓	✓	✓	✓																																																																																																																																							
MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2024 Software: 8100.27 SN 80-016236 Test Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.3727</td><td></td></tr> </table>	Air Blank	0.000	12:43	Control Test	0.050	12:43	Air Blank	0.000	12:44	Control Test	0.048	12:45	Air Blank	0.000	12:45	Control Test	0.048	12:46	Air Blank	0.000	12:47	Control Test Stats			Average	0.0487		Std Dev	0.0010		Rel Std Dev(%)	2.3727		MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2024 Software: 8100.27 SN 80-016236 Test Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </table>	Air Blank	0.000	12:57	Control Test	0.078	12:57	Air Blank	0.000	12:58	Control Test	0.077	12:59	Air Blank	0.000	12:59	Control Test	0.078	13:00	Air Blank	0.000	13:01	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2024 Software: 8100.27 SN 80-016236 Test Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.192</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.194</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1947</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5932</td><td></td></tr> </table>	Air Blank	0.000	12:50	Control Test	0.196	12:51	Air Blank	0.000	12:51	Control Test	0.192	12:52	Air Blank	0.000	12:53	Control Test	0.194	12:53	Air Blank	0.000	12:54	Control Test Stats			Average	0.1947		Std Dev	0.0012		Rel Std Dev(%)	0.5932		MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2024 Software: 8100.27 SN 80-016236 Test Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </table>	Air Blank	0.000	12:37	Control Test	0.079	12:37	Air Blank	0.000	12:38	Control Test	0.080	12:38	Air Blank	0.000	12:38	Control Test	0.080	12:39	Air Blank	0.000	12:39	Control Test	0.080	12:39	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Air Blank	0.000	12:43																																																																																																																																								
Control Test	0.050	12:43																																																																																																																																								
Air Blank	0.000	12:44																																																																																																																																								
Control Test	0.048	12:45																																																																																																																																								
Air Blank	0.000	12:45																																																																																																																																								
Control Test	0.048	12:46																																																																																																																																								
Air Blank	0.000	12:47																																																																																																																																								
Control Test Stats																																																																																																																																										
Average	0.0487																																																																																																																																									
Std Dev	0.0010																																																																																																																																									
Rel Std Dev(%)	2.3727																																																																																																																																									
Air Blank	0.000	12:57																																																																																																																																								
Control Test	0.078	12:57																																																																																																																																								
Air Blank	0.000	12:58																																																																																																																																								
Control Test	0.077	12:59																																																																																																																																								
Air Blank	0.000	12:59																																																																																																																																								
Control Test	0.078	13:00																																																																																																																																								
Air Blank	0.000	13:01																																																																																																																																								
Control Test Stats																																																																																																																																										
Average	0.0777																																																																																																																																									
Std Dev	0.0006																																																																																																																																									
Rel Std Dev(%)	0.7434																																																																																																																																									
Air Blank	0.000	12:50																																																																																																																																								
Control Test	0.196	12:51																																																																																																																																								
Air Blank	0.000	12:51																																																																																																																																								
Control Test	0.192	12:52																																																																																																																																								
Air Blank	0.000	12:53																																																																																																																																								
Control Test	0.194	12:53																																																																																																																																								
Air Blank	0.000	12:54																																																																																																																																								
Control Test Stats																																																																																																																																										
Average	0.1947																																																																																																																																									
Std Dev	0.0012																																																																																																																																									
Rel Std Dev(%)	0.5932																																																																																																																																									
Air Blank	0.000	12:37																																																																																																																																								
Control Test	0.079	12:37																																																																																																																																								
Air Blank	0.000	12:38																																																																																																																																								
Control Test	0.080	12:38																																																																																																																																								
Air Blank	0.000	12:38																																																																																																																																								
Control Test	0.080	12:39																																																																																																																																								
Air Blank	0.000	12:39																																																																																																																																								
Control Test	0.080	12:39																																																																																																																																								
Control Test Stats																																																																																																																																										
Average	0.0797																																																																																																																																									
Std Dev	0.0006																																																																																																																																									
Rel Std Dev(%)	0.7247																																																																																																																																									
Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>																																																																																																																																							

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 14:46

Date of Inspection: 07/15/2024

Serial Number: 80-006236
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.048	0.078	0.195	0.078
0.000	0.049	0.077	0.195	0.079
0.000	0.049	0.078	0.195	0.078
0.000	0.049	0.078	0.194	0.079
0.000	0.050	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.195	0.078
0.000	0.049	0.078	0.195	0.078
0.000	0.049	0.078	0.194	0.078

Standard Deviations	0.0004	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/15/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006236, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006236</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/15/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:46</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/15/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality