



INSTRUMENT PROCESSING SHEET

Agency Pinecrest PDS/N 80-006473Florida Department of
Law EnforcementDate In 03/20/2024 DI Completion Date 03/20/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																																																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☒ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Dropped off with DGS cylinder. Missing support pin for left side of keyboard.		03/20/2024	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>144</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		03/20/2024	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																			
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	<table border="1"><thead><tr><th colspan="2">Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement</td><td></td></tr><tr><td colspan="2">☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td colspan="2">☐ Breath Tube Replacement</td><td></td></tr><tr><td colspan="2">☐ Other _____</td><td></td></tr></tbody></table>			Maintenance By		Date	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____																																					
Simulator	Serial #	Lot #/Exp																																																																							
0.050	MP6286	202303K 03/29/2025																																																																							
0.080	MP6287	202303L 03/29/2025																																																																							
0.200	MP6288	202304C 04/05/2025																																																																							
0.080 DGS	N/A	01923080A3 02/05/2025																																																																							
Maintenance By		Date																																																																							
☐ Battery Replacement																																																																									
☐ Dry Gas Regulator Replacement																																																																									
☐ Breath Tube Replacement																																																																									
☐ Other _____																																																																									
Calibration Adjustment By _____			Department Inspection By <u>TDG</u>																																																																						
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1019</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																						
0.000		N/A	N/A																																																																						
0.040																																																																									
0.100																																																																									
0.200																																																																									
0.300																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial #	Lot #	Expiration																																																																						
0.050																																																																									
0.080																																																																									
0.200																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial Number																																																																								
0.000	MP6284																																																																								
Interferent	MP6285																																																																								
0.050	MP6286																																																																								
0.080	MP6287																																																																								
0.200	MP6288																																																																								
☒ Form 41	☐ Post-Stability Checks																																																																								
☒ Stability Checks	☐ Flow Calibration																																																																								
☒ Calibration Certificate	☐ Form 40																																																																								
☐ Calibration Adjustment	☐ Other _____																																																																								
Notes/Suggested Service: _____ _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Phil Digitally signed by Phil Nicodemo Shayla Digitally signed by Shayla Platt Nicodemo Date: 2024.03.20 14:30:24 -04'00' Platt Date: 2024.03.20 21:02:38 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																																						

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
PINECREST PD Intoxilyzer - Alcohol Analyzer Model 8000 03/20/2024 Software: 8100.27 SN 80-006473 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:02	Control Test	0.050	11:03	Air Blank	0.000	11:04	Control Test	0.050	11:04	Air Blank	0.000	11:05	Control Test	0.050	11:06	Air Blank	0.000	11:06	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PINECREST PD Intoxilyzer - Alcohol Analyzer Model 8000 03/20/2024 Software: 8100.27 SN 80-006473 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:09	Control Test	0.080	11:10	Air Blank	0.000	11:11	Control Test	0.080	11:11	Air Blank	0.000	11:12	Control Test	0.080	11:13	Air Blank	0.000	11:13	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PINECREST PD Intoxilyzer - Alcohol Analyzer Model 8000 03/20/2024 Software: 8100.27 SN 80-006473 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1977</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:17	Control Test	0.198	11:18	Air Blank	0.000	11:18	Control Test	0.198	11:19	Air Blank	0.000	11:19	Control Test	0.197	11:20	Air Blank	0.000	11:21	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		065 PINECREST PD Intoxilyzer - Alcohol Analyzer Model 8000 03/20/2024 Software: 8100.27 SN 80-006473 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.083</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.082</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test</td><td>0.082</td><td>10:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0823</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7012</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	10:51	Control Test	0.083	10:51	Air Blank	0.000	10:51	Control Test	0.082	10:52	Air Blank	0.000	10:52	Control Test	0.082	10:53	Air Blank	0.000	10:53	Control Test Stats			Average	0.0823		Std Dev	0.0006		Rel Std Dev(%)	0.7012	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.050	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.050	11:04																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test	0.050	11:06																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test	0.080	11:10																																																																																																																																																	
Air Blank	0.000	11:11																																																																																																																																																	
Control Test	0.080	11:11																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test	0.080	11:13																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.198	11:18																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.198	11:19																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.197	11:20																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.083	10:51																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.082	10:52																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.082	10:53																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0823																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7012																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINECREST PD

Time of Inspection: 13:27

Date of Inspection: 03/20/2024

Serial Number: 80-006473

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.080	0.197	0.082
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.080	0.197	0.082
0.000	0.050	0.080	0.196	0.081
0.000	0.051	0.079	0.196	0.082
0.000	0.051	0.080	0.196	0.082
0.000	0.050	0.080	0.197	0.082
0.000	0.050	0.079	0.197	0.082


Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/20/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006473, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006473</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PINECREST PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/20/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:27</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

TAYLOR D GUTSCHOW,
Department Inspector

03/20/2024

Date

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality