



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-006736Florida Department of
Law EnforcementDate In 11/08/2024 DI Completion Date 11/12/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ Date <u>11/08/2024</u> ☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Cleared records.</u>	Quality Checks By TDG _____ Date <u>11/08/2024</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>117</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132*</u> (.139 - .169) 36 mm <u>0.152*</u> (.156 - .190) 53 mm <u>0.226*</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Flow Calibration By TDG _____ Date <u>11/08/2024</u> Flow Column # <u>ATP106</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>118</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP6286	202303K 03/29/2025																																																
0.080	MP6287	202303L 03/29/2025																																																
0.200	MP6288	202304C 04/05/2025																																																
0.080 DGS	N/A	01923080A3 02/05/2025																																																
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service: <u>*Out of nominal range. (TDG)</u>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection By TDG _____ Barometric Pressure ID# <u>26932</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 47</u> Instrument Compliance ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2024.11.18 11:36:22 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2024.11.19 13:35:32 -05'00' Tech Review / Date _____ Admin Review / Date _____			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																																				
Simulator	Serial Number																																																	
0.000	MP6284																																																	
Interferent	MP6285																																																	
0.050	MP6286																																																	
0.080	MP6287																																																	
0.200	MP6288																																																	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:56

Date of Inspection: 11/08/2024

Serial Number: 80-006736
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined

*ML-
11/08/2024*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2024
Date

FL HIGHWAY PATROL

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-006736

11/08/2024

Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 6.164

2: Rate (Liters/min) = 15

SQRT(Diff)) = 10.816

3: Rate (Liters/min) = 30

SQRT(Diff)) = 20.637

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 663

Rounded Intercept = -460477

Correlation = 0.99610

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27 SN 80-006736 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.052</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> <tr><td>Control Test</td><td>0.052</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.053</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0523</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1032</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.052	13:23	Air Blank	0.000	13:23	Control Test	0.052	13:24	Air Blank	0.000	13:24	Control Test	0.053	13:25	Air Blank	0.000	13:26	Control Test Stats			Average	0.0523		Std Dev	0.0006		Rel Std Dev(%)	1.1032		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27 SN 80-006736 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.083</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.8704</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:29	Control Test	0.080	13:29	Air Blank	0.000	13:30	Control Test	0.082	13:31	Air Blank	0.000	13:31	Control Test	0.083	13:32	Air Blank	0.000	13:32	Control Test Stats			Average	0.0817		Std Dev	0.0015		Rel Std Dev(%)	1.8704		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27 SN 80-006736 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5025</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:35	Control Test	0.200	13:36	Air Blank	0.000	13:36	Control Test	0.199	13:37	Air Blank	0.000	13:38	Control Test	0.198	13:38	Air Blank	0.000	13:39	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel Std Dev(%)	0.5025		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27 SN 80-006736 065 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.083</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.083</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0827</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.6984</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:09	Control Test	0.083	13:09	Air Blank	0.000	13:10	Control Test	0.082	13:10	Air Blank	0.000	13:10	Control Test	0.083	13:11	Air Blank	0.000	13:11	Control Test Stats			Average	0.0827		Std Dev	0.0006		Rel Std Dev(%)	0.6984	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test	0.052	13:23																																																																																																																																																	
Air Blank	0.000	13:23																																																																																																																																																	
Control Test	0.052	13:24																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.053	13:25																																																																																																																																																	
Air Blank	0.000	13:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0523																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1032																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:29																																																																																																																																																	
Control Test	0.080	13:29																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.082	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.083	13:32																																																																																																																																																	
Air Blank	0.000	13:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0817																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	1.8704																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:35																																																																																																																																																	
Control Test	0.200	13:36																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.199	13:37																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.198	13:38																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.083	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.082	13:10																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.083	13:11																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0827																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.6984																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:19

Date of Inspection: 11/12/2024

Serial Number: 80-006736
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.051	0.080	0.195	0.084
0.000	0.052	0.081	0.195	0.083
0.000	0.051	0.081	0.196	0.084
0.000	0.052	0.082	0.196	0.083
0.000	0.051	0.081	0.195	0.084
0.000	0.052	0.081	0.195	0.084
0.000	0.051	0.081	0.196	0.083
0.000	0.051	0.081	0.196	0.083
0.000	0.051	0.080	0.197	0.083
0.000	0.051	0.081	0.197	0.084


Standard Deviations	0.0004	0.0005	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/12/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006736, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006736</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/12/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:19</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



11/12/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-006736

OWNING AGENCY: Florida Highway Patrol

DATE OF DEPARTMENT INSPECTION: 11/12/2024

AGENCY INSPECTOR: Andrew Sackmann

ADDRESS: 1011 NW 111 Avenue

CITY, STATE, ZIP: Miami, Florida 33172

TELEPHONE NUMBER: 239-293-5529

FAX NUMBER: N/A

EMAIL ADDRESS (if available): AndrewSackmann@flhsmv.gov

For Program Office Use Only:



- ☒ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database