



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-006758Florida Department of
Law EnforcementDate In 04/11/2024 DI Completion Date 04/15/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ Date <u>04/15/2024</u>		Quality Checks By TDG _____ Date <u>04/15/2024</u>		Flow Calibration By TDG _____ Date <u>04/15/2024</u>																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Dry gas elbow broken off and included in plastic baggie.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>224</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.183*</u> (.139 - .169) 36 mm <u>0.203*</u> (.156 - .190) 53 mm <u>0.261</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203 08/10/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG222203 08/10/2024	☒ Flow Column # <u>ATP106</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>225</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) Maintenance By TDG _____ Date <u>04/15/2024</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced dry gas elbow prior to Quality Checks.</u>																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP6286	202303K 03/29/2025																																																															
0.080	MP6287	202303L 03/29/2025																																																															
0.200	MP6288	202304C 04/05/2025																																																															
0.080 DGS	N/A	AG222203 08/10/2024																																																															
Calibration Adjustment By _____		Department Inspection By TDG _____																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>68639</u> Gauge <u>1020</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemmo Digitally signed by Phil Nicodemmo Date: 2024.04.17 13:08:46 -04'00' Shayla Platt Digitally signed by Shayla Platt Date: 2024.04.18 09:16:04 -04'00' Tech Review / Date _____ Admin Review / Date _____				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6284																																																																
Interferent	MP6285																																																																
0.050	MP6286																																																																
0.080	MP6287																																																																
0.200	MP6288																																																																
Notes/Suggested Service: <u>*Flow values outside nominal range. (TDG)</u> <u>Added plastic caps for return to agency. (TDG)</u>																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:19

Date of Inspection: 04/15/2024

Serial Number: 80-006758
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined ^{ML} 04/15/2024

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

04/15/2024
Date

Flow Cal Adjust

MB

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006758
04/15/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.926

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.789

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.855

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 694

Rounded Intercept = -677087

Correlation = 0.99835

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																																																		
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																																																		
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 04/15/2024 SN 80-006758 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:23	Control Test	0.050	11:24	Air Blank	0.000	11:25	Control Test	0.049	11:25	Air Blank	0.000	11:26	Control Test	0.050	11:27	Air Blank	0.000	11:27	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 04/15/2024 SN 80-006758 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:29	Control Test	0.080	11:30	Air Blank	0.000	11:31	Control Test	0.080	11:31	Air Blank	0.000	11:32	Control Test	0.079	11:33	Air Blank	0.000	11:33	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 04/15/2024 SN 80-006758 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr><tr><td>Control Test</td><td>0.196</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.195	11:36	Air Blank	0.000	11:37	Control Test	0.196	11:37	Air Blank	0.000	11:38	Control Test	0.195	11:39	Air Blank	0.000	11:39	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 04/15/2024 SN 80-006758 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr><tr><td>Control Test</td><td>0.196</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.195	11:36	Air Blank	0.000	11:37	Control Test	0.196	11:37	Air Blank	0.000	11:38	Control Test	0.195	11:39	Air Blank	0.000	11:39	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 04/15/2024 SN 80-006758 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr><tr><td>Control Test</td><td>0.196</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.195	11:36	Air Blank	0.000	11:37	Control Test	0.196	11:37	Air Blank	0.000	11:38	Control Test	0.195	11:39	Air Blank	0.000	11:39	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951	
Test	g/210L	Time																																																																																																																																																																																						
Air Blank	0.000	11:23																																																																																																																																																																																						
Control Test	0.050	11:24																																																																																																																																																																																						
Air Blank	0.000	11:25																																																																																																																																																																																						
Control Test	0.049	11:25																																																																																																																																																																																						
Air Blank	0.000	11:26																																																																																																																																																																																						
Control Test	0.050	11:27																																																																																																																																																																																						
Air Blank	0.000	11:27																																																																																																																																																																																						
Control Test Stats																																																																																																																																																																																								
Average	0.0497																																																																																																																																																																																							
Std Dev	0.0006																																																																																																																																																																																							
Rel Std Dev(%)	1.1625																																																																																																																																																																																							
Test	g/210L	Time																																																																																																																																																																																						
Air Blank	0.000	11:29																																																																																																																																																																																						
Control Test	0.080	11:30																																																																																																																																																																																						
Air Blank	0.000	11:31																																																																																																																																																																																						
Control Test	0.080	11:31																																																																																																																																																																																						
Air Blank	0.000	11:32																																																																																																																																																																																						
Control Test	0.079	11:33																																																																																																																																																																																						
Air Blank	0.000	11:33																																																																																																																																																																																						
Control Test Stats																																																																																																																																																																																								
Average	0.0797																																																																																																																																																																																							
Std Dev	0.0006																																																																																																																																																																																							
Rel Std Dev(%)	0.7247																																																																																																																																																																																							
Test	g/210L	Time																																																																																																																																																																																						
Air Blank	0.000	11:35																																																																																																																																																																																						
Control Test	0.195	11:36																																																																																																																																																																																						
Air Blank	0.000	11:37																																																																																																																																																																																						
Control Test	0.196	11:37																																																																																																																																																																																						
Air Blank	0.000	11:38																																																																																																																																																																																						
Control Test	0.195	11:39																																																																																																																																																																																						
Air Blank	0.000	11:39																																																																																																																																																																																						
Control Test Stats																																																																																																																																																																																								
Average	0.1957																																																																																																																																																																																							
Std Dev	0.0006																																																																																																																																																																																							
Rel Std Dev(%)	0.2951																																																																																																																																																																																							
Test	g/210L	Time																																																																																																																																																																																						
Air Blank	0.000	11:35																																																																																																																																																																																						
Control Test	0.195	11:36																																																																																																																																																																																						
Air Blank	0.000	11:37																																																																																																																																																																																						
Control Test	0.196	11:37																																																																																																																																																																																						
Air Blank	0.000	11:38																																																																																																																																																																																						
Control Test	0.195	11:39																																																																																																																																																																																						
Air Blank	0.000	11:39																																																																																																																																																																																						
Control Test Stats																																																																																																																																																																																								
Average	0.1957																																																																																																																																																																																							
Std Dev	0.0006																																																																																																																																																																																							
Rel Std Dev(%)	0.2951																																																																																																																																																																																							
Test	g/210L	Time																																																																																																																																																																																						
Air Blank	0.000	11:35																																																																																																																																																																																						
Control Test	0.195	11:36																																																																																																																																																																																						
Air Blank	0.000	11:37																																																																																																																																																																																						
Control Test	0.196	11:37																																																																																																																																																																																						
Air Blank	0.000	11:38																																																																																																																																																																																						
Control Test	0.195	11:39																																																																																																																																																																																						
Air Blank	0.000	11:39																																																																																																																																																																																						
Control Test Stats																																																																																																																																																																																								
Average	0.1957																																																																																																																																																																																							
Std Dev	0.0006																																																																																																																																																																																							
Rel Std Dev(%)	0.2951																																																																																																																																																																																							

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006758
04/15/2024
Software: 8100.27

NLS

Test	g/210L	Time
Air Blānk	0.000	11:20
Control Test	0.082	11:20
Air Blank	0.000	11:21
Control Test	0.082	11:21
Air Blank	0.000	11:22
Control Test	0.081	11:22
Air Blank	0.000	11:23
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:05

Date of Inspection: 04/15/2024

Serial Number: 80-006758
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.050	0.079	0.197	0.082
0.000	0.050	0.078	0.196	0.081
0.000	0.049	0.079	0.196	0.081
0.000	0.050	0.079	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.050	0.079	0.196	0.081
0.000	0.049	0.079	0.197	0.080

Standard Deviations	0.0004	0.0005	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/15/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006758, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006758</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/15/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:05</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/15/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality