



INSTRUMENT PROCESSING SHEET

Agency Collier CSO

S/N 80-007077

Florida Department of
Law EnforcementDate In 06/21/2024 DI Completion Date 06/25/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date <u>06/24/2024</u>	Quality Checks	By TDG	Date <u>06/25/2024</u>	Flow Calibration	By TDG	Date <u>06/25/2024</u>																																	
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>139</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132*</u> (.139 - .169) 36 mm <u>0.152*</u> (.156 - .190) 53 mm <u>0.222*</u> (.228 - .278) 103 mm <u>0.484</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks			Flow Column # <u>ATP106</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>142</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547)																																			
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5099</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5100</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5101</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5099	202303K 03/29/2025	0.080	MP5100	202303L 03/29/2025	0.200	MP5101	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	<table border="1"><thead><tr><th>Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other</td><td></td></tr><tr><td></td><td></td></tr><tr><td></td><td></td></tr><tr><td></td><td></td></tr><tr><td></td><td></td></tr></tbody></table>					Maintenance By	Date	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other									
Simulator	Serial #	Lot #/Exp																																							
0.050	MP5099	202303K 03/29/2025																																							
0.080	MP5100	202303L 03/29/2025																																							
0.200	MP5101	202304C 04/05/2025																																							
0.080 DGS	N/A	01923080A3 02/05/2025																																							
Maintenance By	Date																																								
☐ Battery Replacement																																									
☐ Dry Gas Regulator Replacement																																									
☐ Breath Tube Replacement																																									
☐ Other																																									

Calibration Adjustment	By	Department Inspection	By TDG																													
Barometric Pressure Gauge ID # _____		Barometric Pressure ID# <u>26932</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A				Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>		
Simulator	Serial #	Lot #	Expiration																													
0.000		N/A	N/A																													
0.040																																
0.100																																
0.200																																
0.300																																
0.080 DGS	N/A																															
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td></tr><tr><td>Interferent</td><td>MP5098</td></tr><tr><td>0.050</td><td>MP5099</td></tr><tr><td>0.080</td><td>MP5100</td></tr><tr><td>0.200</td><td>MP5101</td></tr></tbody></table>	Simulator	Serial Number	0.000	MP5097	Interferent	MP5098	0.050	MP5099	0.080	MP5100	0.200	MP5101																		
Simulator	Serial Number																															
0.000	MP5097																															
Interferent	MP5098																															
0.050	MP5099																															
0.080	MP5100																															
0.200	MP5101																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				<table border="1"><thead><tr><th>Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td></tr><tr><td>☒ Stability Checks</td></tr><tr><td>☒ Calibration Certificate</td></tr><tr><td>☐ Calibration Adjustment</td></tr><tr><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Flow Calibration</td></tr><tr><td>☐ Form 40</td></tr><tr><td>☐ Other</td></tr></tbody></table>	Attachments	☒ Form 41	☒ Stability Checks	☒ Calibration Certificate	☐ Calibration Adjustment	☐ Post-Stability Checks	☒ Flow Calibration	☐ Form 40	☐ Other	
Simulator	Serial #	Lot #	Expiration																													
0.050																																
0.080																																
0.200																																
0.080 DGS	N/A																															
Attachments																																
☒ Form 41																																
☒ Stability Checks																																
☒ Calibration Certificate																																
☐ Calibration Adjustment																																
☐ Post-Stability Checks																																
☒ Flow Calibration																																
☐ Form 40																																
☐ Other																																

Notes/Suggested Service: *Flow values were outside the nominal range. (TDG)

- ☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

Phil
NicodemoDigitally signed by Phil Nicodemo
Date: 2024.06.27 13:12:05 -0400

Tech Review / Date

Shayla
PlattShayla Platt
Date: 2024.06.27
15:36:04 -0400

Admin Review / Date

Flow Calibration Adjustment(s)

Performed by TDG

COLLIER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007077
06/25/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.781

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.617

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.613

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 699

Rounded Intercept = -660610

Correlation = 0.99838

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/25/2024 Software: 8100.27 SN 80-007077 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:38	Control Test	0.050	10:38	Air Blank	0.000	10:39	Control Test	0.050	10:40	Air Blank	0.000	10:40	Control Test	0.049	10:41	Air Blank	0.000	10:41	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/25/2024 Software: 8100.27 SN 80-007077 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:48</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:44	Control Test	0.079	10:45	Air Blank	0.000	10:46	Control Test	0.079	10:46	Air Blank	0.000	10:47	Control Test	0.079	10:48	Air Blank	0.000	10:48	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/25/2024 Software: 8100.27 SN 80-007077 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5025</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:51	Control Test	0.200	10:51	Air Blank	0.000	10:52	Control Test	0.199	10:53	Air Blank	0.000	10:53	Control Test	0.198	10:54	Air Blank	0.000	10:54	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel Std Dev(%)	0.5025		COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/25/2024 Software: 8100.27 SN 80-007077 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:19	Control Test	0.077	10:20	Air Blank	0.000	10:20	Control Test	0.078	10:20	Air Blank	0.000	10:21	Control Test	0.078	10:21	Air Blank	0.000	10:22	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test	0.050	10:38																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.050	10:40																																																																																																																																																	
Air Blank	0.000	10:40																																																																																																																																																	
Control Test	0.049	10:41																																																																																																																																																	
Air Blank	0.000	10:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:44																																																																																																																																																	
Control Test	0.079	10:45																																																																																																																																																	
Air Blank	0.000	10:46																																																																																																																																																	
Control Test	0.079	10:46																																																																																																																																																	
Air Blank	0.000	10:47																																																																																																																																																	
Control Test	0.079	10:48																																																																																																																																																	
Air Blank	0.000	10:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.200	10:51																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.199	10:53																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.198	10:54																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test	0.077	10:20																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test	0.078	10:20																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.078	10:21																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 13:28

Date of Inspection: 06/25/2024

Serial Number: 80-007077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.080	0.198	0.077
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.199	0.077
0.000	0.050	0.079	0.199	0.077
0.000	0.050	0.080	0.199	0.077


Standard Deviations	0.0004	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/25/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007077</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/25/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/25/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality