



Agency Pensacola PD

S/N 80-007159

Date In 4/2/2024DI Completion Date 4/10/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

ID

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:By SP

Acetone Stock Solution Lot # 2023-B

Attachments

- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC

- ☐
- Remain Out of Evidentiary Use**

- ### Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by Taylor Gutschow
Date: 2024.04.11 10:37:21 -0400

Digitally signed by Phil
Nicodemo
Date: 2024.04.12 10:14:39 -04'00'

Tech Review / Date

Admin Review / Date

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:01
Control Test	0.050	13:02
Air Blank	0.000	13:03
Control Test	0.051	13:03
Air Blank	0.000	13:04
Control Test	0.050	13:05
Air Blank	0.000	13:05
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	



Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:07
Control Test	0.080	13:07
Air Blank	0.000	13:08
Control Test	0.080	13:09
Air Blank	0.000	13:09
Control Test	0.079	13:10
Air Blank	0.000	13:10
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	



Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:12
Control Test	0.197	13:12
Air Blank	0.000	13:13
Control Test	0.198	13:14
Air Blank	0.000	13:14
Control Test	0.198	13:15
Air Blank	0.000	13:15
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	



Operator's Signature

DGS

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:17
Control Test	0.082	13:17
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:19
Control Test	0.081	13:19
Air Blank	0.000	13:20
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7099	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA P.D.
Time of Inspection: 14:03

Date of Inspection: 04/10/2024

Serial Number: 80-007159
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.049	0.055 / 0.079	0.196	0.081
0.000	0.049	0.077 / 0.079	0.196	0.081
0.000	0.049	0.077 / 0.079	0.196	0.081
0.000	0.050	0.077 / 0.079	0.196	0.081
0.000	0.050	0.078 / 0.079	0.195	0.081
0.000	0.050	0.078 / 0.079	0.195	0.081
0.000	0.050	0.078 / 0.079	0.195	0.081
0.000	0.050	0.078 / 0.079	0.195	0.080
0.000	0.051	0.078 / 0.079	0.196	0.080
0.000	0.051	0.079 / 0.079	0.196	0.081

Standard Deviations	0.0007	0.0072 / 0.0000	0.0005	0.0004
---------------------	--------	-----------------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance- SIM JAR LOOSE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/10/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007159, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007159</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PENSACOLA P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/10/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:03</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Shayla Platt
SHAYLA D PLATT,
Department Inspector

04/10/2024

Date

Service • Integrity • Respect • Quality