

INSTRUMENT PROCESSING SHEET

Agency Hardee CSO

S/N 80-007160

Florida Department of
Law Enforcement

Date In 10/15/2024 DI Completion Date 10/17/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE[illegible]

Barometric Pressure Gauge _____ ID # _____				By _____	
Simulator	Serial #	Lot #	Expiration		
0.000		N/A	N/A		
0.040					
0.100					
0.200					
0.300					
0.080 DGS	N/A				
☐ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050					
0.080					
0.200					
0.080 DGS	N/A				
Notes/Suggested Service: <u>Added static bag. Checked breath tube screen and replaced o-rings on 10/16. Finished Quality Checks on 10/17. (TDG)</u>					

Barometric Pressure ID# 26932		By TDG	
Gauge 1018		Instrument 1017	
Mouth Alcohol Solution Lot # 2023-A			
Acetone Stock Solution Lot # 2023-B			
Simulator	Serial Number		
0.000	MP6284		
Interferent	MP6285		
0.050	MP6286		
0.080	MP6287		
0.200	MP6288		
Attachments			
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC			
☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use			
☒ Conduct an Agency Inspection Before Evidentiary Use			
Digitally signed by Shayla Platt		Digitally signed by Shayla Platt	
Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2024.10.16 14:45:28 -0400		Date: 2024.10.18 16:01:15 -0400	
Tech Review / Date		Admin Review / Date	

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
☒	☒	☒	☒																																																																																																																																																
✓ HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/17/2024 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:34	Control Test	0.050	13:35	Air Blank	0.000	13:36	Control Test	0.050	13:36	Air Blank	0.000	13:37	Control Test	0.050	13:38	Air Blank	0.000	13:38	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/17/2024 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:41	Control Test	0.079	13:42	Air Blank	0.000	13:42	Control Test	0.078	13:43	Air Blank	0.000	13:44	Control Test	0.078	13:44	Air Blank	0.000	13:45	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		✓ HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/17/2024 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:23	Control Test	0.196	13:24	Air Blank	0.000	13:24	Control Test	0.196	13:25	Air Blank	0.000	13:25	Control Test	0.197	13:26	Air Blank	0.000	13:27	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		✓ HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/17/2024 Software: 8100.27 NS <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:39	Control Test	0.079	11:40	Air Blank	0.000	11:40	Control Test	0.080	11:41	Air Blank	0.000	11:41	Control Test	0.079	11:41	Air Blank	0.000	11:42	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:34																																																																																																																																																	
Control Test	0.050	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.050	13:36																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test	0.050	13:38																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.079	13:42																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.078	13:43																																																																																																																																																	
Air Blank	0.000	13:44																																																																																																																																																	
Control Test	0.078	13:44																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:23																																																																																																																																																	
Control Test	0.196	13:24																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.196	13:25																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.197	13:26																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test	0.079	11:40																																																																																																																																																	
Air Blank	0.000	11:40																																																																																																																																																	
Control Test	0.080	11:41																																																																																																																																																	
Air Blank	0.000	11:41																																																																																																																																																	
Control Test	0.079	11:41																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 15:41

Date of Inspection: 10/17/2024

Serial Number: 80-007160
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.078	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.051	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.051	0.079	0.197	0.079


Standard Deviations	0.0004	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/17/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007160, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007160</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HARDEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/17/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:41</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

10/17/2024

Date

TAYLOR D GUTSCHOW,
Department Inspector

Service • Integrity • Respect • Quality