



INSTRUMENT PROCESSING SHEET

Agency Broward CSOS/N 80-007434Florida Department of
Law EnforcementDate In 03/13/2024 DI Completion Date 03/14/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		03/14/2024	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>244</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		03/14/2024	Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm 36 mm 53 mm 103 mm																																
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	<table border="1"><thead><tr><th>Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other</td><td></td></tr></tbody></table>					Maintenance By	Date	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other						
Simulator	Serial #	Lot #/Exp																																				
0.050	MP6286	202303K 03/29/2025																																				
0.080	MP6287	202303L 03/29/2025																																				
0.200	MP6288	202304C 04/05/2025																																				
0.080 DGS	N/A	01923080A3 02/05/2025																																				
Maintenance By	Date																																					
☐ Battery Replacement																																						
☐ Dry Gas Regulator Replacement																																						
☐ Breath Tube Replacement																																						
☐ Other																																						
Calibration Adjustment			Department Inspection																																			
Barometric Pressure Gauge ID #			Barometric Pressure ID# <u>26932</u>																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>							
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																		
Simulator	Serial Number																																					
0.000	MP6284																																					
Interferent	MP6285																																					
0.050	MP6286																																					
0.080	MP6287																																					
0.200	MP6288																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other</td></tr></tbody></table>						Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
Attachments																																						
☒ Form 41	☐ Post-Stability Checks																																					
☒ Stability Checks	☐ Flow Calibration																																					
☒ Calibration Certificate	☐ Form 40																																					
☐ Calibration Adjustment	☐ Other																																					
Notes/Suggested Service:			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																			
			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.03.15 09:09:26 -04'00'																																			
			Shayla Platt Digitally signed by Shayla Platt Date: 2024.03.20 21:36:33 -04'00'																																			
			Tech Review / Date Admin Review / Date																																			

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																	
✓		✓		✓		✓																																																																																																																																																	
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/14/2024 SN 80-007434 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	10:35	Control Test	0.049	10:36	Air Blank	0.000	10:36	Control Test	0.049	10:37	Air Blank	0.000	10:38	Control Test	0.050	10:38	Air Blank	0.000	10:39	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/14/2024 SN 80-007434 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	10:42	Control Test	0.079	10:43	Air Blank	0.000	10:43	Control Test	0.080	10:44	Air Blank	0.000	10:44	Control Test	0.080	10:45	Air Blank	0.000	10:46	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/14/2024 SN 80-007434 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test</td><td>0.202</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2013</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	10:49	Control Test	0.201	10:50	Air Blank	0.000	10:50	Control Test	0.202	10:51	Air Blank	0.000	10:51	Control Test	0.201	10:52	Air Blank	0.000	10:53	Control Test Stats			Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/14/2024 SN 80-007434 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	10:54	Control Test	0.079	10:54	Air Blank	0.000	10:55	Control Test	0.079	10:55	Air Blank	0.000	10:56	Control Test	0.079	10:56	Air Blank	0.000	10:56	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:35																																																																																																																																																					
Control Test	0.049	10:36																																																																																																																																																					
Air Blank	0.000	10:36																																																																																																																																																					
Control Test	0.049	10:37																																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																																					
Control Test	0.050	10:38																																																																																																																																																					
Air Blank	0.000	10:39																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0493																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	1.1703																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:42																																																																																																																																																					
Control Test	0.079	10:43																																																																																																																																																					
Air Blank	0.000	10:43																																																																																																																																																					
Control Test	0.080	10:44																																																																																																																																																					
Air Blank	0.000	10:44																																																																																																																																																					
Control Test	0.080	10:45																																																																																																																																																					
Air Blank	0.000	10:46																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0797																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7247																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:49																																																																																																																																																					
Control Test	0.201	10:50																																																																																																																																																					
Air Blank	0.000	10:50																																																																																																																																																					
Control Test	0.202	10:51																																																																																																																																																					
Air Blank	0.000	10:51																																																																																																																																																					
Control Test	0.201	10:52																																																																																																																																																					
Air Blank	0.000	10:53																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.2013																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2868																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:54																																																																																																																																																					
Control Test	0.079	10:54																																																																																																																																																					
Air Blank	0.000	10:55																																																																																																																																																					
Control Test	0.079	10:55																																																																																																																																																					
Air Blank	0.000	10:56																																																																																																																																																					
Control Test	0.079	10:56																																																																																																																																																					
Air Blank	0.000	10:56																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0790																																																																																																																																																						
Std Dev	0.0000																																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																																						
Operator's Signature		Operator's Signature		Operator's Signature		Operator's Signature																																																																																																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:45

Date of Inspection: 03/14/2024

Serial Number: 80-007434
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080

Standard Deviations	0.0003	0.0004	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/14/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007434, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007434</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/14/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:45</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



03/14/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality