



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 11:51

Date of Inspection: 05/30/2024

Serial Number: 80-007439  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.196	0.079
Standard Deviations	0.0000	0.0004	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO


Signature and Printed Name

05/30/2024  
Date

STABILITY CHECKS


BRKER COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007439  
 05/30/2024  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.049	09:17
Air Blank	0.000	09:18
Control Test	0.049	09:18
Air Blank	0.000	09:19
Control Test	0.049	09:20
Air Blank	0.000	09:20
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
 -----  
 Operator's Signature


BRKER COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007439  
 05/30/2024  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.080	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:23
Air Blank	0.000	09:24
Control Test	0.079	09:24
Air Blank	0.000	09:25
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

  
 -----  
 Operator's Signature

BRKER COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007439  
 05/30/2024  
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:27
Control Test	0.197	09:27
Air Blank	0.000	09:28
Control Test	0.196	09:29
Air Blank	0.000	09:29
Control Test	0.197	09:30
Air Blank	0.000	09:30
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

  
 -----  
 Operator's Signature

BRKER COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007439  
 05/30/2024  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.079	09:33
Air Blank	0.000	09:33
Control Test	0.079	09:34
Air Blank	0.000	09:34
Control Test	0.079	09:34
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D65

  
 -----  
 Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 09:05

Date of Inspection: 05/30/2024

Serial Number: 80-007439  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

05/30/2024  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/30/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:51</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/30/2024

Date

  
PHIL NICODEMO,  
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality