



Agency **Florida Highway Patrol**

S/N 80-007441

Date In 12/10/2024 DI Completion Date 12/16/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2024
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Stability Checks 80-007441 12/10/24 DA

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007441
12/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.048	13:51
Air Blank	0.000	13:52
Control Test	0.048	13:53
Air Blank	0.000	13:53
Control Test	0.048	13:54
Air Blank	0.000	13:54
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007441
12/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:35
Control Test	0.080	13:36
Air Blank	0.000	13:36
Control Test	0.079	13:37
Air Blank	0.000	13:38
Control Test	0.079	13:38
Air Blank	0.000	13:39
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

wet

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007441
12/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:46
Control Test	0.199	13:46
Air Blank	0.000	13:47
Control Test	0.199	13:48
Air Blank	0.000	13:48
Control Test	0.199	13:49
Air Blank	0.000	13:49
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007441
12/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:57
Control Test	0.000	13:57
Air Blank	0.000	13:57
Control Test	0.000	13:58
Air Blank	0.000	13:58
Control Test	0.000	13:58
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007441
12/10/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:59
Control Test	0.079	14:00
Air Blank	0.000	14:00
Control Test	0.079	14:01
Air Blank	0.000	14:01
Control Test	0.079	14:01
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

I didn't have the DGS connector plugged into calibration inlet port on first DGS attempt. Ensured connection and repeated with nominal results.

12/10/24 DA

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 10:42

Date of Inspection: 12/16/2024

Serial Number: 80-007441

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.048	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080

Standard Deviations	0.0003	0.0003	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

12/16/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007441, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007441</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/16/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/16/2024

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality