



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: FFWCC- Fort Myers Instrument Serial Number: 80-007169 and 80-007494

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>Josh Arbogast</u>	The following records were audited: <input checked="" type="checkbox"/> Instrument Registration <input checked="" type="checkbox"/> Agency Inspection Reports <input checked="" type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input checked="" type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____
Date of Notification: <u>08/08/2024</u>	
Comments: <u>Compliant with 11D-8, FAC.</u> _____ _____ _____	

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input checked="" type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
Comments: <u>ARS and DGS were non-expired. Most recent Agency Inspections, breath test procedure, and</u> <u>agency inspection procedures were posted.</u> _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		

Taylor Gutschow Digitally signed by Taylor
Gutschow
Date: 2024.08.09 15:12:35 -04'00'

Signature of Alcohol Testing Program Staff Member

08/09/2024

Date