



Agency Florida Department of Law Enforcement

S/N 80-000227

Date In N/ADI Completion Date 05/20/2025

Shi

☐ P/U

☐ H/D

Calibration Adjustment

By_

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Evidential loaner will be retained
by FDLE until requested by an agency. (TDG 5/19/25)

Department Inspection

By TDG

Attachments

Instrument Complies with Chapter 11D-8, FAC

☐ Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

☐ **Remain Out of Evidentiary Use**

■ Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by Shayla

Phil
Nicodemo

Digitally signed by Phil Nicodem
Date: 2025.05.29 11:15:44 -04'00'

Shayla Platt

Platt

Date: 2025.05.31 14:52:17
-04'00'

Tech Review / Date

Admin Review / Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓ FILE ProxiLyzer - Proxiol Analyzer Model: 8100 05/19/2025 Software: 8100.27 Test Time Air Blank 0.000 14:11 Control Test 0.049 14:12 Air Blank 0.000 14:13 Control Test 0.049 14:13 Air Blank 0.000 14:14 Control Test 0.048 14:15 Air Blank 0.000 14:15 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel. Std Dev(%) 1.1863 Operator's Signature 	FILE ProxiLyzer - Proxiol Analyzer Model: 8100 05/19/2025 Software: 8100.27 Test Time Air Blank 0.000 14:14 Control Test 0.082 14:14 Air Blank 0.000 14:15 Control Test 0.079 14:16 Air Blank 0.000 14:16 Control Test 0.079 14:17 Air Blank 0.000 14:17 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel. Std Dev(%) 0.7277 Operator's Signature 	FILE ProxiLyzer - Proxiol Analyzer Model: 8100 05/19/2025 Software: 8100.27 Test Time Air Blank 0.000 14:18 Control Test 0.200 14:19 Air Blank 0.000 14:19 Control Test 0.198 14:20 Air Blank 0.000 14:21 Control Test 0.198 14:21 Air Blank 0.000 14:22 Control Test Stats Average 0.1987 Std Dev 0.0012 Rel. Std Dev(%) 0.5812 Operator's Signature 	FILE ProxiLyzer - Proxiol Analyzer Model: 8100 05/19/2025 Software: 8100.27 Test Time Air Blank 0.000 13:56 Control Test 0.081 13:57 Air Blank 0.000 13:57 Control Test 0.080 13:57 Air Blank 0.000 13:58 Control Test 0.079 13:58 Air Blank 0.000 13:59 Control Test Stats Average 0.0800 Std Dev 0.0010 Rel. Std Dev(%) 1.2500 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 15:18

Date of Inspection: 05/20/2025

Serial Number: 80-000227

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078


Standard Deviations	0.0004	0.0004	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/20/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000227, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000227</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/20/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:18</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.05.22 13:13:59
-04'00'

05/20/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality